

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3483

Client ID : SSTDCCC0.4

Date Analyzed: 12/01/2025

Lab File ID: BN038222.D

Time Analyzed: 10:23

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7356	7.768	22525	10.56	12376	14.43
UPPER LIMIT	14712	8.268	45050	11.062	24752	14.93
LOWER LIMIT	3678	7.268	11262.5	10.062	6188	13.93
EPA SAMPLE NO.						
01 HW1025-PT-PAH-SOIL	8240	7.77	26555	10.56	15074	14.43
02 HW1025-PT-PAH-SOILDL	9014	7.78	29234	10.57	16868	14.43
03 PB170590BL	8151	7.77	22998	10.57	11468	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.