

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3483

Client ID : SSTDCCC0.4

Date Analyzed: 12/01/2025

Lab File ID: BN038230.D

Time Analyzed: 17:51

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8054	7.768	24560	10.56	13292	14.43
UPPER LIMIT	16108	8.268	49120	11.062	26584	14.93
LOWER LIMIT	4027	7.268	12280	10.062	6646	13.93
EPA SAMPLE NO.						
01 PB170590BS	9117	7.77	25112	10.56	10775	14.43
02 EME-GENERAL-FILLMS	8533.74	7.77	25803.8	10.56	13370.1	14.43
03 EME-GENERAL-FILLMSD	8233.96	7.76	24781.5	10.56	12691.2	14.43

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.