

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038211.D
Instrument ID: BNA_N

Contract: ALLI03
SDG NO.: Q3483
DFTPP Injection Date: 11/26/2025
DFTPP Injection Time: 19:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	78.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.4 (20.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN038212.D	11/26/2025	20:35
SSTDICC0.2	SSTDICC0.2	BN038213.D	11/26/2025	21:12
SSTDICCC0.4	SSTDICCC0.4	BN038214.D	11/26/2025	21:48
SSTDICC0.8	SSTDICC0.8	BN038215.D	11/26/2025	22:25
SSTDICC1.6	SSTDICC1.6	BN038216.D	11/26/2025	23:01
SSTDICC3.2	SSTDICC3.2	BN038217.D	11/26/2025	23:37
SSTDICC5.0	SSTDICC5.0	BN038218.D	11/27/2025	00:13