

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ALLI03</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3483</u>
Instrument ID:	<u>BNA_G</u>	Calibration Date/Time:	<u>11/18/2025 12:10</u>
Lab File ID:	<u>BG064799.D</u>	Init. Calib. Date(s):	<u>11/12/2025 11/12/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>11:17 16:04</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.145	1.128		-1.5	
Phenol-d6	1.523	1.523		0.0	
Nitrobenzene-d5	0.405	0.405		0.0	
2-Fluorobiphenyl	1.433	1.382		-3.6	
2,4,6-Tribromophenol	0.390	0.387		-0.8	
Atrazine	0.253	0.253		0.0	
Terphenyl-d14	1.017	1.058		4.0	

All other compounds must meet a minimum RRF of 0.010.