

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ALLI03</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3483</u>
Instrument ID:	<u>BNA_P</u>	Calibration Date/Time:	<u>11/26/2025 14:04</u>
Lab File ID:	<u>BP026190.D</u>	Init. Calib. Date(s):	<u>11/18/2025 11/18/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>14:31 21:22</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.246	1.185		-4.9	
Phenol-d6	1.586	1.536		-3.2	
Nitrobenzene-d5	0.352	0.339		-3.7	
2-Fluorobiphenyl	1.365	1.389		1.8	
2,4,6-Tribromophenol	0.259	0.283		9.3	
Atrazine	0.233	0.229		-1.7	
Terphenyl-d14	0.981	0.995		1.4	

All other compounds must meet a minimum RRF of 0.010.