

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG111225

Review By	rahul	Review On	11/12/2025 2:27:04 PM
Supervise By	Sohil	Supervise On	11/14/2025 12:24:17 PM
SubDirectory	BG111225	HP Acquire Method	BNA_G
		HP Processing Method	BG111225

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6910,SP6911,SP6912,SP6913,SP6914,SP6915,SP6916,SP6917
CCC	SP6913
Internal Standard/PEM	S13182,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064744.D	12 Nov 2025 10:37		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BG064745.D	12 Nov 2025 11:17		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BG064746.D	12 Nov 2025 11:58		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BG064747.D	12 Nov 2025 12:39		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BG064748.D	12 Nov 2025 13:20		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BG064749.D	12 Nov 2025 14:00	Compound#32,54,70 kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BG064750.D	12 Nov 2025 14:41		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BG064751.D	12 Nov 2025 15:23		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BG064752.D	12 Nov 2025 16:04		RC/JU	Ok,M
10	SSTDICV040	ICVBG111225	BG064753.D	12 Nov 2025 16:45		RC/JU	Ok,M
11	PB170478BL	PB170478BL	BG064754.D	12 Nov 2025 17:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG111825

Review By	Jagrut	Review On	11/19/2025 10:55:50 AM
Supervise By	Sohil	Supervise On	11/19/2025 3:40:03 PM
SubDirectory	BG111825	HP Acquire Method	BNA_G
		HP Processing Method	BG111225

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	P6910,SP6911,SP6912,SP6913,SP6914,SP6915,SP6916,SP6917
CCC	SP6913
Internal Standard/PEM	S13183,10ul/1000ul sample
ICV/I.BLK	SP6928
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064798.D	18 Nov 2025 11:30		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064799.D	18 Nov 2025 12:10		RC/JU	Ok
3	PB170598BL	PB170598BL	BG064800.D	18 Nov 2025 12:53		RC/JU	Ok
4	PB170598BS	PB170598BS	BG064801.D	18 Nov 2025 13:34		RC/JU	Ok,M
5	Q3657-01	SEWAREN-CONCRET	BG064802.D	18 Nov 2025 14:18		RC/JU	Ok
6	Q3641-01RX	TR-05-111325RX	BG064803.D	18 Nov 2025 14:59	Surrogate Fail	RC/JU	Confirms
7	Q3658-01	NB-07-111725	BG064804.D	18 Nov 2025 15:40		RC/JU	Ok,M
8	Q3658-01MS	NB-07-111725MS	BG064805.D	18 Nov 2025 16:21		RC/JU	Ok,M
9	Q3658-01MSD	NB-07-111725MSD	BG064806.D	18 Nov 2025 17:01		RC/JU	Ok,M
10	Q3652-01	TP-1	BG064807.D	18 Nov 2025 17:42		RC/JU	Ok,M
11	Q3660-01	FRAC-TANK-M21012	BG064808.D	18 Nov 2025 18:23		RC/JU	Ok
12	Q3656-01	T2-GRAVEL	BG064809.D	18 Nov 2025 19:04		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP111825

Review By	Jagrut	Review On	11/19/2025 10:53:24 AM
Supervise By	Sohil	Supervise On	11/19/2025 3:40:33 PM
SubDirectory	BP111825	HP Acquire Method	BNA_P
		HP Processing Method	BP111825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6910,SP6911,SP6912,SP6913,SP6914,SP6915,SP6916,SP6917
CCC	SP6913
Internal Standard/PEM	S13183,10ul/1000ul sample
ICV/I.BLK	SP6928
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026142.D	18 Nov 2025 13:45		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP026143.D	18 Nov 2025 14:31		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP026144.D	18 Nov 2025 15:12		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BP026145.D	18 Nov 2025 15:54		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BP026146.D	18 Nov 2025 16:35		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP026147.D	18 Nov 2025 17:57		RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP026148.D	18 Nov 2025 18:38		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BP026149.D	18 Nov 2025 20:00		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP026150.D	18 Nov 2025 21:22		RC/JU	Ok,M
10	SSTDICV040	ICVBP111825	BP026151.D	18 Nov 2025 23:25		RC/JU	Ok,M
11	PB170493TB	PB170493TB	BP026152.D	19 Nov 2025 00:06		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP112725

Review By	Jagrut	Review On	12/2/2025 4:09:34 PM
Supervise By	mohammad	Supervise On	12/3/2025 1:12:32 AM
SubDirectory	BP112725	HP Acquire Method	BNA_P
		HP Processing Method	BP111825

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6910,SP6911,SP6912,SP6913,SP6914,SP6915,SP6916,SP6917
CCC	SP6913
Internal Standard/PEM	S13184,10ul/1000ul sample
ICV/I.BLK	SP6928
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026189.D	26 Nov 2025 13:23		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP026190.D	26 Nov 2025 14:04		RC/JU	Ok
3	PB170598BL	PB170598BL	BP026191.D	26 Nov 2025 15:07		RC/JU	Ok
4	Q3483-11	HW1025-PT-BNA-SOIL	BP026192.D	26 Nov 2025 15:48	HW1025-PT-BNA-SOIL - Need 5X dilution	RC/JU	Dilution
5	Q3483-11DL	HW1025-PT-BNA-SOIL	BP026193.D	26 Nov 2025 18:32		RC/JU	Ok
6	Q3483-12	HW1025-PT-TRIAZINE	BP026194.D	26 Nov 2025 19:14		RC/JU	Ok
7	SSTDCCC040	SSTDCCC040EC	BP026195.D	26 Nov 2025 19:55		RC/JU	Ok,M

M : Manual Integration