

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3483

Client ID : SSTDCCC040

Date Analyzed: 11/18/2025

Lab File ID: BG064799.D

Time Analyzed: 12:10

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	48235	8.164	186828	10.98	151237	14.77
UPPER LIMIT	96470	8.664	373656	11.478	302474	15.274
LOWER LIMIT	24117.5	7.664	93414	10.478	75618.5	14.274
EPA SAMPLE NO.						
01 PB170598BL	39789	8.16	147034	10.98	118356	14.78
02 PB170598BS	43218	8.16	173831	10.98	141529	14.77
03 NB-07-111725MS	48188	8.16	195252	10.98	155870	14.77
04 NB-07-111725MSD	45859	8.16	185864	10.98	153717	14.77

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.