

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3483

Client ID: SSTDCCC040

Date Analyzed: 11/26/2025

Lab File ID: BP026190.D

Time Analyzed: 14:04

Instrument ID: BNA\_P

GC Column: ZB-GR ID: 0.25 (mm)

|                            | IS4 (PHN)<br>AREA # | RT #   | IS5 (CRY)<br>AREA # | RT #   | IS6 (PRY)<br>AREA # | RT #   |
|----------------------------|---------------------|--------|---------------------|--------|---------------------|--------|
| 12 HOUR STD                | 1501840             | 17.095 | 1698650             | 21.554 | 1917740             | 24.854 |
| UPPER LIMIT                | 3003680             | 17.595 | 3397300             | 22.054 | 3835480             | 25.354 |
| LOWER LIMIT                | 750920              | 16.595 | 849325              | 21.054 | 958870              | 24.354 |
| EPA SAMPLE NO.             |                     |        |                     |        |                     |        |
| 01 HW1025-PT-TRIAZINE-SOIL | 1239080             | 17.11  | 1354640             | 21.55  | 1546710             | 24.84  |

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.