

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120125\
 Data File : BN038230.D
 Acq On : 01 Dec 2025 17:51
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 01 18:21:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 28 21:04:10 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.428	-7.0	109	0.00
3	n-Nitrosodimethylamine	0.400	0.416	-4.0	107	0.00
4 S	2-Fluorophenol	0.400	0.404	-1.0	103	0.00
5 S	Phenol-d6	0.400	0.412	-3.0	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.428	-7.0	108	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	105	-0.01
8 S	Nitrobenzene-d5	0.400	0.396	1.0	107	0.00
9	Naphthalene	0.400	0.424	-6.0	107	0.00
10	Hexachlorobutadiene	0.400	0.442	-10.5	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.416	-4.0	105	0.00
12	2-Methylnaphthalene	0.400	0.419	-4.7	106	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.400	0.440	-10.0	113	-0.01
15 S	2-Fluorobiphenyl	0.400	0.429	-7.2	105	0.00
16	Acenaphthylene	0.400	0.421	-5.2	105	0.00
17	Acenaphthene	0.400	0.424	-6.0	103	-0.01
18	Fluorene	0.400	0.416	-4.0	102	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.563	-40.7#	160	-0.01
21	4-Bromophenyl-phenylether	0.400	0.430	-7.5	101	0.00
22	Hexachlorobenzene	0.400	0.446	-11.5	102	0.00
23	Atrazine	0.400	0.405	-1.3	99	-0.01
24	Pentachlorophenol	0.400	0.448	-12.0	120	0.00
25	Phenanthrene	0.400	0.416	-4.0	96	0.00
26	Anthracene	0.400	0.414	-3.5	100	0.00
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	92	0.00
28	Fluoranthene	0.400	0.406	-1.5	96	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	91	0.00
30	Pyrene	0.400	0.433	-8.2	95	0.00
31 S	Terphenyl-d14	0.400	0.416	-4.0	91	0.00
32	Benzo(a)anthracene	0.400	0.428	-7.0	97	0.00
33	Chrysene	0.400	0.440	-10.0	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.431	-7.7	91	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.425	-6.2	103	-0.01
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	96	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	93	0.00
39 C	Benzo(a)pyrene	0.400	0.405	-1.3	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.428	-7.0	107	-0.01
41	Benzo(g,h,i)perylene	0.400	0.418	-4.5	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4