



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.:	Q3483	Analytical Method:	SW8270-Modified
Client:	Alliance Technical Group, LLC - Newark	DataFile:	BN038234.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3618-02MS	Client Sample ID:	EME-GENERAL-FILLMS								
Atrazine	15.1	0	12.9	ug/Kg	85				20	150	