



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: ALLI03
Lab Code: ACE SDG No.: Q3483
Instrument ID: MSVOA_Y Calibration Date/Time: 11/06/2025 08:42
Lab File ID: VY023692.D Init. Calib. Date(s): 11/04/2025 11/04/2025
Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:19 11:27
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Naphthalene	50.000	42.79	14.428	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.