

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ALLI03</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3483</u>
Lab File ID:	<u>VY023692.D</u>	Date Analyzed:	<u>11/06/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>08:42</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	343829	13.346				
UPPER LIMIT	687658	13.846				
LOWER LIMIT	171915	12.846				
EPA SAMPLE NO.						
HW1025-PT-VOA-SOIL	354034	13.35				
HW1025-PT-VOA-SOILDL	374577	13.35				
VY1106SBL01	415434	13.35				
VY1106SBS01	370507	13.35				
VY1106SBSD01	361529	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.