

Duplicate Sample Summary

Client: Aramark Uniforms	SDG No.: Q3493
Project: Monthly 2025	Sample ID: LB137710BS
Client ID: LB137710BSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.8		17.0		1	1.18		10/30/2025

Duplicate Sample Summary

Client: Aramark Uniforms	SDG No.: Q3493
Project: Monthly 2025	Sample ID: Q3477-03
Client ID: SOUTH-MAHWAH-WATERDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	7.40		7.50		1	1.34		10/31/2025

Duplicate Sample Summary

Client:	Aramark Uniforms	SDG No.:	Q3493
Project:	Monthly 2025	Sample ID:	Q3493-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	4680		4690		1	0.26		10/31/2025