

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3494

Analytical Method: 8270-Modified

Client: G Environmental

DataFile: BN038133.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB170381BSD	Benzo(a)anthracene	0.4	0.28	ug/L	70	22	*		70 (54)	130 (130)	20 (20)
	Benzo(b)fluoranthene	0.4	0.29	ug/L	73	24	*		70 (65)	130 (121)	20 (20)
	Benzo(k)fluoranthene	0.4	0.31	ug/L	78	20			70 (72)	130 (119)	20 (20)
	Benzo(a)pyrene	0.4	0.30	ug/L	75	26	*		70 (68)	130 (120)	20 (20)
	Benzo(g,h,i)perylene	0.4	0.32	ug/L	80	34	*		70 (76)	130 (117)	20 (20)