

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3494

Instrument ID: BNA_P

Calibration Date(s): 10/29/2025 10/29/2025

Calibration Time(s): 09:26 14:13

LAB FILE ID:		RRF2.5 = BP026028.D			RRF005 = BP026029.D		RRF010 = BP026030.D	
		RRF020 = BP026031.D			RRF040 = BP026032.D		RRF050 = BP026033.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.055	1.255	1.234	1.249	1.123	1.212	7.4
Benzaldehyde			0.855	0.938	0.809	0.670	0.802	11.5
Phenol-d6		1.366	1.580	1.574	1.606	1.425	1.543	6.8
bis(2-Chloroethyl)ether		1.271	1.434	1.372	1.373	1.228	1.352	5.6
2,2-oxybis(1-Chloropropane)		1.978	2.210	2.102	2.099	1.846	2.043	5.8
Acetophenone		0.481	0.548	0.519	0.518	0.455	0.506	6.1
n-Nitroso-di-n-propylamine	0.795	0.894	1.013	1.016	1.027	0.904	0.958	8.8
Nitrobenzene-d5		0.256	0.317	0.327	0.338	0.306	0.321	10.3
Hexachloroethane		0.481	0.562	0.535	0.552	0.490	0.533	6.5
Nitrobenzene		0.282	0.345	0.346	0.357	0.317	0.339	8.8
Isophorone		0.606	0.721	0.705	0.720	0.632	0.687	7.0
bis(2-Chloroethoxy)methane		0.406	0.466	0.436	0.447	0.396	0.433	5.8
Naphthalene		1.039	1.181	1.121	1.115	0.978	1.091	6.3
4-Chloroaniline		0.374	0.440	0.427	0.437	0.384	0.419	6.8
Hexachlorobutadiene		0.198	0.231	0.209	0.216	0.190	0.211	6.7
Caprolactam			0.099	0.103	0.113	0.099	0.107	7.0
2-Methylnaphthalene		0.717	0.823	0.771	0.795	0.694	0.763	6.0
Hexachlorocyclopentadiene			0.200	0.198	0.237	0.218	0.227	11.5
2-Fluorobiphenyl		1.350	1.549	1.454	1.421	1.231	1.392	7.5
1,1-Biphenyl		1.454	1.673	1.577	1.560	1.378	1.531	6.4
2-Chloronaphthalene		1.130	1.304	1.248	1.229	1.090	1.205	6.2
2-Nitroaniline		0.166	0.226	0.269	0.304	0.277	0.272	21.9
Dimethylphthalate		1.282	1.574	1.491	1.522	1.323	1.455	7.8
Acenaphthylene		1.547	1.845	1.832	1.832	1.618	1.756	7.2
2,6-Dinitrotoluene		0.142	0.196	0.234	0.264	0.245	0.237	22.6
3-Nitroaniline		0.177	0.254	0.296	0.324	0.295	0.291	20.7
Acenaphthene		1.101	1.285	1.247	1.249	1.089	1.206	6.8
Dibenzofuran		1.725	1.997	1.886	1.869	1.614	1.822	6.9
2,4-Dinitrotoluene		0.195	0.288	0.340	0.393	0.360	0.347	24.0
Diethylphthalate		1.240	1.570	1.484	1.549	1.350	1.463	8.7
4-Chlorophenyl-phenylether		0.699	0.803	0.736	0.731	0.623	0.713	8.3
Fluorene		1.347	1.621	1.501	1.450	1.260	1.427	8.8
4-Nitroaniline		0.207	0.294	0.328	0.342	0.298	0.311	16.9
n-Nitrosodiphenylamine		0.574	0.678	0.641	0.647	0.567	0.625	6.8

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG No.: Q3494

Instrument ID: BNA_P

Calibration Date(s): 10/29/2025 10/29/2025

Calibration Time(s): 09:26 14:13

LAB FILE ID:		RRF2.5 = BP026028.D			RRF005 = BP026029.D		RRF010 = BP026030.D	
		RRF020 = BP026031.D			RRF040 = BP026032.D		RRF050 = BP026033.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,4,6-Tribromophenol		0.155	0.211	0.218	0.236	0.209	0.216	14.2
4-Bromophenyl-phenylether		0.206	0.234	0.222	0.228	0.204	0.222	5.8
Hexachlorobenzene		0.239	0.273	0.256	0.258	0.226	0.253	6.5
Atrazine		0.177	0.213	0.221	0.219	0.193	0.211	9.6
Phenanthrene		1.133	1.288	1.209	1.189	1.031	1.164	7.1
Anthracene		1.064	1.262	1.195	1.177	1.044	1.154	7.1
Carbazole		0.985	1.172	1.164	1.129	0.969	1.092	7.7
Di-n-butylphthalate		0.937	1.219	1.241	1.333	1.196	1.222	11.6
Fluoranthene		1.259	1.469	1.426	1.411	1.224	1.358	7.1
Pyrene		1.201	1.488	1.395	1.427	1.217	1.353	8.2
Terphenyl-d14		0.953	1.125	1.013	1.014	0.860	0.984	8.8
Butylbenzylphthalate		0.235	0.337	0.407	0.496	0.468	0.432	26.4
3,3-Dichlorobenzidine			0.390	0.424	0.462	0.416	0.443	8.9
Benzo (a) anthracene		1.279	1.497	1.407	1.420	1.224	1.374	7.0
Chrysene		1.210	1.424	1.348	1.345	1.169	1.302	6.9
Bis(2-ethylhexyl)phthalate		0.453	0.613	0.690	0.829	0.768	0.718	20.3
Di-n-octyl phthalate			0.822	1.025	1.287	1.238	1.191	18.9
Benzo (b) fluoranthene		1.074	1.277	1.278	1.306	1.149	1.242	7.7
Benzo (k) fluoranthene		1.119	1.365	1.306	1.331	1.141	1.268	7.9
Benzo (a) pyrene		0.941	1.169	1.171	1.210	1.054	1.141	9.4
Indeno (1,2,3-cd) pyrene		1.190	1.486	1.539	1.564	1.371	1.476	10.1
Dibenzo (a,h) anthracene		0.978	1.219	1.249	1.274	1.117	1.202	9.7
Benzo (g,h,i) perylene		0.966	1.167	1.219	1.212	1.068	1.156	8.9
1,2,4,5-Tetrachlorobenzene		0.574	0.670	0.637	0.632	0.565	0.621	6.3
1,4-Dioxane		0.472	0.567	0.519	0.515	0.453	0.511	7.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1