

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q3494
Lab File ID: VX048445.D Date Analyzed: 11/03/2025
Instrument ID: MSVOA_X Time Analyzed: 09:14
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	182398	5.53	256609	6.74	229985	10.04
UPPER LIMIT	364796	6.032	513218	7.239	459970	10.537
LOWER LIMIT	91199	5.032	128305	6.239	114993	9.537
EPA SAMPLE NO.						
Field	147384	5.54	222925	6.75	210225	10.04
VX1103WBL01	160467	5.53	268071	6.74	290044	10.04
VX1103WBS01	172640	5.53	288363	6.74	217740	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.