



SAMPLE DATA

Report of Analysis

Client: Remington & Vernick Engineers
Project: Maclearie Park
Client Sample ID: SB-1025-2
Lab Sample ID: Q3495-02

Date Collected: 10/27/25 10:45
Date Received: 10/29/25
SDG No.: Q3495
Matrix: SOIL
% Solid: 75.1

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Cyanide	0.17	J	1	0.054	0.32	mg/Kg	11/03/25 10:10	11/03/25 15:18	9012B

Comments: _____

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client: Remington & Vernick Engineers
Project: Maclearie Park
Client Sample ID: TW-1025-1
Lab Sample ID: Q3495-03

Date Collected: 10/27/25 11:15
Date Received: 10/29/25
SDG No.: Q3495
Matrix: WATER
% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Cyanide	0.0024	J	1	0.0012	0.0050	mg/L	11/04/25 08:25	11/04/25 12:20	SM 4500-CN C-21 plus E-21

Comments: _____

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

RunNo.: LB137751

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.095	0.099	96	90-110	11/03/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	11/03/2025
Sample ID: CCV2 Cyanide	mg/L	0.25	0.25	100	90-110	11/03/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	11/03/2025

Initial and Continuing Calibration Verification

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

RunNo.: LB137759

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.093	0.099	94	85-115	11/04/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	11/04/2025
Sample ID: CCV2 Cyanide	mg/L	0.26	0.25	104	90-110	11/04/2025
Sample ID: CCV3 Cyanide	mg/L	0.26	0.25	104	90-110	11/04/2025

Initial and Continuing Calibration Blank Summary

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

RunNo.: LB137751

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0021	0.0025	J	0.00096	0.005	11/03/2025
Sample ID: CCB1 Cyanide	mg/L	0.0017	0.0025	J	0.00096	0.005	11/03/2025
Sample ID: CCB2 Cyanide	mg/L	0.0018	0.0025	J	0.00096	0.005	11/03/2025
Sample ID: CCB3 Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	11/03/2025

Initial and Continuing Calibration Blank Summary

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

RunNo.: LB137759

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/04/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/04/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/04/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/04/2025

Preparation Blank Summary

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: Cyanide	PB170366BL mg/Kg	0.085	0.1250	J	0.042	0.25	11/03/2025
Sample ID: Cyanide	PB170367BL mg/L	0.0019	0.0025	J	0.0012	0.005	11/04/2025

Matrix Spike Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3495
Project:	Maclearie Park	Sample ID:	Q3495-03
Client ID:	TW-1025-1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.043		0.0024	J	0.04	1	102		11/04/2025

Matrix Spike Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3495
Project:	Maclearie Park	Sample ID:	Q3495-03
Client ID:	TW-1025-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.047		0.0024	J	0.04	1	112		11/04/2025

Matrix Spike Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3495
Project:	Maclearie Park	Sample ID:	Q3520-01
Client ID:	EO-02-103125MS	Percent Solids for Spike Sample:	91.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.20		0.24	J	2.1	1	93		11/03/2025

Matrix Spike Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3495
Project:	Maclearie Park	Sample ID:	Q3520-01
Client ID:	EO-02-103125MSD	Percent Solids for Spike Sample:	91.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.30		0.24	J	2.1	1	98		11/03/2025

Duplicate Sample Summary

Client: Remington & Vernick Engineers Project: Maclearie Park Client ID: TW-1025-1DUP	SDG No.: Q3495 Sample ID: Q3495-03 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0024	J	0.0023	J	1	4		11/04/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3495
Project:	Maclearie Park	Sample ID:	Q3495-03
Client ID:	TW-1025-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.043		0.047		1	9		11/04/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3495
Project:	Maclearie Park	Sample ID:	Q3520-01
Client ID:	EO-02-103125DUP	Percent Solids for Spike Sample:	91.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	0.24	J	0.25	J	1	4		11/03/2025

Duplicate Sample Summary

Client:	Remington & Vernick Engineers	SDG No.:	Q3495
Project:	Maclearie Park	Sample ID:	Q3520-01
Client ID:	EO-02-103125MSD	Percent Solids for Spike Sample:	91.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	2.20		2.30		1	4		11/03/2025

Laboratory Control Sample Summary

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

Run No.: LB137751

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170366BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	11/03/2025

Laboratory Control Sample Summary

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

Run No.: LB137759

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170367BS							
Cyanide	mg/L	0.1	0.097		97	1	85-115	11/04/2025

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137751

Review By	rubina	Review On	11/4/2025 10:13:00 AM
Supervise By	Iwona	Supervise On	11/4/2025 10:45:59 AM
SubDirectory	LB137751	Test	Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP115412,WP115413,WP115414,WP115415,WP115416,WP115417,WP115418		
ICV Standard	W3012		
CCV Standard	WP115413		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115420		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPBCN	0.0PPBCN	CAL1	11/03/25 09:49		rubina	OK
2	5.0PPBCN	5.0PPBCN	CAL2	11/03/25 09:49		rubina	OK
3	10PPBCN	10PPBCN	CAL3	11/03/25 09:49		rubina	OK
4	50PPBCN	50PPBCN	CAL4	11/03/25 09:49		rubina	OK
5	100PPBCN	100PPBCN	CAL5	11/03/25 09:49		rubina	OK
6	250PPBCN	250PPBCN	CAL6	11/03/25 09:49		rubina	OK
7	500PPBCN	500PPBCN	CAL7	11/03/25 09:49		rubina	OK
8	ICV1	ICV1	ICV	11/03/25 15:10		rubina	OK
9	ICB1	ICB1	ICB	11/03/25 15:10		rubina	OK
10	CCV1	CCV1	CCV	11/03/25 15:10		rubina	OK
11	CCB1	CCB1	CCB	11/03/25 15:10		rubina	OK
12	PB170366BL	PB170366BL	MB	11/03/25 15:10		rubina	OK
13	PB170366BS	PB170366BS	LCS	11/03/25 15:10		rubina	OK
14	LOWPB170366	LOWPB170366	SAM	11/03/25 15:18		rubina	OK
15	HIGHPB170366	HIGHPB170366	SAM	11/03/25 15:18		rubina	OK
16	Q3495-02	SB-1025-2	SAM	11/03/25 15:18		rubina	OK
17	Q3515-01	SU-04-103125	SAM	11/03/25 15:18		rubina	OK
18	Q3516-01	HD-01-103125	SAM	11/03/25 15:18		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137751

Review By	rubina	Review On	11/4/2025 10:13:00 AM
Supervise By	Iwona	Supervise On	11/4/2025 10:45:59 AM
SubDirectory	LB137751	Test	Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP115412,WP115413,WP115414,WP115415,WP115416,WP115417,WP115418		
ICV Standard	W3012		
CCV Standard	WP115413		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115420		

19	Q3519-01	PL-01-103125	SAM	11/03/25 15:18		rubina	OK
20	Q3520-01	EO-02-103125	SAM	11/03/25 15:25		rubina	OK
21	Q3520-01DUP	EO-02-103125DUP	DUP	11/03/25 15:25		rubina	OK
22	CCV2	CCV2	CCV	11/03/25 15:25		rubina	OK
23	CCB2	CCB2	CCB	11/03/25 15:33		rubina	OK
24	Q3520-01MS	EO-02-103125MS	MS	11/03/25 15:33		rubina	OK
25	Q3520-01MSD	EO-02-103125MSD	MSD	11/03/25 15:33		rubina	OK
26	CCV3	CCV3	CCV	11/03/25 15:45		rubina	OK
27	CCB3	CCB3	CCB	11/03/25 15:45		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137759

Review By	rubina	Review On	11/5/2025 10:43:35 AM
Supervise By	Iwona	Supervise On	11/5/2025 10:48:00 AM
SubDirectory	LB137759	Test	Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP115425,WP115426,WP115427,WP115428,WP115429,WP115430,WP115431		
ICV Standard	W3012		
CCV Standard	WP115426		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115432,wp115433		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPBCN	0.0PPBCN	CAL1	11/04/25 11:23		rubina	OK
2	5.0PPBCN	5.0PPBCN	CAL2	11/04/25 11:23		rubina	OK
3	10PPBCN	10PPBCN	CAL3	11/04/25 11:23		rubina	OK
4	50PPBCN	50PPBCN	CAL4	11/04/25 11:23		rubina	OK
5	100PPBCN	100PPBCN	CAL5	11/04/25 11:23		rubina	OK
6	250PPBCN	250PPBCN	CAL6	11/04/25 11:23		rubina	OK
7	500PPBCN	500PPBCN	CAL7	11/04/25 11:23		rubina	OK
8	ICV1	ICV1	ICV	11/04/25 12:12		rubina	OK
9	ICB1	ICB1	ICB	11/04/25 12:12		rubina	OK
10	CCV1	CCV1	CCV	11/04/25 12:12		rubina	OK
11	CCB1	CCB1	CCB	11/04/25 12:13		rubina	OK
12	RL	RL	LOQ	11/04/25 12:13		rubina	OK
13	PB170367BL	PB170367BL	MB	11/04/25 12:20		rubina	OK
14	PB170367BS	PB170367BS	LCS	11/04/25 12:20		rubina	OK
15	MIDPB170367	MIDPB170367	SAM	11/04/25 12:20		rubina	OK
16	Q3495-03	TW-1025-1	SAM	11/04/25 12:20		rubina	OK
17	Q3495-03DUP	TW-1025-1DUP	DUP	11/04/25 12:20		rubina	OK
18	Q3495-03MS	TW-1025-1MS	MS	11/04/25 12:28		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB137759

Review By	rubina	Review On	11/5/2025 10:43:35 AM
Supervise By	Iwona	Supervise On	11/5/2025 10:48:00 AM
SubDirectory	LB137759	Test	Cyanide
STD. NAME	STD REF.#		
ICAL Standard	WP115425,WP115426,WP115427,WP115428,WP115429,WP115430,WP115431		
ICV Standard	W3012		
CCV Standard	WP115426		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP113838		
Chk Standard	WP115157,WP114324,WP115432,wp115433		

19	Q3495-03MSD	TW-1025-1MSD	MSD	11/04/25 12:28		rubina	OK
20	Q3530-11	LOQ-WATER-05-QT4	LOQ	11/04/25 12:32		rubina	OK
21	CCV2	CCV2	CCV	11/04/25 12:32		rubina	OK
22	CCB2	CCB2	CCB	11/04/25 12:32		rubina	OK
23	Q3530-10	LOD-MDL-WATER-04	SAM	11/04/25 12:40		rubina	OK
24	Q3530-10RE	LOD-MDL-WATER-04	SAM	11/04/25 13:31		rubina	OK
25	CCV3	CCV3	CCV	11/04/25 13:31		rubina	OK
26	CCB3	CCB3	CCB	11/04/25 13:31		rubina	OK

LAB CHRONICLE

OrderID:	Q3495	OrderDate:	10/30/2025 10:45:41 AM
Client:	Remington & Vernick Engineers	Project:	Maclearie Park
Contact:	Kyle Carlson	Location:	D51,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3495-02	SB-1025-2	SOIL			10/27/25 10:45			10/29/25
			Cyanide	9012B		11/03/25	11/03/25 15:18	
Q3495-03	TW-1025-1	WATER			10/27/25 11:15			10/29/25
			Cyanide	SM4500-CN C,E		11/04/25	11/04/25 12:20	

SOP ID : M9012B-Total, Amenable and Reactive Cyanide-21

SDG No : N/A

Start Digest Date: 11/03/2025 Time : 10:10 Temp : 124 °C

Matrix : SOIL

End Digest Date: 11/03/2025 Time : 11:40 Temp : 126 °C

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : MC-1,MC-2

Filter paper ID : N/A

Prep Technician Signature: 

Weigh By : JP

pH Meter ID : N/A

Supervisor Signature: 


Standardized Name	MLS USED	STD REF. # FROM LOG
LCSS	1.0ML	WP113838
MS/MSD SPIKE SOL.	0.40ML	WP113837
PBS003	50.0ML	W3112
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
0.25N NaOH	50.0ML	WP113836
50% v/v H2SO4	5.0ML	WP115334
51% w/v MgCL2	2.0ML	WP115335
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
S0	S0	N/A	N/A
S5.0	S5.0	N/A	N/A
S10.0	S10.0	N/A	N/A
S100.0	S100.0	N/A	N/A
S250.0	S250.0	N/A	N/A
S500.0	S500.0	N/A	N/A
ICV	ICV	0.5ML	W3012
ICB	ICB	N/A	N/A
CCV	CCV	N/A	N/A
CCB	CCB	N/A	N/A
Midrange	Midrange	N/A	N/A
HIGHSTD	HIGHSTD	5.0ML	WP113837
LOWSTD	LOWSTD	0.1ML	WP113837

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/03/2025 11:55		RIM
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB170366BL	PBS366	1.00	50	N/A	N/A	N/A	N/A	N/A	N/A
PB170366BS	LCS366	1.00	50	N/A	N/A	N/A	N/A	N/A	N/A
Q3495-02	SB-1025-2	1.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q3515-01	SU-04-103125	1.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q3516-01	HD-01-103125	1.02	50	N/A	N/A	N/A	N/A	N/A	N/A
Q3519-01	PL-01-103125	1.01	50	N/A	N/A	N/A	N/A	N/A	N/A
Q3520-01	EO-02-103125	1.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q3520-01DUP	EO-02-103125DUP	1.03	50	N/A	N/A	N/A	N/A	N/A	N/A
Q3520-01MS	EO-02-103125MS	1.04	50	N/A	N/A	N/A	N/A	N/A	N/A
Q3520-01MSD	EO-02-103125MSD	1.04	50	N/A	N/A	N/A	N/A	N/A	N/A

SOP ID : MSM4500-CN C,E-Cyanide-13

SDG No : N/A

Start Digest Date: 11/04/2025 Time : 08:25 Temp : 124 °C

Matrix : WATER

End Digest Date: 11/04/2025 Time : 09:55 Temp : 126 °C

Pipette ID : WC

Balance ID : N/A

Hood ID : HOOD#1

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : MC-1, MC-2

Filter paper ID : N/A

 Prep Technician Signature: *[Signature]*

Weigh By : N/A

pH Meter ID : N/A

 Supervisor Signature: *17*

Standard Name	MLS USED	STD REF. # FROM LOG
LCSW	1.0ML	WP113838
MS/MSD SPIKE SOL.	0.40ML	WP113837
RL CHECK	50.0ML	WP115430
PBW	50.0ML	W3112
LOQ	2.5ML	WP115433

Chemical Used	ML/SAMPLE USED	Lot Number
0.25N NaOH	50.0ML	WP113836
50% v/v H2SO4	5.0ML	WP115334
51% w/v MgCL2	2.0ML	WP115335
pH Paper 0-14	N/A	W3241
Nitrate/Nitrite Strip	N/A	W3182
Lead Acetate strip	N/A	W3134
KI-starch paper	N/A	W3155
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
S0	S0	N/A	N/A
S5.0	S5.0	N/A	N/A
S10.0	S10.0	N/A	N/A
S100.0	S100.0	N/A	N/A
S250.0	S250.0	N/A	N/A
S500.0	S500.0	N/A	N/A
ICV	ICV	0.5ML	W3012
ICB	ICB	N/A	N/A
CCV	CCV	N/A	N/A
CCB	CCB	N/A	N/A
Midrange	Midrange	2.5ML	WP113837
HIGHSTD	HIGHSTD	N/A	N/A
LOWSTD	LOWSTD	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

LOD WP115433 2.0ML, LOD WP115433 1.25ML

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/04/2025 10:10	<i>[Signature]</i> / CDC	RM (WC)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB170367BL	PBW367	50	50	>12	Negative	Negative	Negative	N/A	N/A
PB170367BS	LCS367	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3495-03DUP	TW-1025-1DUP	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3495-03MS	TW-1025-1MS	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3495-03MSD	TW-1025-1MSD	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3495-03	TW-1025-1	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3530-10	LOD-MDL-WATER-04-QT4-2025	50	50	>12	Negative	Negative	Negative	N/A	N/A
Q3530-11	LOQ-WATER-05-QT4-2025	50	50	>12	Negative	Negative	Negative	N/A	N/A