

### Initial and Continuing Calibration Verification

**Client:** Remington & Vernick Engineers

**SDG No.:** Q3495

**Project:** Maclearie Park

**RunNo.:** LB137751

| Analyte               |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|-----------------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:<br>Cyanide | ICV1 | mg/L  | 0.095  | 0.099      | 96            | 90-110                    | 11/03/2025       |
| Sample ID:<br>Cyanide | CCV1 | mg/L  | 0.24   | 0.25       | 96            | 90-110                    | 11/03/2025       |
| Sample ID:<br>Cyanide | CCV2 | mg/L  | 0.25   | 0.25       | 100           | 90-110                    | 11/03/2025       |
| Sample ID:<br>Cyanide | CCV3 | mg/L  | 0.25   | 0.25       | 100           | 90-110                    | 11/03/2025       |

## Initial and Continuing Calibration Verification

**Client:** Remington & Vernick Engineers

**SDG No.:** Q3495

**Project:** Maclearie Park

**RunNo.:** LB137759

| Analyte                           | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|-----------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>Cyanide | mg/L  | 0.093  | 0.099      | 94         | 85-115                 | 11/04/2025    |
| Sample ID: <b>CCV1</b><br>Cyanide | mg/L  | 0.24   | 0.25       | 96         | 90-110                 | 11/04/2025    |
| Sample ID: <b>CCV2</b><br>Cyanide | mg/L  | 0.26   | 0.25       | 104        | 90-110                 | 11/04/2025    |
| Sample ID: <b>CCV3</b><br>Cyanide | mg/L  | 0.26   | 0.25       | 104        | 90-110                 | 11/04/2025    |