

Initial and Continuing Calibration Blank Summary

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

RunNo.: LB137751

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0021	0.0025	J	0.00096	0.005	11/03/2025
Sample ID: CCB1 Cyanide	mg/L	0.0017	0.0025	J	0.00096	0.005	11/03/2025
Sample ID: CCB2 Cyanide	mg/L	0.0018	0.0025	J	0.00096	0.005	11/03/2025
Sample ID: CCB3 Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	11/03/2025

Initial and Continuing Calibration Blank Summary

Client: Remington & Vernick Engineers

SDG No.: Q3495

Project: Maclearie Park

RunNo.: LB137759

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/04/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/04/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/04/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	11/04/2025