

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH QC BATCHID: LB137750

| | NA | NO | YES |
|---|-------|--------|--------|
| 1.Calibration Summary met criteria. | _____ | _____ | _____✓ |
| 2.ICP Interference Check Sample Results Summary Submitted. | _____ | _____ | _____✓ |
| 3. Serial Dilution Summary (if applicable) Submitted. | _____ | _____ | _____✓ |
| 4.Laboratory Control Sample Summary (if applicable) Submitted.: | _____ | _____ | _____✓ |
| 5.Blank Contamination - If yes, list compounds and concentrations in each blank: | _____ | _____✓ | _____ |
| 6.Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. | _____ | _____ | _____✓ |
| 7.Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. | _____ | _____ | _____✓ |
| 8.Digestion Holding Time Met If not met, list number of days exceeded for each sample: | _____ | _____ | _____✓ |
| 9.Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range. | _____ | _____ | _____✓ |

ADDITIONAL COMMENTS:

Q3500-04 IS REPORTED WITH DIRECT DILUTION AS SAMPLE IS OILY AND JELLY LIKE MATRIX SO, TO AVOID CLOGGING OF SAMPLE INTRODUCTION SAMPLE IS RUN WITH 5 TIMES DILUTION.

Analyst Janvi

Date 11/4/2025 12:00:00 AM

QA REVIEW

Date