

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137725

Review By	mohan	Review On	10/31/2025 3:28:32 PM
Supervise By	jaswal	Supervise On	10/31/2025 3:31:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP87801,MP87802,MP87803,MP87804,MP87805,MP87806
ICV Standard	MP87807
CCV Standard	MP87809
ICSA Standard	
CRI Standard	MP87811
LCS Standard	
Chk Standard	MP87808,MP87810,MP87812,MP87814

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/31/25 10:56		mohan	OK
2	S0.2	S0.2	CAL2	10/31/25 10:58		mohan	OK
3	S2.5	S2.5	CAL3	10/31/25 11:01		mohan	OK
4	S5	S5	CAL4	10/31/25 11:03		mohan	OK
5	S7.5	S7.5	CAL5	10/31/25 11:09		mohan	OK
6	S10	S10	CAL6	10/31/25 11:15		mohan	OK
7	ICV95	ICV95	ICV	10/31/25 11:25		mohan	OK
8	ICB95	ICB95	ICB	10/31/25 11:27		mohan	OK
9	CCV14	CCV14	CCV	10/31/25 11:30		mohan	OK
10	CCB14	CCB14	CCB	10/31/25 11:32		mohan	OK
11	CRA	CRA	CRDL	10/31/25 11:35		mohan	OK
12	HighStd	HighStd	HIGH STD	10/31/25 11:40		mohan	OK
13	ChkStd	ChkStd	SAM	10/31/25 11:42		mohan	OK
14	PB170349BL	PB170349BL	MB	10/31/25 11:45		mohan	OK
15	PB170349BS	PB170349BS	LCS	10/31/25 11:47		mohan	OK
16	Q3495-01	SB-1025-1	SAM	10/31/25 11:55		mohan	OK
17	Q3495-01DUP	SB-1025-1DUP	DUP	10/31/25 11:57		mohan	OK
18	Q3495-01MS	SB-1025-1MS	MS	10/31/25 12:00		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137725

Review By	mohan	Review On	10/31/2025 3:28:32 PM
Supervise By	jaswal	Supervise On	10/31/2025 3:31:58 PM

STD. NAME	STD REF.#
ICAL Standard	MP87801,MP87802,MP87803,MP87804,MP87805,MP87806
ICV Standard	MP87807
CCV Standard	MP87809
ICSA Standard	
CRI Standard	MP87811
LCS Standard	
Chk Standard	MP87808,MP87810,MP87812,MP87814

19	Q3495-01MSD	SB-1025-1MSD	MSD	10/31/25 12:02		mohan	OK
20	Q3495-02	SB-1025-2	SAM	10/31/25 12:05		mohan	OK
21	CCV15	CCV15	CCV	10/31/25 12:07		mohan	OK
22	CCB15	CCB15	CCB	10/31/25 12:10		mohan	OK
23	Q3496-01	DRILL-CUTTINGS	SAM	10/31/25 12:12		mohan	OK
24	Q3497-01	OK-01-103025	SAM	10/31/25 12:15		mohan	OK
25	Q3501-01	461	SAM	10/31/25 12:17		mohan	OK
26	Q3502-04	PARTS-CLEANER	SAM	10/31/25 12:19		mohan	OK
27	Q3495-01L	SB-1025-1L	SD	10/31/25 12:25		mohan	OK
28	Q3495-01A	SB-1025-1A	PS	10/31/25 12:27		mohan	OK
29	CCV16	CCV16	CCV	10/31/25 12:29		mohan	OK
30	CCB16	CCB16	CCB	10/31/25 12:32		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137738

Review By	mohan	Review On	11/6/2025 8:29:04 AM
Supervise By	jaswal	Supervise On	11/6/2025 10:57:56 AM

STD. NAME	STD REF.#
ICAL Standard	MP87801,MP87802,MP87803,MP87804,MP87805,MP87806
ICV Standard	MP87807
CCV Standard	MP87809
ICSA Standard	
CRI Standard	MP87811
LCS Standard	
Chk Standard	MP87808,MP87810,MP87812,MP87816

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/03/25 09:30		mohan	OK
2	S0.2	S0.2	CAL2	11/03/25 09:35		mohan	OK
3	S2.5	S2.5	CAL3	11/03/25 09:37		mohan	OK
4	S5	S5	CAL4	11/03/25 09:39		mohan	OK
5	S7.5	S7.5	CAL5	11/03/25 09:44		mohan	OK
6	S10	S10	CAL6	11/03/25 09:49		mohan	OK
7	ICV96	ICV96	ICV	11/03/25 10:01		mohan	OK
8	ICB96	ICB96	ICB	11/03/25 10:03		mohan	OK
9	CCV17	CCV17	CCV	11/03/25 10:09		mohan	OK
10	CCB17	CCB17	CCB	11/03/25 10:11		mohan	OK
11	CRA	CRA	CRDL	11/03/25 10:14		mohan	OK
12	HighStd	HighStd	HIGH STD	11/03/25 10:16		mohan	OK
13	ChkStd	ChkStd	SAM	11/03/25 10:18		mohan	OK
14	PB170354BL	PB170354BL	MB	11/03/25 10:21		mohan	OK
15	PB170354BS	PB170354BS	LCS	11/03/25 10:26		mohan	OK
16	Q3484-01	TRE-25-0118	SAM	11/03/25 10:28		mohan	OK
17	Q3484-01DUP	TRE-25-0118DUP	DUP	11/03/25 10:35		mohan	OK
18	Q3492-01	252611	SAM	11/03/25 10:42		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137738

Review By	mohan	Review On	11/6/2025 8:29:04 AM
Supervise By	jaswal	Supervise On	11/6/2025 10:57:56 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87801,MP87802,MP87803,MP87804,MP87805,MP87806		
ICV Standard	MP87807		
CCV Standard	MP87809		
ICSA Standard			
CRI Standard	MP87811		
LCS Standard			
Chk Standard	MP87808,MP87810,MP87812,MP87816		

19	CCV18	CCV18	CCV	11/03/25 10:49		mohan	OK
20	CCB18	CCB18	CCB	11/03/25 10:52		mohan	OK
21	Q3495-03	TW-1025-1	SAM	11/03/25 10:54		mohan	OK
22	Q3500-02	462	SAM	11/03/25 10:56		mohan	OK
23	Q3500-03	463	SAM	11/03/25 10:58		mohan	OK
24	Q3500-04	465	SAM	11/03/25 11:01		mohan	OK
25	Q3500-05	461-LIQUID	SAM	11/03/25 11:03		mohan	OK
26	Q3502-02	PARTS-CLEANER	SAM	11/03/25 11:05		mohan	OK
27	Q3503-07	320	SAM	11/03/25 11:08		mohan	OK
28	PB170361BL	PB170361BL	MB	11/03/25 11:10		mohan	OK
29	PB170361BS	PB170361BS	LCS	11/03/25 11:15		mohan	OK
30	Q3488-03	MOO-25-0299	SAM	11/03/25 11:17		mohan	OK
31	CCV19	CCV19	CCV	11/03/25 11:20		mohan	OK
32	CCB19	CCB19	CCB	11/03/25 11:22		mohan	OK
33	Q3488-05	MOO-25-0302-0310	SAM	11/03/25 11:24		mohan	OK
34	Q3488-07	MOO-25-0311-0313	SAM	11/03/25 11:26		mohan	OK
35	Q3490-02	LAW-25-0151	SAM	11/03/25 11:29		mohan	OK
36	Q3499-02	RBR251235	SAM	11/03/25 11:31		mohan	OK
37	Q3501-02	461	SAM	11/03/25 11:33		mohan	OK
38	Q3502-01	CATCH-CAN-USED-C	SAM	11/03/25 11:35		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137738

Review By	mohan	Review On	11/6/2025 8:29:04 AM
Supervise By	jaswal	Supervise On	11/6/2025 10:57:56 AM

STD. NAME	STD REF.#
ICAL Standard	MP87801,MP87802,MP87803,MP87804,MP87805,MP87806
ICV Standard	MP87807
CCV Standard	MP87809
ICSA Standard	
CRI Standard	MP87811
LCS Standard	
Chk Standard	MP87808,MP87810,MP87812,MP87816

39	Q3503-02	322	SAM	11/03/25 11:38		mohan	OK
40	Q3503-04	323	SAM	11/03/25 11:40		mohan	OK
41	Q3503-06	324	SAM	11/03/25 11:42		mohan	OK
42	Q3504-02	72-11986	SAM	11/03/25 11:45		mohan	OK
43	CCV20	CCV20	CCV	11/03/25 11:47		mohan	OK
44	CCB20	CCB20	CCB	11/03/25 11:49		mohan	OK
45	Q3504-02DUP	72-11986DUP	DUP	11/03/25 11:51		mohan	OK
46	Q3504-02MS	72-11986MS	MS	11/03/25 12:02		mohan	OK
47	Q3504-02MSD	72-11986MSD	MSD	11/03/25 12:05		mohan	OK
48	PB170333TB	PB170333TB	MB	11/03/25 12:07		mohan	OK
49	Q3484-01L	TRE-25-0118L	SD	11/03/25 12:09		mohan	OK
50	Q3484-01A	TRE-25-0118A	PS	11/03/25 12:11		mohan	OK
51	Q3504-02L	72-11986L	SD	11/03/25 12:14		mohan	OK
52	Q3504-02A	72-11986A	PS	11/03/25 12:21		mohan	OK
53	Q3484-01MS	TRE-25-0118MS	MS	11/03/25 12:23		mohan	OK
54	Q3484-01MSD	TRE-25-0118MSD	MSD	11/03/25 12:26		mohan	OK
55	CCV21	CCV21	CCV	11/03/25 12:28		mohan	OK
56	CCB21	CCB21	CCB	11/03/25 12:31		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Janvi	Review On	11/4/2025 1:50:12 PM
Supervise By	jaswal	Supervise On	11/5/2025 10:46:19 AM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/03/25 11:43		Jaswal	OK
2	S1	S1	CAL2	11/03/25 11:47		Jaswal	OK
3	S2	S2	CAL3	11/03/25 11:51		Jaswal	OK
4	S3	S3	CAL4	11/03/25 11:55		Jaswal	OK
5	S4	S4	CAL5	11/03/25 11:59		Jaswal	OK
6	S5	S5	CAL6	11/03/25 12:04		Jaswal	OK
7	ICV01	ICV01	ICV	11/03/25 12:08		Jaswal	OK
8	LLICV01	LLICV01	LLICV	11/03/25 12:27		Jaswal	OK
9	ICB01	ICB01	ICB	11/03/25 12:31		Jaswal	OK
10	CRI01	CRI01	CRDL	11/03/25 12:35		Jaswal	OK
11	ICSA01	ICSA01	ICSA	11/03/25 12:40		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	11/03/25 12:44		Jaswal	OK
13	ICSADL	ICSADL	ICSA	11/03/25 12:48		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	11/03/25 12:52		Jaswal	OK
15	CCV01	CCV01	CCV	11/03/25 12:56		Jaswal	OK
16	CCB01	CCB01	CCB	11/03/25 13:01		Jaswal	OK
17	LR-1	LR-1	HIGH STD	11/03/25 13:06		Jaswal	OK
18	LR-2	LR-2	HIGH STD	11/03/25 13:10		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Janvi	Review On	11/4/2025 1:50:12 PM
Supervise By	jaswal	Supervise On	11/5/2025 10:46:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	LR-3	LR-3	HIGH STD	11/03/25 13:17		Jaswal	OK
20	PB170192BL	PB170192BL	MB	11/03/25 13:22		Jaswal	OK
21	PB170295TB	PB170295TB	MB	11/03/25 13:36		Jaswal	OK
22	PB170192BS	PB170192BS	LCS	11/03/25 13:45		Jaswal	OK
23	PB170333TB	PB170333TB	MB	11/03/25 13:50		Jaswal	OK
24	Q3482-08	WC-1	SAM	11/03/25 13:54		Jaswal	OK
25	Q3482-09	WC-2	SAM	11/03/25 13:59		Jaswal	OK
26	Q3482-10	WC-3	SAM	11/03/25 14:03		Jaswal	OK
27	CCV02	CCV02	CCV	11/03/25 14:07		Jaswal	OK
28	CCB02	CCB02	CCB	11/03/25 14:11		Jaswal	OK
29	Q3482-11	WC-4	SAM	11/03/25 14:16		Jaswal	OK
30	Q3482-12	WC-5	SAM	11/03/25 14:20		Jaswal	OK
31	Q3482-13	WC-6	SAM	11/03/25 14:25		Jaswal	OK
32	Q3482-14	WC-7	SAM	11/03/25 14:29		Jaswal	OK
33	Q3488-03	MOO-25-0299	SAM	11/03/25 14:34		Jaswal	OK
34	Q3488-05	MOO-25-0302-0310	SAM	11/03/25 14:38		Jaswal	OK
35	Q3488-07	MOO-25-0311-0313	SAM	11/03/25 14:42		Jaswal	OK
36	Q3490-02	LAW-25-0151	SAM	11/03/25 14:47		Jaswal	OK
37	Q3499-02	RBR251235	SAM	11/03/25 14:51		Jaswal	OK
38	Q3501-02	461	SAM	11/03/25 14:55		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Janvi	Review On	11/4/2025 1:50:12 PM
Supervise By	jaswal	Supervise On	11/5/2025 10:46:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

39	CCV03	CCV03	CCV	11/03/25 15:00		Jaswal	OK
40	CCB03	CCB03	CCB	11/03/25 15:04		Jaswal	OK
41	PB170355BL	PB170355BL	MB	11/03/25 15:08		Jaswal	OK
42	PB170355BS	PB170355BS	LCS	11/03/25 15:12		Jaswal	OK
43	Q3502-01DL	CATCH-CAN-USED-C	SAM	11/03/25 15:16	Report as Straight dilution	Jaswal	OK
44	Q3503-02	322	SAM	11/03/25 15:21		Jaswal	OK
45	Q3503-04	323	SAM	11/03/25 15:25	Cd is high, Need dilution	Jaswal	Dilution
46	Q3503-06	324	SAM	11/03/25 15:30		Jaswal	OK
47	Q3504-02	72-11986	SAM	11/03/25 15:34		Jaswal	OK
48	Q3504-02DUP	72-11986DUP	DUP	11/03/25 15:38		Jaswal	OK
49	Q3504-02L	72-11986L	SD	11/03/25 15:43		Jaswal	OK
50	Q3504-02MS	72-11986MS	MS	11/03/25 15:47		Jaswal	OK
51	CCV04	CCV04	CCV	11/03/25 15:51		Jaswal	OK
52	CCB04	CCB04	CCB	11/03/25 15:55		Jaswal	OK
53	Q3504-02MSD	72-11986MSD	MSD	11/03/25 16:00		Jaswal	OK
54	Q3504-02A	72-11986A	PS	11/03/25 16:04	0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
55	PB170357BL	PB170357BL	MB	11/03/25 16:08		Jaswal	OK
56	PB170357BS	PB170357BS	LCS	11/03/25 16:12		Jaswal	OK
57	Q3500-02	462	SAM	11/03/25 16:16		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Janvi	Review On	11/4/2025 1:50:12 PM
Supervise By	jaswal	Supervise On	11/5/2025 10:46:19 AM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

58	Q3500-03	463	SAM	11/03/25 16:20		Jaswal	OK
59	Q3500-04DL	465DL	SAM	11/03/25 16:25	Report as Straight dilution	Jaswal	OK
60	Q3500-05	461-LIQUID	SAM	11/03/25 16:29		Jaswal	OK
61	Q3497-01	OK-01-103025	SAM	11/03/25 16:33		Jaswal	OK
62	Q3501-01	461	SAM	11/03/25 16:37		Jaswal	OK
63	CCV05	CCV05	CCV	11/03/25 16:41		Jaswal	OK
64	CCB05	CCB05	CCB	11/03/25 16:45		Jaswal	OK
65	Q3502-04	PARTS-CLEANER	SAM	11/03/25 16:50		Jaswal	OK
66	Q3495-03	TW-1025-1	SAM	11/03/25 16:54		Jaswal	OK
67	Q3495-03DUP	TW-1025-1DUP	DUP	11/03/25 16:58		Jaswal	OK
68	Q3495-03L	TW-1025-1L	SD	11/03/25 17:03		Jaswal	OK
69	Q3495-03MS	TW-1025-1MS	MS	11/03/25 17:07		Jaswal	OK
70	Q3495-03MSD	TW-1025-1MSD	MSD	11/03/25 17:11		Jaswal	OK
71	Q3495-03A	TW-1025-1A	PS	11/03/25 17:15	0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
72	PB170353BL	PB170353BL	MB	11/03/25 17:19		Jaswal	OK
73	PB170353BS	PB170353BS	LCS	11/03/25 17:23		Jaswal	OK
74	Q3496-01	DRILL-CUTTINGS	SAM	11/03/25 17:27		Jaswal	OK
75	CCV06	CCV06	CCV	11/03/25 17:32		Jaswal	OK
76	CCB06	CCB06	CCB	11/03/25 17:36		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Janvi	Review On	11/4/2025 1:50:12 PM
Supervise By	jaswal	Supervise On	11/5/2025 10:46:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

77	Q3495-01	SB-1025-1	SAM	11/03/25 17:40	Q3495-01DUP bd injection	Jaswal	Not Ok
78	Q3495-01DUP	SB-1025-1DUP	DUP	11/03/25 17:44	Q3495-01DUP bad injection	Jaswal	Not Ok
79	Q3495-01L	SB-1025-1L	SD	11/03/25 17:48	Q3495-01DUP bad injection	Jaswal	Not Ok
80	Q3495-01MS	SB-1025-1MS	MS	11/03/25 17:53	Q3495-01DUP bad injection	Jaswal	Not Ok
81	Q3495-01MSD	SB-1025-1MSD	MSD	11/03/25 17:56	Q3495-01DUP bad injection	Jaswal	Not Ok
82	Q3495-01A	SB-1025-1A	PS	11/03/25 18:00	Q3495-01DUP bad injection (0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.)	Jaswal	Not Ok
83	Q3495-02	SB-1025-2	SAM	11/03/25 18:04	Not use	Jaswal	Not Ok
84	Q3475-01DL	GW-DRUMDL	SAM	11/03/25 18:08	Not use	Jaswal	Not Ok
85	Q3502-01	CATCH-CAN-USED-C	SAM	11/03/25 18:13	Not use	Jaswal	Not Ok
86	Q3519-01	PL-01-103125	SAM	11/03/25 18:17	Not use	Jaswal	Not Ok
87	CCV07	CCV07	CCV	11/03/25 18:21		Jaswal	OK
88	CCB07	CCB07	CCB	11/03/25 18:25		Jaswal	OK
89	PB170374BL	PB170374BL	MB	11/03/25 18:29	Not use	Jaswal	Not Ok
90	PB170374BS	PB170374BS	LCS	11/03/25 18:34	Not use	Jaswal	Not Ok
91	Q3512-01	10200	SAM	11/03/25 18:38	Not use	Jaswal	Not Ok
92	Q3515-01	SU-04-103125	SAM	11/03/25 18:42	Not use	Jaswal	Not Ok
93	Q3516-01	HD-01-103125	SAM	11/03/25 18:46	Not use	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Janvi	Review On	11/4/2025 1:50:12 PM
Supervise By	jaswal	Supervise On	11/5/2025 10:46:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

94	Q3520-01	EO-02-103125	SAM	11/03/25 18:51	Not use	Jaswal	Not Ok
95	Q3520-01DUP	EO-02-103125DUP	DUP	11/03/25 18:55	Not use	Jaswal	Not Ok
96	Q3520-01L	EO-02-103125L	SD	11/03/25 18:59	Not use	Jaswal	Not Ok
97	Q3520-01MS	EO-02-103125MS	MS	11/03/25 19:03	Not use	Jaswal	Not Ok
98	Q3520-01MSD	EO-02-103125MSD	MSD	11/03/25 19:07	Not use	Jaswal	Not Ok
99	CCV08	CCV08	CCV	11/03/25 19:11		Jaswal	OK
100	CCB08	CCB08	CCB	11/03/25 19:15		Jaswal	OK
101	Q3520-01A	EO-02-103125A	PS	11/03/25 19:20	Not use, (0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.)	Jaswal	Not Ok
102	Q3503-04DL	323DL	SAM	11/03/25 19:24	Not use	Jaswal	Not Ok
103	PB170382BL	PB170382BL	MB	11/03/25 19:28	Not use	Jaswal	Not Ok
104	PB170382BS	PB170382BS	LCS	11/03/25 19:33	Not use	Jaswal	Not Ok
105	Q3527-01	RW8-SP100-2025103	SAM	11/03/25 19:37	Not use	Jaswal	Not Ok
106	Q3527-02	RW8-SP303-2025103	SAM	11/03/25 19:41	Not use	Jaswal	Not Ok
107	Q3527-02DUP	RW8-SP303-2025103	DUP	11/03/25 19:45	Not use	Jaswal	Not Ok
108	Q3527-02L	RW8-SP303-2025103	SD	11/03/25 19:49	Not use	Jaswal	Not Ok
109	Q3527-02MS	RW8-SP303-2025103	MS	11/03/25 19:54	Not use	Jaswal	Not Ok
110	Q3527-02MSD	RW8-SP303-2025103	MSD	11/03/25 19:58	Not use	Jaswal	Not Ok
111	CCV09	CCV09	CCV	11/03/25 20:02		Jaswal	OK
112	CCB09	CCB09	CCB	11/03/25 20:06		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Janvi	Review On	11/4/2025 1:50:12 PM
Supervise By	jaswal	Supervise On	11/5/2025 10:46:19 AM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

113	Q3527-02A	RW8-SP303-2025103	PS	11/03/25 20:10	Not use, (0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.)	Jaswal	Not Ok
114	CCV10	CCV10	CCV	11/03/25 20:15		Jaswal	OK
115	CCB10	CCB10	CCB	11/03/25 20:19		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137761

Review By	Janvi	Review On	11/5/2025 10:46:17 AM
Supervise By	jaswal	Supervise On	11/7/2025 12:56:11 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/04/25 11:33		Jaswal	OK
2	S1	S1	CAL2	11/04/25 11:37		Jaswal	OK
3	S2	S2	CAL3	11/04/25 11:42		Jaswal	OK
4	S3	S3	CAL4	11/04/25 11:46		Jaswal	OK
5	S4	S4	CAL5	11/04/25 11:50		Jaswal	OK
6	S5	S5	CAL6	11/04/25 11:54		Jaswal	OK
7	ICV01	ICV01	ICV	11/04/25 11:58		Jaswal	OK
8	LLICV01	LLICV01	LLICV	11/04/25 12:16		Jaswal	OK
9	ICB01	ICB01	ICB	11/04/25 12:20		Jaswal	OK
10	CRI01	CRI01	CRDL	11/04/25 12:24		Jaswal	OK
11	ICSA01	ICSA01	ICSA	11/04/25 12:34		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	11/04/25 12:38		Jaswal	OK
13	ICSADL	ICSADL	ICSA	11/04/25 12:48		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	11/04/25 12:52		Jaswal	OK
15	CCV01	CCV01	CCV	11/04/25 12:56		Jaswal	OK
16	CCB01	CCB01	CCB	11/04/25 13:00		Jaswal	OK
17	LR-1	LR-1	HIGH STD	11/04/25 13:05		Jaswal	OK
18	LR-2	LR-2	HIGH STD	11/04/25 13:09		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137761

Review By	Janvi	Review On	11/5/2025 10:46:17 AM
Supervise By	jaswal	Supervise On	11/7/2025 12:56:11 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	LR-3	LR-3	HIGH STD	11/04/25 13:14		Jaswal	OK
20	Q3475-01DL	GW-DRUMDL	SAM	11/04/25 13:18	5x for Fe, Still Ag oversaturated	Jaswal	Dilution
21	Q3503-04DL	323DL	SAM	11/04/25 13:29	NOT USE	Jaswal	Not Ok
22	Q3519-01	PL-01-103125	SAM	11/04/25 13:33		Jaswal	OK
23	Q3475-01DL2	GW-DRUMDL2	SAM	11/04/25 13:39	25x for Ag	Jaswal	Confirms
24	Q3503-04DL2	323DL2	SAM	11/04/25 13:43	10X for Cd	Jaswal	Confirms
25	Q3495-01	SB-1025-1	SAM	11/04/25 13:47	NOT USE	Jaswal	Not Ok
26	Q3495-01DUP	SB-1025-1DUP	DUP	11/04/25 13:51	NOT USE	Jaswal	Not Ok
27	CCV02	CCV02	CCV	11/04/25 13:55		Jaswal	OK
28	CCB02	CCB02	CCB	11/04/25 14:03		Jaswal	OK
29	Q3495-01L	SB-1025-1L	SD	11/04/25 14:08	NOT USE	Jaswal	Not Ok
30	Q3495-01MS	SB-1025-1MS	MS	11/04/25 14:12	NOT USE	Jaswal	Not Ok
31	Q3495-01MSD	SB-1025-1MSD	MSD	11/04/25 14:16	NOT USE	Jaswal	Not Ok
32	Q3495-01A	SB-1025-1A	PS	11/04/25 14:20	NOT USE	Jaswal	Not Ok
33	PB170374BL	PB170374BL	MB	11/04/25 14:24		Jaswal	OK
34	PB170374BS	PB170374BS	LCS	11/04/25 14:28		Jaswal	OK
35	Q3515-01	SU-04-103125	SAM	11/04/25 14:32		Jaswal	OK
36	Q3516-01	HD-01-103125	SAM	11/04/25 14:36		Jaswal	OK
37	Q3520-01	EO-02-103125	SAM	11/04/25 14:40		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137761

Review By	Janvi	Review On	11/5/2025 10:46:17 AM
Supervise By	jaswal	Supervise On	11/7/2025 12:56:11 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

38	Q3520-01DUP	EO-02-103125DUP	DUP	11/04/25 14:44		Jaswal	OK
39	CCV03	CCV03	CCV	11/04/25 15:06		Jaswal	OK
40	CCB03	CCB03	CCB	11/04/25 15:10		Jaswal	OK
41	Q3520-01L	EO-02-103125L	SD	11/04/25 15:46		Jaswal	OK
42	Q3520-01MS	EO-02-103125MS	MS	11/04/25 15:50		Jaswal	OK
43	Q3520-01MSD	EO-02-103125MSD	MSD	11/04/25 15:54		Jaswal	OK
44	Q3520-01A	EO-02-103125A	PS	11/04/25 15:58	0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK
45	PB170382BL	PB170382BL	MB	11/04/25 16:02		Jaswal	OK
46	PB170382BS	PB170382BS	LCS	11/04/25 16:06		Jaswal	OK
47	CCV04	CCV04	CCV	11/04/25 16:10		Jaswal	OK
48	CCB04	CCB04	CCB	11/04/25 16:14		Jaswal	OK
49	Q3527-01	RW8-SP100-2025103	SAM	11/04/25 16:18		Jaswal	OK
50	Q3527-02	RW8-SP303-2025103	SAM	11/04/25 16:23		Jaswal	OK
51	Q3527-02DUP	RW8-SP303-2025103	DUP	11/04/25 16:27		Jaswal	OK
52	Q3527-02L	RW8-SP303-2025103	SD	11/04/25 16:31		Jaswal	OK
53	Q3527-02MS	RW8-SP303-2025103	MS	11/04/25 16:35		Jaswal	OK
54	Q3527-02MSD	RW8-SP303-2025103	MSD	11/04/25 16:39		Jaswal	OK
55	Q3527-02A	RW8-SP303-2025103	PS	11/04/25 16:43	0.1 ML OF EACH M6008 AND M6017 IN 10ML SAMPLE.	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137761

Review By	Janvi	Review On	11/5/2025 10:46:17 AM
Supervise By	jaswal	Supervise On	11/7/2025 12:56:11 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

56	Q3495-02	SB-1025-2	SAM	11/04/25 16:47		Jaswal	OK
57	Q3512-01	10200	SAM	11/04/25 16:51		Jaswal	OK
58	CCV05	CCV05	CCV	11/04/25 16:55		Jaswal	OK
59	CCB05	CCB05	CCB	11/04/25 16:59		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137798

Review By	Janvi	Review On	11/6/2025 3:46:12 PM
Supervise By	jaswal	Supervise On	11/7/2025 12:58:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/05/25 11:07		Jaswal	OK
2	S1	S1	CAL2	11/05/25 11:11		Jaswal	OK
3	S2	S2	CAL3	11/05/25 11:15		Jaswal	OK
4	S3	S3	CAL4	11/05/25 11:19		Jaswal	OK
5	S4	S4	CAL5	11/05/25 11:23		Jaswal	OK
6	S5	S5	CAL6	11/05/25 11:27		Jaswal	OK
7	ICV01	ICV01	ICV	11/05/25 11:32		Jaswal	OK
8	LLICV01	LLICV01	LLICV	11/05/25 11:36		Jaswal	OK
9	ICB01	ICB01	ICB	11/05/25 11:41		Jaswal	OK
10	CRI01	CRI01	CRDL	11/05/25 11:49		Jaswal	OK
11	ICSA01	ICSA01	ICSA	11/05/25 11:53		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	11/05/25 12:11		Jaswal	OK
13	ICSADL	ICSADL	ICSA	11/05/25 12:15		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	11/05/25 12:19		Jaswal	OK
15	CCV01	CCV01	CCV	11/05/25 12:23		Jaswal	OK
16	CCB01	CCB01	CCB	11/05/25 12:33		Jaswal	OK
17	LR-1	LR-1	HIGH STD	11/05/25 12:39		Jaswal	OK
18	LR-2	LR-2	HIGH STD	11/05/25 12:44		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137798

Review By	Janvi	Review On	11/6/2025 3:46:12 PM
Supervise By	jaswal	Supervise On	11/7/2025 12:58:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	LR-3	LR-3	HIGH STD	11/05/25 12:48		Jaswal	OK
20	Q3512-01DL	10200DL	SAM	11/05/25 12:53		Jaswal	OK
21	Q3495-01	SB-1025-1	SAM	11/05/25 12:57		Jaswal	OK
22	Q3495-01DUP	SB-1025-1DUP	DUP	11/05/25 13:01		Jaswal	OK
23	Q3495-01L	SB-1025-1L	SD	11/05/25 13:05		Jaswal	OK
24	Q3495-01MS	SB-1025-1MS	MS	11/05/25 13:09		Jaswal	OK
25	Q3495-01MSD	SB-1025-1MSD	MSD	11/05/25 13:13		Jaswal	OK
26	Q3495-01A	SB-1025-1A	PS	11/05/25 13:17	0.1 ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
27	CCV02	CCV02	CCV	11/05/25 13:21		Jaswal	OK
28	CCB02	CCB02	CCB	11/05/25 13:25		Jaswal	OK
29	PB170376BL	PB170376BL	MB	11/05/25 13:32		Jaswal	OK
30	PB170376BS	PB170376BS	LCS	11/05/25 13:36		Jaswal	OK
31	Q3524-01	TH-430-DI-1	SAM	11/05/25 13:40		Jaswal	OK
32	Q3524-02	TH-430-DI-2	SAM	11/05/25 13:45		Jaswal	OK
33	Q3524-03	TH-430-DI-3	SAM	11/05/25 13:49		Jaswal	OK
34	Q3524-04	TH-404-DI-4	SAM	11/05/25 13:53		Jaswal	OK
35	Q3524-04DUP	TH-404-DI-4DUP	DUP	11/05/25 13:58		Jaswal	OK
36	Q3524-04L	TH-404-DI-4L	SD	11/05/25 14:02		Jaswal	OK
37	Q3524-04MS	TH-404-DI-4MS	MS	11/05/25 14:06		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137798

Review By	Janvi	Review On	11/6/2025 3:46:12 PM
Supervise By	jaswal	Supervise On	11/7/2025 12:58:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

38	Q3524-04MSD	TH-404-DI-4MSD	MSD	11/05/25 14:10		Jaswal	OK
39	CCV03	CCV03	CCV	11/05/25 14:14		Jaswal	OK
40	CCB03	CCB03	CCB	11/05/25 14:18		Jaswal	OK
41	Q3524-04A	TH-404-DI-4A	PS	11/05/25 14:23	0.1 ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
42	PBS	PBS	LCS	11/05/25 14:27	Not Use	Jaswal	Not Ok
43	LCSS	LCSS	SAM	11/05/25 14:31	Not Use	Jaswal	Not Ok
44	Q3483-11	HW1025-PT-BNA-SO	SAM	11/05/25 14:38	Not Use	Jaswal	Not Ok
45	Q3483-12	HW1025-PT-TRIAZIN	SAM	11/05/25 14:42	Not Use	Jaswal	Not Ok
46	Q3483-11DL	HW1025-PT-BNA-SO	SAM	11/05/25 14:48	Not Use	Jaswal	Not Ok
47	Q3483-12DL	HW1025-PT-TRIAZIN	SAM	11/05/25 14:52	Not Use	Jaswal	Not Ok
48	CCV04	CCV04	CCV	11/05/25 15:16		Jaswal	OK
49	CCB04	CCB04	CCB	11/05/25 15:21		Jaswal	OK
50	PB170392BL	PB170392BL	MB	11/05/25 15:25		Jaswal	OK
51	PB170392BS	PB170392BS	LCS	11/05/25 15:29		Jaswal	OK
52	Q3534-01	MW-6	SAM	11/05/25 15:33		Jaswal	OK
53	Q3534-02	MW-1	SAM	11/05/25 15:38	Na high, need dilution	Jaswal	Dilution
54	Q3534-03	MW-11	SAM	11/05/25 15:42	Na high, need dilution	Jaswal	Dilution
55	Q3534-04	MW-5	SAM	11/05/25 15:47		Jaswal	OK
56	Q3534-05	MW-EB-1	SAM	11/05/25 15:51		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137798

Review By	Janvi	Review On	11/6/2025 3:46:12 PM
Supervise By	jaswal	Supervise On	11/7/2025 12:58:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

57	Q3534-06	MW-EB-2	SAM	11/05/25 15:55	Na high, need dilution	Jaswal	Dilution
58	Q3534-07	MW-EB-3	SAM	11/05/25 16:00	Na high, need dilution	Jaswal	Dilution
59	Q3534-09	FB-1	SAM	11/05/25 16:04		Jaswal	OK
60	CCV05	CCV05	CCV	11/05/25 16:17		Jaswal	OK
61	CCB05	CCB05	CCB	11/05/25 16:25		Jaswal	OK
62	Q3537-03	MW-17-PURGE-WAT	SAM	11/05/25 16:30	CCB,CCV Fail for Na	Jaswal	Not Ok
63	Q3541-01	N48969	SAM	11/05/25 16:34	CCB,CCV Fail for Na	Jaswal	Not Ok
64	Q3542-01	AUD-25-182-194	SAM	11/05/25 16:38	CCB,CCV Fail for Na	Jaswal	Not Ok
65	Q3542-01DUP	AUD-25-182-194DUP	DUP	11/05/25 16:42	CCB,CCV Fail for Na	Jaswal	Not Ok
66	Q3542-01L	AUD-25-182-194L	SD	11/05/25 16:47	CCB,CCV Fail for Na	Jaswal	Not Ok
67	Q3542-01MS	AUD-25-182-194MS	MS	11/05/25 16:51	CCB,CCV Fail for Na	Jaswal	Not Ok
68	Q3542-01MSD	AUD-25-182-194MSD	MSD	11/05/25 16:55	CCB,CCV Fail for Na	Jaswal	Not Ok
69	Q3542-01A	AUD-25-182-194A	PS	11/05/25 17:00	CCB,CCV Fail for Na (0.1 ml of each m6008 and m6017 in 10ml sample.)	Jaswal	Not Ok
70	CCV06	CCV06	CCV	11/05/25 17:04		Jaswal	OK
71	CCB06	CCB06	CCB	11/05/25 17:08		Jaswal	OK