



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB95	Mercury	0.076	+/-0.2	U	0.20	CV	10/31/2025	11:27	LB137725

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB14	Mercury	0.076	+/-0.2	U	0.20	CV	10/31/2025	11:32	LB137725
CCB15	Mercury	0.076	+/-0.2	U	0.20	CV	10/31/2025	12:10	LB137725
CCB16	Mercury	0.076	+/-0.2	U	0.20	CV	10/31/2025	12:32	LB137725

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB96	Mercury	0.076	+/-0.2	U	0.20	CV	11/03/2025	10:03	LB137738

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB17	Mercury	0.076	+/-0.2	U	0.20	CV	11/03/2025	10:11	LB137738
CCB18	Mercury	0.076	+/-0.2	U	0.20	CV	11/03/2025	10:52	LB137738
CCB19	Mercury	0.076	+/-0.2	U	0.20	CV	11/03/2025	11:22	LB137738
CCB20	Mercury	0.076	+/-0.2	U	0.20	CV	11/03/2025	11:49	LB137738
CCB21	Mercury	0.076	+/-0.2	U	0.20	CV	11/03/2025	12:31	LB137738

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	11/03/2025	12:31	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	12:31	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	12:31	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	12:31	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	12:31	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	12:31	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	12:31	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	12:31	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	12:31	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	12:31	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	12:31	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	12:31	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	12:31	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	12:31	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	12:31	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	12:31	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	12:31	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	12:31	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	12:31	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	12:31	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	12:31	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	12:31	LB137750

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	11/03/2025	13:01	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	13:01	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	13:01	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	13:01	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	13:01	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	13:01	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	13:01	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	13:01	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	13:01	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	13:01	LB137750
	Iron	41.6	+/-50	J	100	P	11/03/2025	13:01	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	13:01	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	13:01	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	13:01	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	13:01	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	13:01	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	13:01	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	13:01	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	13:01	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	13:01	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	13:01	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	13:01	LB137750
CCB02	Aluminum	26.9	+/-50	J	100	P	11/03/2025	14:11	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	14:11	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	14:11	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	14:11	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	14:11	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	14:11	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	14:11	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	14:11	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	14:11	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	14:11	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	14:11	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	14:11	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	14:11	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	14:11	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	14:11	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	14:11	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	14:11	LB137750

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/03/2025	14:11	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	14:11	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	14:11	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	14:11	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	14:11	LB137750
CCB03	Aluminum	13.7	+/-50	J	100	P	11/03/2025	15:04	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	15:04	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	15:04	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	15:04	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	15:04	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	15:04	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	15:04	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	15:04	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	15:04	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	15:04	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	15:04	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	15:04	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	15:04	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	15:04	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	15:04	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	15:04	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	15:04	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	15:04	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	15:04	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	15:04	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	15:04	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	15:04	LB137750
CCB04	Aluminum	11.3	+/-50	U	100	P	11/03/2025	15:55	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	15:55	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	15:55	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	15:55	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	15:55	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	15:55	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	15:55	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	15:55	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	15:55	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	15:55	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	15:55	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	15:55	LB137750

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/03/2025	15:55	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	15:55	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	15:55	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	15:55	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	15:55	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	15:55	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	15:55	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	15:55	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	15:55	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	15:55	LB137750
CCB05	Aluminum	11.3	+/-50	U	100	P	11/03/2025	16:45	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	16:45	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	16:45	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	16:45	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	16:45	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	16:45	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	16:45	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	16:45	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	16:45	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	16:45	LB137750
	Iron	49.8	+/-50	J	100	P	11/03/2025	16:45	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	16:45	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	16:45	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	16:45	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	16:45	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	16:45	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	16:45	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	16:45	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	16:45	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	16:45	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	16:45	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	16:45	LB137750
CCB06	Aluminum	11.3	+/-50	U	100	P	11/03/2025	17:36	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	17:36	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	17:36	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	17:36	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	17:36	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	17:36	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	17:36	LB137750

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	17:36	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	17:36	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	17:36	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	17:36	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	17:36	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	17:36	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	17:36	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	17:36	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	17:36	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	17:36	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	17:36	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	17:36	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	17:36	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	17:36	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	17:36	LB137750
CCB07	Aluminum	11.3	+/-50	U	100	P	11/03/2025	18:25	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	18:25	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	18:25	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	18:25	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	18:25	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	18:25	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	18:25	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	18:25	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	18:25	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	18:25	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	18:25	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	18:25	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	18:25	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	18:25	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	18:25	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	18:25	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	18:25	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	18:25	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	18:25	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	18:25	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	18:25	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	18:25	LB137750
CCB08	Aluminum	11.3	+/-50	U	100	P	11/03/2025	19:15	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	19:15	LB137750

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	19:15	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	19:15	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	19:15	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	19:15	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	19:15	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	19:15	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	19:15	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	19:15	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	19:15	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	19:15	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	19:15	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	19:15	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	19:15	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	19:15	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	19:15	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	19:15	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	19:15	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	19:15	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	19:15	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	19:15	LB137750
CCB09	Aluminum	11.3	+/-50	U	100	P	11/03/2025	20:06	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	20:06	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	20:06	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	20:06	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	20:06	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	20:06	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	20:06	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	20:06	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	20:06	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	20:06	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	20:06	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	20:06	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	20:06	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	20:06	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	20:06	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	20:06	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	20:06	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	20:06	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	20:06	LB137750

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	20:06	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	20:06	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	20:06	LB137750
CCB10	Aluminum	11.3	+/-50	U	100	P	11/03/2025	20:19	LB137750
	Antimony	6.76	+/-25	U	50.0	P	11/03/2025	20:19	LB137750
	Arsenic	5.12	+/-10	U	20.0	P	11/03/2025	20:19	LB137750
	Barium	14.6	+/-50	U	100	P	11/03/2025	20:19	LB137750
	Beryllium	0.56	+/-3	U	6.00	P	11/03/2025	20:19	LB137750
	Cadmium	0.50	+/-3	U	6.00	P	11/03/2025	20:19	LB137750
	Calcium	234	+/-1000	U	2000	P	11/03/2025	20:19	LB137750
	Chromium	2.12	+/-5	U	10.0	P	11/03/2025	20:19	LB137750
	Cobalt	2.26	+/-15	U	30.0	P	11/03/2025	20:19	LB137750
	Copper	4.60	+/-10	U	20.0	P	11/03/2025	20:19	LB137750
	Iron	23.4	+/-50	U	100	P	11/03/2025	20:19	LB137750
	Lead	2.30	+/-6	U	12.0	P	11/03/2025	20:19	LB137750
	Magnesium	244	+/-1000	U	2000	P	11/03/2025	20:19	LB137750
	Manganese	5.94	+/-10	U	20.0	P	11/03/2025	20:19	LB137750
	Nickel	3.06	+/-20	U	40.0	P	11/03/2025	20:19	LB137750
	Potassium	918	+/-1000	U	2000	P	11/03/2025	20:19	LB137750
	Selenium	9.64	+/-10	U	20.0	P	11/03/2025	20:19	LB137750
	Silver	1.62	+/-5	U	10.0	P	11/03/2025	20:19	LB137750
	Sodium	868	+/-1000	U	2000	P	11/03/2025	20:19	LB137750
	Thallium	4.38	+/-20	U	40.0	P	11/03/2025	20:19	LB137750
	Vanadium	6.26	+/-20	U	40.0	P	11/03/2025	20:19	LB137750
	Zinc	16.7	+/-20	U	40.0	P	11/03/2025	20:19	LB137750

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	11/04/2025	12:20	LB137761
	Antimony	6.76	+/-25	U	50.0	P	11/04/2025	12:20	LB137761
	Arsenic	5.12	+/-10	U	20.0	P	11/04/2025	12:20	LB137761
	Barium	14.6	+/-50	U	100	P	11/04/2025	12:20	LB137761
	Beryllium	0.56	+/-3	U	6.00	P	11/04/2025	12:20	LB137761
	Cadmium	0.50	+/-3	U	6.00	P	11/04/2025	12:20	LB137761
	Calcium	234	+/-1000	U	2000	P	11/04/2025	12:20	LB137761
	Chromium	2.12	+/-5	U	10.0	P	11/04/2025	12:20	LB137761
	Cobalt	2.26	+/-15	U	30.0	P	11/04/2025	12:20	LB137761
	Copper	4.60	+/-10	U	20.0	P	11/04/2025	12:20	LB137761
	Iron	23.4	+/-50	U	100	P	11/04/2025	12:20	LB137761
	Lead	2.30	+/-6	U	12.0	P	11/04/2025	12:20	LB137761
	Magnesium	244	+/-1000	U	2000	P	11/04/2025	12:20	LB137761
	Manganese	5.94	+/-10	U	20.0	P	11/04/2025	12:20	LB137761
	Nickel	3.06	+/-20	U	40.0	P	11/04/2025	12:20	LB137761
	Potassium	918	+/-1000	U	2000	P	11/04/2025	12:20	LB137761
	Selenium	9.64	+/-10	U	20.0	P	11/04/2025	12:20	LB137761
	Silver	1.62	+/-5	U	10.0	P	11/04/2025	12:20	LB137761
	Sodium	868	+/-1000	U	2000	P	11/04/2025	12:20	LB137761
	Thallium	4.38	+/-20	U	40.0	P	11/04/2025	12:20	LB137761
	Vanadium	6.26	+/-20	U	40.0	P	11/04/2025	12:20	LB137761
	Zinc	16.7	+/-20	U	40.0	P	11/04/2025	12:20	LB137761

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	11/04/2025	13:00	LB137761
	Antimony	6.76	+/-25	U	50.0	P	11/04/2025	13:00	LB137761
	Arsenic	5.12	+/-10	U	20.0	P	11/04/2025	13:00	LB137761
	Barium	14.6	+/-50	U	100	P	11/04/2025	13:00	LB137761
	Beryllium	0.56	+/-3	U	6.00	P	11/04/2025	13:00	LB137761
	Cadmium	0.50	+/-3	U	6.00	P	11/04/2025	13:00	LB137761
	Calcium	234	+/-1000	U	2000	P	11/04/2025	13:00	LB137761
	Chromium	2.12	+/-5	U	10.0	P	11/04/2025	13:00	LB137761
	Cobalt	2.26	+/-15	U	30.0	P	11/04/2025	13:00	LB137761
	Copper	4.60	+/-10	U	20.0	P	11/04/2025	13:00	LB137761
	Iron	31.4	+/-50	J	100	P	11/04/2025	13:00	LB137761
	Lead	2.30	+/-6	U	12.0	P	11/04/2025	13:00	LB137761
	Magnesium	244	+/-1000	U	2000	P	11/04/2025	13:00	LB137761
	Manganese	5.94	+/-10	U	20.0	P	11/04/2025	13:00	LB137761
	Nickel	3.06	+/-20	U	40.0	P	11/04/2025	13:00	LB137761
	Potassium	918	+/-1000	U	2000	P	11/04/2025	13:00	LB137761
	Selenium	9.64	+/-10	U	20.0	P	11/04/2025	13:00	LB137761
	Silver	1.62	+/-5	U	10.0	P	11/04/2025	13:00	LB137761
	Sodium	868	+/-1000	U	2000	P	11/04/2025	13:00	LB137761
	Thallium	4.38	+/-20	U	40.0	P	11/04/2025	13:00	LB137761
	Vanadium	6.26	+/-20	U	40.0	P	11/04/2025	13:00	LB137761
	Zinc	16.7	+/-20	U	40.0	P	11/04/2025	13:00	LB137761
CCB02	Aluminum	11.3	+/-50	U	100	P	11/04/2025	14:03	LB137761
	Antimony	6.76	+/-25	U	50.0	P	11/04/2025	14:03	LB137761
	Arsenic	5.12	+/-10	U	20.0	P	11/04/2025	14:03	LB137761
	Barium	14.6	+/-50	U	100	P	11/04/2025	14:03	LB137761
	Beryllium	0.56	+/-3	U	6.00	P	11/04/2025	14:03	LB137761
	Cadmium	0.50	+/-3	U	6.00	P	11/04/2025	14:03	LB137761
	Calcium	234	+/-1000	U	2000	P	11/04/2025	14:03	LB137761
	Chromium	2.12	+/-5	U	10.0	P	11/04/2025	14:03	LB137761
	Cobalt	2.26	+/-15	U	30.0	P	11/04/2025	14:03	LB137761
	Copper	4.60	+/-10	U	20.0	P	11/04/2025	14:03	LB137761
	Iron	23.4	+/-50	U	100	P	11/04/2025	14:03	LB137761
	Lead	2.30	+/-6	U	12.0	P	11/04/2025	14:03	LB137761
	Magnesium	244	+/-1000	U	2000	P	11/04/2025	14:03	LB137761
	Manganese	5.94	+/-10	U	20.0	P	11/04/2025	14:03	LB137761
	Nickel	3.06	+/-20	U	40.0	P	11/04/2025	14:03	LB137761
	Potassium	918	+/-1000	U	2000	P	11/04/2025	14:03	LB137761
	Selenium	9.64	+/-10	U	20.0	P	11/04/2025	14:03	LB137761

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/04/2025	14:03	LB137761
	Sodium	868	+/-1000	U	2000	P	11/04/2025	14:03	LB137761
	Thallium	4.38	+/-20	U	40.0	P	11/04/2025	14:03	LB137761
	Vanadium	6.26	+/-20	U	40.0	P	11/04/2025	14:03	LB137761
	Zinc	16.7	+/-20	U	40.0	P	11/04/2025	14:03	LB137761
CCB03	Aluminum	16.9	+/-50	J	100	P	11/04/2025	15:10	LB137761
	Antimony	6.76	+/-25	U	50.0	P	11/04/2025	15:10	LB137761
	Arsenic	5.12	+/-10	U	20.0	P	11/04/2025	15:10	LB137761
	Barium	14.6	+/-50	U	100	P	11/04/2025	15:10	LB137761
	Beryllium	0.56	+/-3	U	6.00	P	11/04/2025	15:10	LB137761
	Cadmium	0.50	+/-3	U	6.00	P	11/04/2025	15:10	LB137761
	Calcium	234	+/-1000	U	2000	P	11/04/2025	15:10	LB137761
	Chromium	2.12	+/-5	U	10.0	P	11/04/2025	15:10	LB137761
	Cobalt	2.26	+/-15	U	30.0	P	11/04/2025	15:10	LB137761
	Copper	4.60	+/-10	U	20.0	P	11/04/2025	15:10	LB137761
	Iron	23.4	+/-50	U	100	P	11/04/2025	15:10	LB137761
	Lead	2.30	+/-6	U	12.0	P	11/04/2025	15:10	LB137761
	Magnesium	244	+/-1000	U	2000	P	11/04/2025	15:10	LB137761
	Manganese	5.94	+/-10	U	20.0	P	11/04/2025	15:10	LB137761
	Nickel	3.06	+/-20	U	40.0	P	11/04/2025	15:10	LB137761
	Potassium	918	+/-1000	U	2000	P	11/04/2025	15:10	LB137761
	Selenium	9.64	+/-10	U	20.0	P	11/04/2025	15:10	LB137761
	Silver	1.62	+/-5	U	10.0	P	11/04/2025	15:10	LB137761
	Sodium	868	+/-1000	U	2000	P	11/04/2025	15:10	LB137761
	Thallium	4.38	+/-20	U	40.0	P	11/04/2025	15:10	LB137761
	Vanadium	6.26	+/-20	U	40.0	P	11/04/2025	15:10	LB137761
	Zinc	16.7	+/-20	U	40.0	P	11/04/2025	15:10	LB137761
CCB04	Aluminum	11.3	+/-50	U	100	P	11/04/2025	16:14	LB137761
	Antimony	6.76	+/-25	U	50.0	P	11/04/2025	16:14	LB137761
	Arsenic	5.12	+/-10	U	20.0	P	11/04/2025	16:14	LB137761
	Barium	14.6	+/-50	U	100	P	11/04/2025	16:14	LB137761
	Beryllium	0.56	+/-3	U	6.00	P	11/04/2025	16:14	LB137761
	Cadmium	0.50	+/-3	U	6.00	P	11/04/2025	16:14	LB137761
	Calcium	234	+/-1000	U	2000	P	11/04/2025	16:14	LB137761
	Chromium	2.12	+/-5	U	10.0	P	11/04/2025	16:14	LB137761
	Cobalt	2.26	+/-15	U	30.0	P	11/04/2025	16:14	LB137761
	Copper	4.60	+/-10	U	20.0	P	11/04/2025	16:14	LB137761
	Iron	23.4	+/-50	U	100	P	11/04/2025	16:14	LB137761
	Lead	2.30	+/-6	U	12.0	P	11/04/2025	16:14	LB137761

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/04/2025	16:14	LB137761
	Manganese	5.94	+/-10	U	20.0	P	11/04/2025	16:14	LB137761
	Nickel	3.06	+/-20	U	40.0	P	11/04/2025	16:14	LB137761
	Potassium	918	+/-1000	U	2000	P	11/04/2025	16:14	LB137761
	Selenium	9.64	+/-10	U	20.0	P	11/04/2025	16:14	LB137761
	Silver	1.62	+/-5	U	10.0	P	11/04/2025	16:14	LB137761
	Sodium	868	+/-1000	U	2000	P	11/04/2025	16:14	LB137761
	Thallium	4.38	+/-20	U	40.0	P	11/04/2025	16:14	LB137761
	Vanadium	6.26	+/-20	U	40.0	P	11/04/2025	16:14	LB137761
	Zinc	16.7	+/-20	U	40.0	P	11/04/2025	16:14	LB137761
CCB05	Aluminum	24.8	+/-50	J	100	P	11/04/2025	16:59	LB137761
	Antimony	6.76	+/-25	U	50.0	P	11/04/2025	16:59	LB137761
	Arsenic	5.12	+/-10	U	20.0	P	11/04/2025	16:59	LB137761
	Barium	14.6	+/-50	U	100	P	11/04/2025	16:59	LB137761
	Beryllium	0.56	+/-3	U	6.00	P	11/04/2025	16:59	LB137761
	Cadmium	0.50	+/-3	U	6.00	P	11/04/2025	16:59	LB137761
	Calcium	234	+/-1000	U	2000	P	11/04/2025	16:59	LB137761
	Chromium	2.12	+/-5	U	10.0	P	11/04/2025	16:59	LB137761
	Cobalt	2.26	+/-15	U	30.0	P	11/04/2025	16:59	LB137761
	Copper	4.60	+/-10	U	20.0	P	11/04/2025	16:59	LB137761
	Iron	23.4	+/-50	U	100	P	11/04/2025	16:59	LB137761
	Lead	2.30	+/-6	U	12.0	P	11/04/2025	16:59	LB137761
	Magnesium	244	+/-1000	U	2000	P	11/04/2025	16:59	LB137761
	Manganese	5.94	+/-10	U	20.0	P	11/04/2025	16:59	LB137761
	Nickel	3.06	+/-20	U	40.0	P	11/04/2025	16:59	LB137761
	Potassium	918	+/-1000	U	2000	P	11/04/2025	16:59	LB137761
	Selenium	9.64	+/-10	U	20.0	P	11/04/2025	16:59	LB137761
	Silver	1.62	+/-5	U	10.0	P	11/04/2025	16:59	LB137761
	Sodium	868	+/-1000	U	2000	P	11/04/2025	16:59	LB137761
	Thallium	4.38	+/-20	U	40.0	P	11/04/2025	16:59	LB137761
	Vanadium	6.26	+/-20	U	40.0	P	11/04/2025	16:59	LB137761
	Zinc	16.7	+/-20	U	40.0	P	11/04/2025	16:59	LB137761

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	18.6	+/-50	J	100	P	11/05/2025	11:41	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	11:41	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	11:41	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	11:41	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	11:41	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	11:41	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	11:41	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	11:41	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	11:41	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	11:41	LB137798
	Iron	29.3	+/-50	J	100	P	11/05/2025	11:41	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	11:41	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	11:41	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	11:41	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	11:41	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	11:41	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	11:41	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	11:41	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	11:41	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	11:41	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	11:41	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	11:41	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	11/05/2025	12:33	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	12:33	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	12:33	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	12:33	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	12:33	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	12:33	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	12:33	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	12:33	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	12:33	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	12:33	LB137798
	Iron	31.0	+/-50	J	100	P	11/05/2025	12:33	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	12:33	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	12:33	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	12:33	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	12:33	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	12:33	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	12:33	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	12:33	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	12:33	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	12:33	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	12:33	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	12:33	LB137798
CCB02	Aluminum	25.4	+/-50	J	100	P	11/05/2025	13:25	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	13:25	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	13:25	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	13:25	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	13:25	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	13:25	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	13:25	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	13:25	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	13:25	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	13:25	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	13:25	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	13:25	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	13:25	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	13:25	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	13:25	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	13:25	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	13:25	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	11/05/2025	13:25	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	13:25	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	13:25	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	13:25	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	13:25	LB137798
CCB03	Aluminum	11.3	+/-50	U	100	P	11/05/2025	14:18	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	14:18	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	14:18	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	14:18	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	14:18	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	14:18	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	14:18	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	14:18	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	14:18	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	14:18	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	14:18	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	14:18	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	14:18	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	14:18	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	14:18	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	14:18	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	14:18	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	14:18	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	14:18	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	14:18	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	14:18	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	14:18	LB137798
CCB04	Aluminum	11.3	+/-50	U	100	P	11/05/2025	15:21	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	15:21	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	15:21	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	15:21	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	15:21	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	15:21	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	15:21	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	15:21	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	15:21	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	15:21	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	15:21	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	15:21	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	11/05/2025	15:21	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	15:21	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	15:21	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	15:21	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	15:21	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	15:21	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	15:21	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	15:21	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	15:21	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	15:21	LB137798
CCB05	Aluminum	11.3	+/-50	U	100	P	11/05/2025	16:25	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	16:25	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	16:25	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	16:25	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	16:25	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	16:25	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	16:25	LB137798
	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	16:25	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	16:25	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	16:25	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	16:25	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	16:25	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	16:25	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	16:25	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	16:25	LB137798
	Potassium	918	+/-1000	U	2000	P	11/05/2025	16:25	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	16:25	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	16:25	LB137798
	Sodium	868	+/-1000	U	2000	P	11/05/2025	16:25	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	16:25	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	16:25	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	16:25	LB137798
CCB06	Aluminum	11.3	+/-50	U	100	P	11/05/2025	17:08	LB137798
	Antimony	6.76	+/-25	U	50.0	P	11/05/2025	17:08	LB137798
	Arsenic	5.12	+/-10	U	20.0	P	11/05/2025	17:08	LB137798
	Barium	14.6	+/-50	U	100	P	11/05/2025	17:08	LB137798
	Beryllium	0.56	+/-3	U	6.00	P	11/05/2025	17:08	LB137798
	Cadmium	0.50	+/-3	U	6.00	P	11/05/2025	17:08	LB137798
	Calcium	234	+/-1000	U	2000	P	11/05/2025	17:08	LB137798

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Remington & Vernick Engineers

SDG No.: Q3495

Contract: REMI02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	11/05/2025	17:08	LB137798
	Cobalt	2.26	+/-15	U	30.0	P	11/05/2025	17:08	LB137798
	Copper	4.60	+/-10	U	20.0	P	11/05/2025	17:08	LB137798
	Iron	23.4	+/-50	U	100	P	11/05/2025	17:08	LB137798
	Lead	2.30	+/-6	U	12.0	P	11/05/2025	17:08	LB137798
	Magnesium	244	+/-1000	U	2000	P	11/05/2025	17:08	LB137798
	Manganese	5.94	+/-10	U	20.0	P	11/05/2025	17:08	LB137798
	Nickel	3.06	+/-20	U	40.0	P	11/05/2025	17:08	LB137798
	Potassium	1830	+/-1000	J*	2000	P	11/05/2025	17:08	LB137798
	Selenium	9.64	+/-10	U	20.0	P	11/05/2025	17:08	LB137798
	Silver	1.62	+/-5	U	10.0	P	11/05/2025	17:08	LB137798
	Sodium	3440	+/-1000	*	2000	P	11/05/2025	17:08	LB137798
	Thallium	4.38	+/-20	U	40.0	P	11/05/2025	17:08	LB137798
	Vanadium	6.26	+/-20	U	40.0	P	11/05/2025	17:08	LB137798
	Zinc	16.7	+/-20	U	40.0	P	11/05/2025	17:08	LB137798