

Continuing Calibration Report for SequenceID : FC103125AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC070104.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 8770301.000  | 60.000  | 3.281  | 6.574  | 146171.683 | 144785.198 | -0.958 |
| Aliphatic C12-C16 | 6495222.000  | 40.000  | 6.575  | 9.976  | 162380.550 | 163782.335 | 0.856  |
| Aliphatic C16-C21 | 9605434.000  | 60.000  | 9.977  | 13.346 | 160090.567 | 161945.502 | 1.145  |
| Aliphatic C21-C28 | 11040604.000 | 80.000  | 13.347 | 17.012 | 138007.550 | 147142.336 | 6.208  |
| Aliphatic C28-C40 | 11953969.000 | 120.000 | 17.013 | 21.974 | 99616.408  | 107599.833 | 7.420  |
| Aliphatic EPH     | 47865530.000 | 360.000 | 3.281  | 21.974 | 132959.806 | 137884.728 | 3.572  |