

Continuing Calibration Report for SequenceID : FE103125AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE056611.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 9629117.000  | 60.000  | 3.227  | 6.853  | 160485.283 | 171162.641 | 6.238  |
| Aliphatic C12-C16 | 6897912.000  | 40.000  | 6.854  | 10.303 | 172447.800 | 178754.319 | 3.528  |
| Aliphatic C16-C21 | 10623622.000 | 60.000  | 10.304 | 13.679 | 177060.367 | 186505.496 | 5.064  |
| Aliphatic C21-C28 | 13373897.000 | 80.000  | 13.680 | 17.352 | 167173.713 | 192766.144 | 13.276 |
| Aliphatic C28-C40 | 16435157.000 | 120.000 | 17.353 | 22.305 | 136959.642 | 159314.647 | 14.032 |
| Aliphatic EPH     | 56959705.000 | 360.000 | 3.227  | 22.305 | 158221.403 | 175414.751 | 9.802  |