

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114798.D **Time Analyzed:** 09:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.498	5.392	5.592	549.210	500.000	9.8
Aroclor-1016-2	5.519	5.414	5.614	512.000	500.000	2.4
Aroclor-1016-3	5.581	5.476	5.676	518.270	500.000	3.7
Aroclor-1016-4	5.677	5.571	5.771	544.470	500.000	8.9
Aroclor-1016-5	5.968	5.863	6.063	545.780	500.000	9.2
Aroclor-1260-1	7.084	6.975	7.175	533.290	500.000	6.7
Aroclor-1260-2	7.337	7.229	7.429	537.580	500.000	7.5
Aroclor-1260-3	7.695	7.585	7.785	482.040	500.000	-3.6
Aroclor-1260-4	7.917	7.807	8.007	513.500	500.000	2.7
Aroclor-1260-5	8.224	8.113	8.313	496.810	500.000	-0.6
Decachlorobiphenyl	9.954	9.834	10.034	49.130	50.000	-1.7
Tetrachloro-m-xylene	4.358	4.255	4.455	52.190	50.000	4.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114798.D **Time Analyzed:** 09:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.643	4.843	449.460	500.000	-10.1
Aroclor-1016-2	4.753	4.662	4.862	449.990	500.000	-10.0
Aroclor-1016-3	4.928	4.836	5.036	472.290	500.000	-5.5
Aroclor-1016-4	4.970	4.879	5.079	481.220	500.000	-3.8
Aroclor-1016-5	5.183	5.092	5.292	477.330	500.000	-4.5
Aroclor-1260-1	6.215	6.125	6.325	501.650	500.000	0.3
Aroclor-1260-2	6.403	6.313	6.513	514.710	500.000	2.9
Aroclor-1260-3	6.556	6.466	6.666	481.740	500.000	-3.7
Aroclor-1260-4	7.027	6.939	7.139	548.660	500.000	9.7
Aroclor-1260-5	7.268	7.179	7.379	543.930	500.000	8.8
Decachlorobiphenyl	8.655	8.571	8.771	49.810	50.000	-0.4
Tetrachloro-m-xylene	3.654	3.560	3.760	44.200	50.000	-11.6