

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114813.D **Time Analyzed:** 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.490	5.392	5.592	548.810	500.000	9.8
Aroclor-1016-2	5.512	5.414	5.614	517.150	500.000	3.4
Aroclor-1016-3	5.574	5.476	5.676	521.040	500.000	4.2
Aroclor-1016-4	5.669	5.571	5.771	550.960	500.000	10.2
Aroclor-1016-5	5.961	5.863	6.063	557.470	500.000	11.5
Aroclor-1260-1	7.075	6.975	7.175	536.020	500.000	7.2
Aroclor-1260-2	7.329	7.229	7.429	529.070	500.000	5.8
Aroclor-1260-3	7.687	7.585	7.785	473.980	500.000	-5.2
Aroclor-1260-4	7.908	7.807	8.007	513.080	500.000	2.6
Aroclor-1260-5	8.215	8.113	8.313	485.160	500.000	-3.0
Decachlorobiphenyl	9.938	9.834	10.034	48.450	50.000	-3.1
Tetrachloro-m-xylene	4.351	4.255	4.455	52.720	50.000	5.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114813.D **Time Analyzed:** 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.731	4.643	4.843	436.480	500.000	-12.7
Aroclor-1016-2	4.751	4.662	4.862	465.100	500.000	-7.0
Aroclor-1016-3	4.926	4.836	5.036	483.540	500.000	-3.3
Aroclor-1016-4	4.968	4.879	5.079	492.740	500.000	-1.5
Aroclor-1016-5	5.180	5.092	5.292	493.280	500.000	-1.3
Aroclor-1260-1	6.212	6.125	6.325	516.380	500.000	3.3
Aroclor-1260-2	6.399	6.313	6.513	525.280	500.000	5.1
Aroclor-1260-3	6.552	6.466	6.666	496.340	500.000	-0.7
Aroclor-1260-4	7.023	6.939	7.139	575.900	500.000	15.2
Aroclor-1260-5	7.263	7.179	7.379	552.550	500.000	10.5
Decachlorobiphenyl	8.649	8.571	8.771	50.390	50.000	0.8
Tetrachloro-m-xylene	3.653	3.560	3.760	45.610	50.000	-8.8