

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076256.D **Time Analyzed:** 16:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.799	5.695	5.895	549.220	500.000	9.8
Aroclor-1016-2	5.820	5.717	5.917	531.540	500.000	6.3
Aroclor-1016-3	5.883	5.780	5.980	536.190	500.000	7.2
Aroclor-1016-4	5.980	5.877	6.077	544.310	500.000	8.9
Aroclor-1016-5	6.272	6.169	6.369	534.530	500.000	6.9
Aroclor-1260-1	7.391	7.287	7.487	567.910	500.000	13.6
Aroclor-1260-2	7.644	7.540	7.740	539.900	500.000	8.0
Aroclor-1260-3	8.002	7.899	8.099	557.630	500.000	11.5
Aroclor-1260-4	8.230	8.127	8.327	589.830	500.000	18.0
Aroclor-1260-5	8.556	8.452	8.652	540.380	500.000	8.1
Decachlorobiphenyl	10.426	10.319	10.519	51.770	50.000	3.5
Tetrachloro-m-xylene	4.646	4.542	4.742	54.410	50.000	8.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/31/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076256.D **Time Analyzed:** 16:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.876	4.778	4.978	532.760	500.000	6.6
Aroclor-1016-2	4.934	4.836	5.036	531.330	500.000	6.3
Aroclor-1016-3	5.054	4.954	5.154	503.330	500.000	0.7
Aroclor-1016-4	5.096	4.996	5.196	550.210	500.000	10.0
Aroclor-1016-5	5.310	5.210	5.410	499.130	500.000	-0.2
Aroclor-1260-1	6.337	6.238	6.438	542.710	500.000	8.5
Aroclor-1260-2	6.523	6.425	6.625	546.460	500.000	9.3
Aroclor-1260-3	6.678	6.578	6.778	557.640	500.000	11.5
Aroclor-1260-4	7.144	7.047	7.247	552.050	500.000	10.4
Aroclor-1260-5	7.387	7.287	7.487	561.460	500.000	12.3
Decachlorobiphenyl	8.781	8.682	8.882	45.840	50.000	-8.3
Tetrachloro-m-xylene	3.782	3.681	3.881	56.140	50.000	12.3