

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO100925

Review By	yogesh	Review On	10/9/2025 1:00:25 PM
Supervise By	mohammad	Supervise On	10/11/2025 1:35:39 AM
SubDirectory	PO100925	HP Acquire Method	HP Processing Method PO100925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114311.D	09 Oct 2025 13:37		YP/AJ	Ok
2	I.BLK	I.BLK	PO114312.D	09 Oct 2025 13:55		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO114313.D	09 Oct 2025 14:14		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO114314.D	09 Oct 2025 14:32		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO114315.D	09 Oct 2025 14:50		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO114316.D	09 Oct 2025 15:09		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO114317.D	09 Oct 2025 15:27		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO114318.D	09 Oct 2025 15:45		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO114319.D	09 Oct 2025 16:04		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO114320.D	09 Oct 2025 16:22		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO114321.D	09 Oct 2025 16:40		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO114322.D	09 Oct 2025 16:59		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO114323.D	09 Oct 2025 17:17		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO114324.D	09 Oct 2025 17:35		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO114325.D	09 Oct 2025 17:54		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO114326.D	09 Oct 2025 18:12		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO114327.D	09 Oct 2025 18:30		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO114328.D	09 Oct 2025 18:49		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO100925

Review By	yogesh	Review On	10/9/2025 1:00:25 PM
Supervise By	mohammad	Supervise On	10/11/2025 1:35:39 AM
SubDirectory	PO100925	HP Acquire Method	HP Processing Method PO100925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO114329.D	09 Oct 2025 19:07		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO114330.D	09 Oct 2025 19:26		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO114331.D	09 Oct 2025 19:44		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO114332.D	09 Oct 2025 20:02		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO114333.D	09 Oct 2025 20:21		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO114334.D	09 Oct 2025 20:39		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO114335.D	09 Oct 2025 20:57		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO114336.D	09 Oct 2025 21:16		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO114337.D	09 Oct 2025 21:34		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO114338.D	09 Oct 2025 21:52		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO114339.D	09 Oct 2025 22:11		YP/AJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PO114340.D	09 Oct 2025 22:29		YP/AJ	Ok,M
31	PO100925ICV500	ICVPO100925	PO114341.D	09 Oct 2025 22:47		YP/AJ	Ok
32	AR1242ICV500	ICVPO100925AR1242	PO114342.D	09 Oct 2025 23:24		YP/AJ	Ok
33	AR1248ICV500	ICVPO100925AR1248	PO114343.D	10 Oct 2025 00:01		YP/AJ	Ok
34	AR1254ICV500	ICVPO100925AR1254	PO114344.D	10 Oct 2025 00:38		YP/AJ	Ok
35	AR1268ICV500	ICVPO100925AR1268	PO114345.D	10 Oct 2025 01:14		YP/AJ	Ok
36	DDT ANALOG	DDT ANALOG	PO114346.D	10 Oct 2025 01:51		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO103125

Review By	yogesh	Review On	10/31/2025 2:38:32 PM
Supervise By	rahul	Supervise On	11/3/2025 8:14:34 AM
SubDirectory	PO103125	HP Acquire Method	HP Processing Method PO100925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO114797.D	31 Oct 2025 09:01		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO114798.D	31 Oct 2025 09:45		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO114799.D	31 Oct 2025 10:02		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO114800.D	31 Oct 2025 10:20		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO114801.D	31 Oct 2025 10:38		YP/AJ	Ok
6	I.BLK	I.BLK	PO114802.D	31 Oct 2025 10:56		YP/AJ	Ok
7	Q3489-26	L65-B	PO114803.D	31 Oct 2025 11:14		YP/AJ	Ok,M
8	Q3489-30	STF24-B	PO114804.D	31 Oct 2025 11:33		YP/AJ	Ok
9	Q3489-15DL	L60-ADL	PO114805.D	31 Oct 2025 11:51	AR1260 Hit	YP/AJ	Ok,M
10	Q3489-17DL	L61-ADL	PO114806.D	31 Oct 2025 12:10	AR1260 Hit	YP/AJ	Ok,M
11	PB170339BL	PB170339BL	PO114807.D	31 Oct 2025 13:22		YP/AJ	Ok
12	PB170339BS	PB170339BS	PO114808.D	31 Oct 2025 13:40		YP/AJ	Ok,M
13	Q3495-02	SB-1025-2	PO114809.D	31 Oct 2025 13:58		YP/AJ	Ok,M
14	Q3495-02MS	SB-1025-2MS	PO114810.D	31 Oct 2025 14:17		YP/AJ	Ok,M
15	Q3495-02MSD	SB-1025-2MSD	PO114811.D	31 Oct 2025 14:35		YP/AJ	Ok,M
16	Q3496-01	DRILL-CUTTINGS	PO114812.D	31 Oct 2025 14:52		YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO114813.D	31 Oct 2025 15:59		YP/AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PO114814.D	31 Oct 2025 16:16		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO103125

Review By	yogesh	Review On	10/31/2025 2:38:32 PM
Supervise By	rahul	Supervise On	11/3/2025 8:14:34 AM
SubDirectory	PO103125	HP Acquire Method	HP Processing Method PO100925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO114815.D	31 Oct 2025 16:33		YP/AJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PO114816.D	31 Oct 2025 16:52		YP/AJ	Ok
21	I.BLK	I.BLK	PO114817.D	31 Oct 2025 17:10		YP/AJ	Ok,M
22	Q3497-01	OK-01-103025	PO114818.D	31 Oct 2025 17:29		YP/AJ	Ok
23	Q3499-01	RBR251235	PO114819.D	31 Oct 2025 17:46		YP/AJ	Ok,M
24	Q3504-01	72-11986	PO114820.D	31 Oct 2025 18:04		YP/AJ	Ok,M
25	Q3505-01	10-10-25-SOIL	PO114821.D	31 Oct 2025 18:21	AR1254 Hit	YP/AJ	Ok,M
26	Q3500-05	461-LIQUID	PO114822.D	31 Oct 2025 18:38		YP/AJ	Ok,M
27	Q3505-04	10-10-25-WATER	PO114823.D	31 Oct 2025 18:57	All surrogate fail	YP/AJ	ReRun
28	Q3505-05	10-16-25-WATER	PO114824.D	31 Oct 2025 19:15		YP/AJ	Ok
29	Q3505-06	10-17-25-WATER	PO114825.D	31 Oct 2025 19:34		YP/AJ	Ok
30	Q3496-03	PURGE-WATER	PO114826.D	31 Oct 2025 19:52		YP/AJ	Ok,M
31	Q3502-04	PARTS-CLEANER	PO114827.D	31 Oct 2025 20:09		YP/AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PO114828.D	31 Oct 2025 21:12		YP/AJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PO114829.D	31 Oct 2025 22:07		YP/AJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PO114830.D	31 Oct 2025 22:24		YP/AJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PO114831.D	31 Oct 2025 22:43		YP/AJ	Ok
36	I.BLK	I.BLK	PO114832.D	31 Oct 2025 23:00		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP075886.D	21 Oct 2025 08:33		YP\AJ	Ok
2	I.BLK	I.BLK	PP075887.D	21 Oct 2025 08:49		YP\AJ	Ok,M
3	AR1660ICC1000	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP075889.D	21 Oct 2025 09:22		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP075890.D	21 Oct 2025 09:38		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP075891.D	21 Oct 2025 09:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP075892.D	21 Oct 2025 10:11		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP075893.D	21 Oct 2025 10:27		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP075894.D	21 Oct 2025 10:43		YP\AJ	Ok,M
10	AR1242ICC1000	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59		YP\AJ	Ok,M
11	AR1242ICC750	AR1242ICC750	PP075896.D	21 Oct 2025 11:16		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP075897.D	21 Oct 2025 11:48		YP\AJ	Ok,M
13	AR1242ICC250	AR1242ICC250	PP075898.D	21 Oct 2025 12:04		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP075899.D	21 Oct 2025 12:21		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41		YP\AJ	Ok,M
16	AR1248ICC750	AR1248ICC750	PP075901.D	21 Oct 2025 12:57		YP\AJ	Ok,M
17	AR1248ICC500	AR1248ICC500	PP075902.D	21 Oct 2025 13:13		YP\AJ	Ok,M
18	AR1248ICC250	AR1248ICC250	PP075903.D	21 Oct 2025 13:30		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PP075904.D	21 Oct 2025 13:46		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP075906.D	21 Oct 2025 14:18		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP075907.D	21 Oct 2025 14:35		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP075908.D	21 Oct 2025 14:51		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP075909.D	21 Oct 2025 15:07		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP075910.D	21 Oct 2025 15:40		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP075912.D	21 Oct 2025 16:12		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP075913.D	21 Oct 2025 16:29		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP075914.D	21 Oct 2025 16:45		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP075915.D	21 Oct 2025 17:01		YPIAJ	Ok
31	PP102125ICV500	ICVPP102125	PP075916.D	21 Oct 2025 17:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP102125AR1242	PP075917.D	21 Oct 2025 18:06		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP102125AR1248	PP075918.D	21 Oct 2025 18:39		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP102125AR1254	PP075919.D	22 Oct 2025 08:49		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP102125AR1268	PP075920.D	22 Oct 2025 09:07		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PP075921.D	22 Oct 2025 09:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM
SubDirectory	PP103125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP076241.D	31 Oct 2025 09:26		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP076242.D	31 Oct 2025 09:43	AR1260-04 high in second column	YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP076243.D	31 Oct 2025 09:59		YPIAJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP076244.D	31 Oct 2025 10:15		YPIAJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP076245.D	31 Oct 2025 10:32		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP076246.D	31 Oct 2025 10:48		YPIAJ	Ok
7	Q3482-04	WC-4	PP076247.D	31 Oct 2025 11:09		YPIAJ	Ok,M
8	Q3500-01	459	PP076248.D	31 Oct 2025 13:21	TCMX high in 2nd column ,	YPIAJ	Ok,M
9	Q3501-01	461	PP076249.D	31 Oct 2025 13:38	DCB low in 2nd and high in 1st column , TCMX low in 2nd column	YPIAJ	Not Ok
10	Q3502-01	CATCH-CAN-USED-OI	PP076250.D	31 Oct 2025 13:54		YPIAJ	Ok,M
11	Q3503-01	322	PP076251.D	31 Oct 2025 14:10		YPIAJ	Ok,M
12	Q3503-03	323	PP076252.D	31 Oct 2025 14:26	DCB Low in 2nd column , AR1260 Hit	YPIAJ	Ok,M
13	Q3503-05	324	PP076253.D	31 Oct 2025 14:43		YPIAJ	Ok,M
14	Q3505-02	10-10-25-PLASTIC	PP076254.D	31 Oct 2025 14:59	DCB Low in 2nd column	YPIAJ	Ok,M
15	Q3505-03	10-17-25-RAGS	PP076255.D	31 Oct 2025 15:15		YPIAJ	Ok,M
16	AR1660CCC500	AR1660CCC500	PP076256.D	31 Oct 2025 16:04		YPIAJ	Ok,M
17	AR1242CCC500	AR1242CCC500	PP076257.D	31 Oct 2025 16:20		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM
SubDirectory	PP103125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1248CCC500	AR1248CCC500	PP076258.D	31 Oct 2025 16:37		YPIAJ	Ok,M
19	AR1254CCC500	AR1254CCC500	PP076259.D	31 Oct 2025 16:53		YPIAJ	Ok,M
20	I.BLK	I.BLK	PP076260.D	31 Oct 2025 17:09		YPIAJ	Ok
21	PB170347BL	PB170347BL	PP076261.D	31 Oct 2025 17:26		YPIAJ	Ok
22	PB170347BS	PB170347BS	PP076262.D	31 Oct 2025 17:42		YPIAJ	Ok,M
23	PB170347BSD	PB170347BSD	PP076263.D	31 Oct 2025 17:58		YPIAJ	Ok,M
24	Q3484-01	TRE-25-0118	PP076264.D	31 Oct 2025 18:14	AR1254 hit	YPIAJ	Ok,M
25	Q3492-01	252611	PP076265.D	31 Oct 2025 18:31		YPIAJ	Ok,M
26	Q3495-03	TW-1025-1	PP076266.D	31 Oct 2025 18:47		YPIAJ	Ok,M
27	Q3500-02	462	PP076267.D	31 Oct 2025 19:03	AR1248+1260 Hit	YPIAJ	Ok,M
28	Q3500-03	463	PP076268.D	31 Oct 2025 19:19	AR1248+1260 Hit	YPIAJ	Ok,M
29	Q3500-04	465	PP076269.D	31 Oct 2025 19:36	DCB Low in 2nd column	YPIAJ	Ok,M
30	Q3503-07	320	PP076270.D	31 Oct 2025 19:52	DCB low and TCMX high in 2nd column	YPIAJ	ReRun
31	AR1660CCC500	AR1660CCC500	PP076271.D	31 Oct 2025 20:57		YPIAJ	Ok
32	AR1242CCC500	AR1242CCC500	PP076272.D	31 Oct 2025 21:46	TCMX high in 2nd column	YPIAJ	Ok,M
33	AR1248CCC500	AR1248CCC500	PP076273.D	31 Oct 2025 22:02		YPIAJ	Ok,M
34	AR1254CCC500	AR1254CCC500	PP076274.D	31 Oct 2025 22:18		YPIAJ	Ok,M
35	I.BLK	I.BLK	PP076275.D	31 Oct 2025 22:34		YPIAJ	Ok
36	PB170358BL	PB170358BL	PP076276.D	31 Oct 2025 22:51		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP103125

Review By	rahul	Review On	11/3/2025 8:18:24 AM
Supervise By	yogesh	Supervise On	11/3/2025 8:49:19 AM
SubDirectory	PP103125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB170358BS	PB170358BS	PP076277.D	31 Oct 2025 23:07		YPIAJ	Ok,M
38	Q3437-15	SS-32(0-0.5)	PP076278.D	31 Oct 2025 23:23	AR1254 Hit	YPIAJ	Ok,M
39	Q3450-04	SS-10R-(3.5-4.0)	PP076279.D	31 Oct 2025 23:40		YPIAJ	Ok,M
40	Q3450-04MS	SS-10R-(3.5-4.0)MS	PP076280.D	31 Oct 2025 23:56		YPIAJ	Ok,M
41	Q3450-04MSD	SS-10R-(3.5-4.0)MSD	PP076281.D	01 Nov 2025 00:12		YPIAJ	Ok,M
42	Q3450-05	SS-13R-(3.5-4.0)	PP076282.D	01 Nov 2025 00:28		YPIAJ	Ok
43	Q3450-06	SS-13R-(5.0-5.5)	PP076283.D	01 Nov 2025 00:45		YPIAJ	Ok,M
44	Q3450-23	SS-22-(0.0-0.5)	PP076284.D	01 Nov 2025 01:01	AR1254 Hit , Need 5x dilution	YPIAJ	Dilution
45	Q3450-25	SS-22-(3.5-4.0)	PP076285.D	01 Nov 2025 01:17	AR1254 Hit	YPIAJ	Ok,M
46	AR1660CCC500	AR1660CCC500	PP076286.D	01 Nov 2025 02:22		YPIAJ	Ok,M
47	AR1242CCC500	AR1242CCC500	PP076287.D	01 Nov 2025 03:11	TCMX high in 2nd column	YPIAJ	Ok,M
48	AR1248CCC500	AR1248CCC500	PP076288.D	01 Nov 2025 03:28	TCMX high in 2nd column	YPIAJ	Ok,M
49	AR1254CCC500	AR1254CCC500	PP076289.D	01 Nov 2025 03:44	TCMX high in 2nd column	YPIAJ	Ok,M
50	I.BLK	I.BLK	PP076290.D	01 Nov 2025 04:00		YPIAJ	Ok

M : Manual Integration