



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3495	Analytical Method:	8082A
Client:	Remington & Vernick Engineers	Datafile :	PO114808.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170339BS (Column 1)	AR1016	166.6	183	ug/kg	110				71	120	
	AR1260	166.6	163	ug/kg	98				65	130	
PB170339BS (Column 2)	AR1016	166.6	157	ug/kg	94				71	120	
	AR1260	166.6	170	ug/kg	102				65	130	