



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3495</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Remington & Vernick Engineers</u>	Datafile :	<u>PP076263.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170347BSD (Column 1)	AR1016	5	4.90	ug/L	98	2			77	107	20
	AR1260	5	4.90	ug/L	98	2			66	113	20
PB170347BSD (Column 2)	AR1016	5	4.30	ug/L	86	5			77	107	20
	AR1260	5	5.00	ug/L	100	2			66	113	20