

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3495

Continuing Calib Date: 10/31/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 11:09

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.08	9.06	8.96	9.16	-0.02
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
alpha-BHC	4.00	3.99	3.89	4.09	-0.01
beta-BHC	4.52	4.51	4.41	4.61	-0.01
delta-BHC	4.77	4.76	4.66	4.86	-0.01
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.92	4.82	5.02	-0.01
Aldrin	5.27	5.26	5.16	5.36	-0.01
Heptachlor epoxide	5.70	5.68	5.58	5.78	-0.02
Endosulfan I	6.08	6.07	5.97	6.17	-0.01
Dieldrin	6.35	6.34	6.24	6.44	-0.01
4,4'-DDE	6.20	6.19	6.09	6.29	-0.01
Endrin	6.58	6.57	6.47	6.67	-0.01
Endosulfan II	6.79	6.78	6.68	6.88	-0.01
4,4'-DDD	6.71	6.70	6.60	6.80	-0.01
Endosulfan sulfate	7.16	7.15	7.05	7.25	-0.01
4,4'-DDT	7.03	7.02	6.92	7.12	-0.01
Methoxychlor	7.50	7.49	7.39	7.59	-0.01
Endrin ketone	7.64	7.63	7.53	7.73	-0.01
Endrin aldehyde	6.92	6.91	6.81	7.01	-0.01
alpha-Chlordane	6.03	6.02	5.92	6.12	-0.01
gamma-Chlordane	5.95	5.94	5.84	6.04	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI02

Lab Code: ACE

SDG NO.: Q3495

Continuing Calib Date: 10/31/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 11:09

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.06	7.96	8.16	-0.01
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.26	4.26	4.16	4.36	0.00
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.01
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.87	4.86	4.76	4.96	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.51	5.50	5.40	5.60	-0.01
4,4'-DDE	5.37	5.37	5.27	5.47	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.48	6.47	6.37	6.57	-0.01
4,4'-DDT	6.18	6.17	6.07	6.27	-0.01
Methoxychlor	6.75	6.75	6.65	6.85	0.00
Endrin ketone	6.99	6.98	6.88	7.08	-0.01
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.12	5.12	5.02	5.22	0.00