

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090926.D **Time Analyzed:** 11:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.712	6.600	6.800	48.540	50.000	-2.9
4,4'-DDE	6.202	6.090	6.290	45.040	50.000	-9.9
4,4'-DDT	7.028	6.915	7.115	51.870	50.000	3.7
Aldrin	5.274	5.164	5.364	45.260	50.000	-9.5
alpha-BHC	4.003	3.894	4.094	45.180	50.000	-9.6
alpha-Chlordane	6.032	5.921	6.121	43.400	50.000	-13.2
beta-BHC	4.522	4.412	4.612	43.330	50.000	-13.3
Decachlorobiphenyl	9.081	8.964	9.164	43.690	50.000	-12.6
delta-BHC	4.770	4.660	4.860	45.000	50.000	-10.0
Dieldrin	6.352	6.240	6.440	46.130	50.000	-7.7
Endosulfan I	6.079	5.968	6.168	45.300	50.000	-9.4
Endosulfan II	6.792	6.681	6.881	46.350	50.000	-7.3
Endosulfan sulfate	7.156	7.045	7.245	45.990	50.000	-8.0
Endrin	6.580	6.468	6.668	46.810	50.000	-6.4
Endrin aldehyde	6.922	6.810	7.010	47.350	50.000	-5.3
Endrin ketone	7.637	7.525	7.725	47.160	50.000	-5.7
gamma-BHC (Lindane)	4.334	4.225	4.425	44.400	50.000	-11.2
gamma-Chlordane	5.951	5.840	6.040	45.120	50.000	-9.8
Heptachlor	4.932	4.823	5.023	56.000	50.000	12.0
Heptachlor epoxide	5.695	5.584	5.784	44.950	50.000	-10.1
Methoxychlor	7.501	7.388	7.588	52.280	50.000	4.6
Tetrachloro-m-xylene	3.552	3.445	3.645	42.890	50.000	-14.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090926.D **Time Analyzed:** 11:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.923	5.821	6.021	50.860	50.000	1.7
4,4'-DDE	5.367	5.266	5.466	50.270	50.000	0.5
4,4'-DDT	6.177	6.074	6.274	57.220	50.000	14.4
Aldrin	4.361	4.261	4.461	49.550	50.000	-0.9
alpha-BHC	3.386	3.287	3.487	49.450	50.000	-1.1
alpha-Chlordane	5.183	5.081	5.281	50.040	50.000	0.1
beta-BHC	4.019	3.919	4.119	48.710	50.000	-2.6
Decachlorobiphenyl	8.066	7.962	8.162	56.820	50.000	13.6
delta-BHC	4.255	4.155	4.355	49.160	50.000	-1.7
Dieldrin	5.506	5.404	5.604	50.670	50.000	1.3
Endosulfan I	5.240	5.138	5.338	48.720	50.000	-2.6
Endosulfan II	6.074	5.972	6.172	52.200	50.000	4.4
Endosulfan sulfate	6.477	6.374	6.574	52.480	50.000	5.0
Endrin	5.783	5.681	5.881	51.560	50.000	3.1
Endrin aldehyde	6.253	6.150	6.350	52.780	50.000	5.6
Endrin ketone	6.985	6.883	7.083	58.590	50.000	17.2
gamma-BHC (Lindane)	3.723	3.623	3.823	49.340	50.000	-1.3
gamma-Chlordane	5.118	5.017	5.217	50.120	50.000	0.2
Heptachlor	4.075	3.975	4.175	51.770	50.000	3.5
Heptachlor epoxide	4.865	4.764	4.964	48.690	50.000	-2.6
Methoxychlor	6.748	6.645	6.845	56.010	50.000	12.0
Tetrachloro-m-xylene	2.873	2.775	2.975	48.290	50.000	-3.4