

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090933.D **Time Analyzed:** 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.702	6.600	6.800	46.530	50.000	-6.9
4,4'-DDE	6.192	6.090	6.290	43.680	50.000	-12.6
4,4'-DDT	7.018	6.915	7.115	50.220	50.000	0.4
Aldrin	5.265	5.164	5.364	44.350	50.000	-11.3
alpha-BHC	3.995	3.894	4.094	44.810	50.000	-10.4
alpha-Chlordane	6.022	5.921	6.121	42.650	50.000	-14.7
beta-BHC	4.513	4.412	4.612	43.330	50.000	-13.3
Decachlorobiphenyl	9.070	8.964	9.164	42.870	50.000	-14.3
delta-BHC	4.761	4.660	4.860	43.970	50.000	-12.1
Dieldrin	6.343	6.240	6.440	45.000	50.000	-10.0
Endosulfan I	6.070	5.968	6.168	44.180	50.000	-11.6
Endosulfan II	6.783	6.681	6.881	44.920	50.000	-10.2
Endosulfan sulfate	7.147	7.045	7.245	44.700	50.000	-10.6
Endrin	6.570	6.468	6.668	45.550	50.000	-8.9
Endrin aldehyde	6.912	6.810	7.010	46.010	50.000	-8.0
Endrin ketone	7.628	7.525	7.725	45.830	50.000	-8.3
gamma-BHC (Lindane)	4.326	4.225	4.425	44.280	50.000	-11.4
gamma-Chlordane	5.942	5.840	6.040	43.940	50.000	-12.1
Heptachlor	4.924	4.823	5.023	54.600	50.000	9.2
Heptachlor epoxide	5.686	5.584	5.784	44.780	50.000	-10.4
Methoxychlor	7.491	7.388	7.588	50.720	50.000	1.4
Tetrachloro-m-xylene	3.545	3.445	3.645	43.080	50.000	-13.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090933.D **Time Analyzed:** 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.921	5.821	6.021	50.070	50.000	0.1
4,4'-DDE	5.366	5.266	5.466	49.650	50.000	-0.7
4,4'-DDT	6.174	6.074	6.274	56.020	50.000	12.0
Aldrin	4.360	4.261	4.461	49.190	50.000	-1.6
alpha-BHC	3.386	3.287	3.487	48.830	50.000	-2.3
alpha-Chlordane	5.181	5.081	5.281	49.170	50.000	-1.7
beta-BHC	4.019	3.919	4.119	48.210	50.000	-3.6
Decachlorobiphenyl	8.062	7.962	8.162	56.070	50.000	12.1
delta-BHC	4.255	4.155	4.355	48.750	50.000	-2.5
Dieldrin	5.505	5.404	5.604	50.000	50.000	0.0
Endosulfan I	5.238	5.138	5.338	47.980	50.000	-4.0
Endosulfan II	6.072	5.972	6.172	50.960	50.000	1.9
Endosulfan sulfate	6.474	6.374	6.574	51.370	50.000	2.7
Endrin	5.781	5.681	5.881	51.480	50.000	3.0
Endrin aldehyde	6.251	6.150	6.350	51.790	50.000	3.6
Endrin ketone	6.983	6.883	7.083	56.530	50.000	13.1
gamma-BHC (Lindane)	3.723	3.623	3.823	48.820	50.000	-2.4
gamma-Chlordane	5.117	5.017	5.217	49.240	50.000	-1.5
Heptachlor	4.075	3.975	4.175	51.650	50.000	3.3
Heptachlor epoxide	4.864	4.764	4.964	48.020	50.000	-4.0
Methoxychlor	6.745	6.645	6.845	56.430	50.000	12.9
Tetrachloro-m-xylene	2.874	2.775	2.975	47.690	50.000	-4.6