

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090949.D **Time Analyzed:** 13:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.600	6.800	51.660	50.000	3.3
4,4'-DDE	6.197	6.090	6.290	48.680	50.000	-2.6
4,4'-DDT	7.022	6.915	7.115	48.520	50.000	-3.0
Aldrin	5.270	5.164	5.364	52.900	50.000	5.8
alpha-BHC	4.000	3.894	4.094	55.710	50.000	11.4
alpha-Chlordane	6.027	5.921	6.121	47.950	50.000	-4.1
beta-BHC	4.519	4.412	4.612	51.550	50.000	3.1
Decachlorobiphenyl	9.074	8.964	9.164	43.990	50.000	-12.0
delta-BHC	4.767	4.660	4.860	54.270	50.000	8.5
Dieldrin	6.347	6.240	6.440	50.090	50.000	0.2
Endosulfan I	6.074	5.968	6.168	49.820	50.000	-0.4
Endosulfan II	6.787	6.681	6.881	48.630	50.000	-2.7
Endosulfan sulfate	7.151	7.045	7.245	45.780	50.000	-8.4
Endrin	6.575	6.468	6.668	49.110	50.000	-1.8
Endrin aldehyde	6.917	6.810	7.010	45.580	50.000	-8.8
Endrin ketone	7.631	7.525	7.725	47.380	50.000	-5.2
gamma-BHC (Lindane)	4.332	4.225	4.425	54.710	50.000	9.4
gamma-Chlordane	5.947	5.840	6.040	50.180	50.000	0.4
Heptachlor	4.929	4.823	5.023	66.730	50.000	33.5
Heptachlor epoxide	5.690	5.584	5.784	49.950	50.000	-0.1
Methoxychlor	7.495	7.388	7.588	50.020	50.000	0.0
Tetrachloro-m-xylene	3.551	3.445	3.645	51.620	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090949.D **Time Analyzed:** 13:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.821	6.021	47.850	50.000	-4.3
4,4'-DDE	5.365	5.266	5.466	49.430	50.000	-1.1
4,4'-DDT	6.174	6.074	6.274	48.230	50.000	-3.5
Aldrin	4.359	4.261	4.461	52.620	50.000	5.2
alpha-BHC	3.386	3.287	3.487	53.750	50.000	7.5
alpha-Chlordane	5.181	5.081	5.281	50.060	50.000	0.1
beta-BHC	4.018	3.919	4.119	52.130	50.000	4.3
Decachlorobiphenyl	8.062	7.962	8.162	37.880	50.000	-24.2
delta-BHC	4.255	4.155	4.355	52.950	50.000	5.9
Dieldrin	5.503	5.404	5.604	49.200	50.000	-1.6
Endosulfan I	5.238	5.138	5.338	48.710	50.000	-2.6
Endosulfan II	6.072	5.972	6.172	47.370	50.000	-5.3
Endosulfan sulfate	6.474	6.374	6.574	43.940	50.000	-12.1
Endrin	5.780	5.681	5.881	47.270	50.000	-5.5
Endrin aldehyde	6.250	6.150	6.350	44.960	50.000	-10.1
Endrin ketone	6.983	6.883	7.083	40.460	50.000	-19.1
gamma-BHC (Lindane)	3.722	3.623	3.823	53.140	50.000	6.3
gamma-Chlordane	5.116	5.017	5.217	50.480	50.000	1.0
Heptachlor	4.074	3.975	4.175	55.600	50.000	11.2
Heptachlor epoxide	4.864	4.764	4.964	51.470	50.000	2.9
Methoxychlor	6.745	6.645	6.845	43.330	50.000	-13.3
Tetrachloro-m-xylene	2.874	2.775	2.975	53.000	50.000	6.0