

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090957.D **Time Analyzed:** 16:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.699	6.600	6.800	56.440	50.000	12.9
4,4'-DDE	6.189	6.090	6.290	53.890	50.000	7.8
4,4'-DDT	7.015	6.915	7.115	53.170	50.000	6.3
Aldrin	5.263	5.164	5.364	54.610	50.000	9.2
alpha-BHC	3.993	3.894	4.094	55.760	50.000	11.5
alpha-Chlordane	6.020	5.921	6.121	52.830	50.000	5.7
beta-BHC	4.511	4.412	4.612	53.250	50.000	6.5
Decachlorobiphenyl	9.064	8.964	9.164	48.210	50.000	-3.6
delta-BHC	4.760	4.660	4.860	55.270	50.000	10.5
Dieldrin	6.340	6.240	6.440	55.040	50.000	10.1
Endosulfan I	6.067	5.968	6.168	54.280	50.000	8.6
Endosulfan II	6.780	6.681	6.881	52.890	50.000	5.8
Endosulfan sulfate	7.143	7.045	7.245	51.440	50.000	2.9
Endrin	6.567	6.468	6.668	54.530	50.000	9.1
Endrin aldehyde	6.909	6.810	7.010	51.240	50.000	2.5
Endrin ketone	7.624	7.525	7.725	51.440	50.000	2.9
gamma-BHC (Lindane)	4.324	4.225	4.425	55.080	50.000	10.2
gamma-Chlordane	5.939	5.840	6.040	53.960	50.000	7.9
Heptachlor	4.922	4.823	5.023	66.530	50.000	33.1
Heptachlor epoxide	5.683	5.584	5.784	54.960	50.000	9.9
Methoxychlor	7.487	7.388	7.588	52.810	50.000	5.6
Tetrachloro-m-xylene	3.544	3.445	3.645	52.920	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL04 **Date Analyzed:** 11/04/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090957.D **Time Analyzed:** 16:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.918	5.821	6.021	54.300	50.000	8.6
4,4'-DDE	5.364	5.266	5.466	52.910	50.000	5.8
4,4'-DDT	6.172	6.074	6.274	52.000	50.000	4.0
Aldrin	4.359	4.261	4.461	52.640	50.000	5.3
alpha-BHC	3.385	3.287	3.487	53.410	50.000	6.8
alpha-Chlordane	5.180	5.081	5.281	52.490	50.000	5.0
beta-BHC	4.018	3.919	4.119	52.430	50.000	4.9
Decachlorobiphenyl	8.060	7.962	8.162	46.000	50.000	-8.0
delta-BHC	4.254	4.155	4.355	53.330	50.000	6.7
Dieldrin	5.502	5.404	5.604	53.290	50.000	6.6
Endosulfan I	5.236	5.138	5.338	51.070	50.000	2.1
Endosulfan II	6.070	5.972	6.172	52.170	50.000	4.3
Endosulfan sulfate	6.472	6.374	6.574	50.680	50.000	1.4
Endrin	5.778	5.681	5.881	53.220	50.000	6.4
Endrin aldehyde	6.248	6.150	6.350	50.540	50.000	1.1
Endrin ketone	6.980	6.883	7.083	49.200	50.000	-1.6
gamma-BHC (Lindane)	3.722	3.623	3.823	53.100	50.000	6.2
gamma-Chlordane	5.115	5.017	5.217	52.640	50.000	5.3
Heptachlor	4.073	3.975	4.175	55.570	50.000	11.1
Heptachlor epoxide	4.862	4.764	4.964	51.570	50.000	3.1
Methoxychlor	6.742	6.645	6.845	51.400	50.000	2.8
Tetrachloro-m-xylene	2.874	2.775	2.975	52.660	50.000	5.3