

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/28/2025 10/28/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097701.D **Time Analyzed:** 11:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.665	6.560	6.760	56.430	50.000	12.9
4,4'-DDE	6.156	6.051	6.251	56.430	50.000	12.9
4,4'-DDT	6.981	6.874	7.074	51.580	50.000	3.2
Aldrin	5.230	5.127	5.327	54.640	50.000	9.3
alpha-BHC	3.971	3.870	4.070	55.870	50.000	11.7
alpha-Chlordane	5.985	5.881	6.081	54.860	50.000	9.7
beta-BHC	4.488	4.386	4.586	55.240	50.000	10.5
Decachlorobiphenyl	9.005	8.893	9.093	48.940	50.000	-2.1
delta-BHC	4.734	4.631	4.831	58.620	50.000	17.2
Dieldrin	6.306	6.200	6.400	53.420	50.000	6.8
Endosulfan I	6.033	5.928	6.128	55.820	50.000	11.6
Endosulfan II	6.746	6.641	6.841	51.310	50.000	2.6
Endosulfan sulfate	7.109	7.003	7.203	52.430	50.000	4.9
Endrin	6.531	6.427	6.627	53.140	50.000	6.3
Endrin aldehyde	6.875	6.769	6.969	51.160	50.000	2.3
Endrin ketone	7.589	7.481	7.681	54.080	50.000	8.2
gamma-BHC (Lindane)	4.299	4.198	4.398	55.510	50.000	11.0
gamma-Chlordane	5.904	5.801	6.001	53.610	50.000	7.2
Heptachlor	4.889	4.788	4.988	55.480	50.000	11.0
Heptachlor epoxide	5.651	5.547	5.747	53.350	50.000	6.7
Methoxychlor	7.455	7.347	7.547	48.460	50.000	-3.1
Tetrachloro-m-xylene	3.523	3.424	3.624	55.300	50.000	10.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/28/2025 10/28/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097701.D **Time Analyzed:** 11:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.847	5.745	5.945	52.580	50.000	5.2
4,4'-DDE	5.293	5.193	5.393	51.700	50.000	3.4
4,4'-DDT	6.099	5.997	6.197	49.420	50.000	-1.2
Aldrin	4.286	4.187	4.387	52.480	50.000	5.0
alpha-BHC	3.323	3.226	3.426	53.470	50.000	6.9
alpha-Chlordane	5.104	5.004	5.204	50.820	50.000	1.6
beta-BHC	3.952	3.854	4.054	52.170	50.000	4.3
Decachlorobiphenyl	7.984	7.879	8.079	43.900	50.000	-12.2
delta-BHC	4.185	4.086	4.286	52.530	50.000	5.1
Dieldrin	5.423	5.324	5.524	51.980	50.000	4.0
Endosulfan I	5.159	5.059	5.259	49.340	50.000	-1.3
Endosulfan II	5.992	5.890	6.090	52.040	50.000	4.1
Endosulfan sulfate	6.394	6.291	6.491	49.350	50.000	-1.3
Endrin	5.699	5.599	5.799	53.140	50.000	6.3
Endrin aldehyde	6.170	6.068	6.268	45.880	50.000	-8.2
Endrin ketone	6.899	6.796	6.996	51.460	50.000	2.9
gamma-BHC (Lindane)	3.655	3.558	3.758	53.110	50.000	6.2
gamma-Chlordane	5.040	4.940	5.140	49.840	50.000	-0.3
Heptachlor	4.003	3.905	4.105	52.530	50.000	5.1
Heptachlor epoxide	4.788	4.689	4.889	51.620	50.000	3.2
Methoxychlor	6.672	6.569	6.769	46.410	50.000	-7.2
Tetrachloro-m-xylene	2.818	2.721	2.921	52.500	50.000	5.0