

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/28/2025 10/28/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097715.D **Time Analyzed:** 16:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.664	6.560	6.760	58.700	50.000	17.4
4,4'-DDE	6.155	6.051	6.251	53.560	50.000	7.1
4,4'-DDT	6.978	6.874	7.074	49.700	50.000	-0.6
Aldrin	5.229	5.127	5.327	51.750	50.000	3.5
alpha-BHC	3.972	3.870	4.070	53.980	50.000	8.0
alpha-Chlordane	5.983	5.881	6.081	52.600	50.000	5.2
beta-BHC	4.488	4.386	4.586	53.670	50.000	7.3
Decachlorobiphenyl	9.000	8.893	9.093	47.600	50.000	-4.8
delta-BHC	4.733	4.631	4.831	54.110	50.000	8.2
Dieldrin	6.303	6.200	6.400	51.850	50.000	3.7
Endosulfan I	6.031	5.928	6.128	52.670	50.000	5.3
Endosulfan II	6.744	6.641	6.841	53.360	50.000	6.7
Endosulfan sulfate	7.107	7.003	7.203	50.710	50.000	1.4
Endrin	6.530	6.427	6.627	50.660	50.000	1.3
Endrin aldehyde	6.873	6.769	6.969	52.540	50.000	5.1
Endrin ketone	7.586	7.481	7.681	51.360	50.000	2.7
gamma-BHC (Lindane)	4.299	4.198	4.398	53.600	50.000	7.2
gamma-Chlordane	5.902	5.801	6.001	51.350	50.000	2.7
Heptachlor	4.891	4.788	4.988	52.360	50.000	4.7
Heptachlor epoxide	5.649	5.547	5.747	52.760	50.000	5.5
Methoxychlor	7.451	7.347	7.547	46.070	50.000	-7.9
Tetrachloro-m-xylene	3.525	3.424	3.624	53.320	50.000	6.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI02
Lab Code: ACE **SDG NO.:** Q3495
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/28/2025 10/28/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097715.D **Time Analyzed:** 16:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.848	5.745	5.945	51.730	50.000	3.5
4,4'-DDE	5.296	5.193	5.393	51.180	50.000	2.4
4,4'-DDT	6.100	5.997	6.197	46.760	50.000	-6.5
Aldrin	4.288	4.187	4.387	52.640	50.000	5.3
alpha-BHC	3.327	3.226	3.426	53.410	50.000	6.8
alpha-Chlordane	5.107	5.004	5.204	50.920	50.000	1.8
beta-BHC	3.955	3.854	4.054	51.150	50.000	2.3
Decachlorobiphenyl	7.984	7.879	8.079	42.150	50.000	-15.7
delta-BHC	4.188	4.086	4.286	52.820	50.000	5.6
Dieldrin	5.426	5.324	5.524	53.500	50.000	7.0
Endosulfan I	5.162	5.059	5.259	49.160	50.000	-1.7
Endosulfan II	5.993	5.890	6.090	49.400	50.000	-1.2
Endosulfan sulfate	6.395	6.291	6.491	48.360	50.000	-3.3
Endrin	5.701	5.599	5.799	52.770	50.000	5.5
Endrin aldehyde	6.171	6.068	6.268	45.730	50.000	-8.5
Endrin ketone	6.899	6.796	6.996	50.160	50.000	0.3
gamma-BHC (Lindane)	3.659	3.558	3.758	53.010	50.000	6.0
gamma-Chlordane	5.043	4.940	5.140	50.840	50.000	1.7
Heptachlor	4.006	3.905	4.105	51.690	50.000	3.4
Heptachlor epoxide	4.791	4.689	4.889	51.480	50.000	3.0
Methoxychlor	6.673	6.569	6.769	45.090	50.000	-9.8
Tetrachloro-m-xylene	2.822	2.721	2.921	51.810	50.000	3.6