

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102825\
 Data File : PL097639.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2025 14:57
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 16:40:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102825.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 28 16:38:04 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.523	2.821	397.0E6	576.3E6	99.892	99.443
2)	SA Decachlor...	8.994	7.978	275.3E6	484.7E6	98.173	98.119
Target Compounds							
2)	A alpha-BHC	3.970	3.326	692.9E6	982.0E6	101.783	100.729
3)	MA gamma-BHC...	4.298	3.658	644.0E6	902.6E6	101.445	100.404
4)	MA Heptachlor	4.888	4.005	631.4E6	867.6E6	100.517	99.104
5)	MB Aldrin	5.227	4.287	605.8E6	843.5E6	100.940	99.768
6)	B beta-BHC	4.486	3.954	248.3E6	370.0E6	99.286	98.784
7)	B delta-BHC	4.731	4.186	597.2E6	900.7E6	101.667	100.580
8)	B Heptachlo...	5.647	4.789	535.0E6	744.4E6	99.545	98.693
9)	A Endosulfan I	6.028	5.160	492.3E6	706.4E6	100.497	98.424
10)	B gamma-Chl...	5.901	5.041	541.9E6	826.6E6	100.565	98.859
11)	B alpha-Chl...	5.980	5.105	531.9E6	795.5E6	100.201	99.006
12)	B 4,4'-DDE	6.151	5.293	466.3E6	760.6E6	102.432	100.110
13)	MA Dieldrin	6.300	5.424	528.4E6	784.9E6	100.458	99.464
14)	MA Endrin	6.526	5.698	445.4E6	669.6E6	102.065	98.974
15)	B Endosulfa...	6.740	5.990	432.7E6	665.5E6	99.250	98.774
16)	A 4,4'-DDD	6.660	5.845	356.8E6	627.6E6	100.683	99.775
17)	MA 4,4'-DDT	6.974	6.098	400.2E6	682.0E6	100.618	100.247
18)	B Endrin al...	6.869	6.168	310.1E6	497.5E6	99.061	97.442
19)	B Endosulfa...	7.102	6.392	375.9E6	633.0E6	99.762	98.377
20)	A Methoxychlor	7.447	6.669	191.6E6	341.4E6	98.564	97.527
21)	B Endrin ke...	7.582	6.896	405.9E6	694.1E6	99.604	98.094
22)	Mirex	8.059	7.085	278.6E6	491.3E6	97.296	97.268

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL102825\
Data File : PL097639.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2025 14:57
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 16:40:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL102825.M
Quant Title : GC Extractables
QLast Update : Tue Oct 28 16:38:04 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

