

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090646.D	17 Oct 2025 10:40	AR\AJ	Ok
2	I.BLK	PD090647.D	17 Oct 2025 10:53	AR\AJ	Ok
3	PEM	PD090648.D	17 Oct 2025 11:07	AR\AJ	Ok,M
4	RESCHK	PD090649.D	17 Oct 2025 11:20	AR\AJ	Ok,M
5	PSTDICC100	PD090650.D	17 Oct 2025 11:34	AR\AJ	Ok
6	PSTDICC075	PD090651.D	17 Oct 2025 11:48	AR\AJ	Ok
7	PSTDICC050	PD090652.D	17 Oct 2025 12:01	AR\AJ	Ok
8	PSTDICC025	PD090653.D	17 Oct 2025 12:15	AR\AJ	Ok
9	PSTDICC005	PD090654.D	17 Oct 2025 12:28	AR\AJ	Ok,M
10	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42	AR\AJ	Ok
11	PCHLORICC750	PD090656.D	17 Oct 2025 12:56	AR\AJ	Ok
12	PCHLORICC500	PD090657.D	17 Oct 2025 13:09	AR\AJ	Ok
13	PCHLORICC250	PD090658.D	17 Oct 2025 13:23	AR\AJ	Ok
14	PCHLORICC050	PD090659.D	17 Oct 2025 13:36	AR\AJ	Ok,M
15	PTOXICC1000	PD090660.D	17 Oct 2025 13:50	AR\AJ	Ok
16	PTOXICC750	PD090661.D	17 Oct 2025 14:03	AR\AJ	Ok
17	PTOXICC500	PD090662.D	17 Oct 2025 14:17	AR\AJ	Ok
18	PTOXICC250	PD090663.D	17 Oct 2025 14:30	AR\AJ	Ok
19	PTOXICC100	PD090664.D	17 Oct 2025 16:18	AR\AJ	Ok
20	PSTDICV050	PD090665.D	17 Oct 2025 16:46	AR\AJ	Ok
21	PCHLORICV500	PD090666.D	17 Oct 2025 17:50	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD090667.D	17 Oct 2025 18:42	AR\AJ	Ok,M
23	I.BLK	PD090668.D	17 Oct 2025 19:50	AR\AJ	Ok
24	PEM	PD090669.D	17 Oct 2025 20:04	AR\AJ	Ok,M
25	PSTDCCC050	PD090670.D	17 Oct 2025 20:18	AR\AJ	Ok
26	PB170142BL	PD090671.D	17 Oct 2025 20:31	AR\AJ	Ok
27	PB170142BS	PD090672.D	17 Oct 2025 20:45	AR\AJ	Ok
28	Q3366-01	PD090673.D	17 Oct 2025 20:59	AR\AJ	Ok,M
29	Q3366-03	PD090674.D	17 Oct 2025 21:12	AR\AJ	Ok,M
30	Q3367-01	PD090675.D	17 Oct 2025 21:26	AR\AJ	Ok,M
31	Q3369-01	PD090676.D	17 Oct 2025 21:40	AR\AJ	Ok,M
32	Q3369-05	PD090677.D	17 Oct 2025 21:53	AR\AJ	Ok,M
33	Q3363-01	PD090678.D	17 Oct 2025 22:48	AR\AJ	Ok,M
34	Q3363-02	PD090679.D	17 Oct 2025 23:02	AR\AJ	Ok,M
35	Q3363-03	PD090680.D	17 Oct 2025 23:15	AR\AJ	Ok,M
36	Q3363-04	PD090681.D	17 Oct 2025 23:29	AR\AJ	Ok,M
37	Q3363-04MS	PD090682.D	17 Oct 2025 23:43	AR\AJ	Ok,M
38	Q3363-04MSD	PD090683.D	17 Oct 2025 23:56	AR\AJ	Ok,M
39	Q3363-05	PD090684.D	18 Oct 2025 00:10	AR\AJ	Ok,M
40	Q3343-04MS	PD090685.D	18 Oct 2025 00:24	AR\AJ	Ok,M
41	Q3343-04MSD	PD090686.D	18 Oct 2025 00:37	AR\AJ	Ok
42	I.BLK	PD090687.D	18 Oct 2025 00:51	AR\AJ	Ok
43	PSTDCCC050	PD090688.D	18 Oct 2025 02:13	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD103125

Review By	rahul	Review On	11/3/2025 8:19:14 AM
Supervise By	yogesh	Supervise On	11/3/2025 8:49:55 AM
SubDirectory	PD103125	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090923.D	31 Oct 2025 09:53	AR\AJ	Ok
2	I.BLK	PD090924.D	31 Oct 2025 10:06	AR\AJ	Ok
3	PEM	PD090925.D	31 Oct 2025 10:20	AR\AJ	Ok,M
4	PSTDCCC050	PD090926.D	31 Oct 2025 11:09	AR\AJ	Ok,M
5	PB170348BL	PD090927.D	31 Oct 2025 14:37	AR\AJ	Ok
6	PB170348BS	PD090928.D	31 Oct 2025 14:51	AR\AJ	Not Ok
7	PB170348BSD	PD090929.D	31 Oct 2025 15:07	AR\AJ	Not Ok
8	Q3477-03	PD090930.D	31 Oct 2025 15:21	AR\AJ	Ok,M
9	Q3495-03	PD090931.D	31 Oct 2025 15:34	AR\AJ	Ok,M
10	I.BLK	PD090932.D	31 Oct 2025 15:48	AR\AJ	Ok
11	PSTDCCC050	PD090933.D	31 Oct 2025 16:01	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD110425

Review By	rahul	Review On	11/5/2025 8:30:12 AM
Supervise By	yogesh	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PD110425	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090934.D	04 Nov 2025 09:04	AR\AJ	Ok
2	I.BLK	PD090935.D	04 Nov 2025 09:17	AR\AJ	Ok
3	PEM	PD090936.D	04 Nov 2025 09:31	AR\AJ	Ok
4	PSTDCCC050	PD090937.D	04 Nov 2025 09:45	AR\AJ	Ok,M
5	Q3515-01	PD090938.D	04 Nov 2025 10:17	AR\AJ	Ok,M
6	Q3516-01	PD090939.D	04 Nov 2025 10:31	AR\AJ	Ok,M
7	Q3519-01	PD090940.D	04 Nov 2025 10:44	AR\AJ	Ok,M
8	Q3519-01MS	PD090941.D	04 Nov 2025 10:58	AR\AJ	Ok,M
9	Q3519-01MSD	PD090942.D	04 Nov 2025 11:12	AR\AJ	Ok,M
10	Q3520-01	PD090943.D	04 Nov 2025 11:25	AR\AJ	Ok,M
11	Q3368-06	PD090944.D	04 Nov 2025 11:39	AR\AJ	Not Ok
12	Q3368-06MS	PD090945.D	04 Nov 2025 11:53	AR\AJ	Not Ok
13	Q3368-06MSD	PD090946.D	04 Nov 2025 12:06	AR\AJ	Not Ok
14	Q3368-07	PD090947.D	04 Nov 2025 12:22	AR\AJ	Not Ok
15	I.BLK	PD090948.D	04 Nov 2025 12:35	AR\AJ	Ok
16	PSTDCCC050	PD090949.D	04 Nov 2025 13:29	AR\AJ	Ok
17	PB170348BSD	PD090950.D	04 Nov 2025 13:52	AR\AJ	Ok
18	PB170371BL	PD090951.D	04 Nov 2025 14:09	AR\AJ	Ok
19	PB170348BS	PD090952.D	04 Nov 2025 14:22	AR\AJ	Ok,M
20	PB170387BS	PD090953.D	04 Nov 2025 15:41	AR\AJ	Ok,M
21	PB170387BL	PD090954.D	04 Nov 2025 16:03	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD110425

Review By	rahul	Review On	11/5/2025 8:30:12 AM
Supervise By	yogesh	Supervise On	11/5/2025 8:40:26 AM
SubDirectory	PD110425	HP Acquire Method	HP Processing Method PD101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB170387TB	PD090955.D	04 Nov 2025 16:17	AR\AJ	Ok
23	I.BLK	PD090956.D	04 Nov 2025 16:31	AR\AJ	Ok
24	PSTDCCC050	PD090957.D	04 Nov 2025 16:44	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL102825

Review By	Abdul	Review On	10/29/2025 8:12:28 AM
Supervise By	Sohil	Supervise On	10/29/2025 4:12:00 PM
SubDirectory	PL102825	HP Acquire Method	HP Processing Method pl102825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097635.D	28 Oct 2025 14:03	AR\AJ	Ok
2	I.BLK	PL097636.D	28 Oct 2025 14:17	AR\AJ	Ok
3	PEM	PL097637.D	28 Oct 2025 14:30	AR\AJ	Ok,M
4	RESCHK	PL097638.D	28 Oct 2025 14:44	AR\AJ	Ok,M
5	PSTDICC100	PL097639.D	28 Oct 2025 14:57	AR\AJ	Ok
6	PSTDICC075	PL097640.D	28 Oct 2025 15:11	AR\AJ	Ok
7	PSTDICC050	PL097641.D	28 Oct 2025 15:25	AR\AJ	Ok
8	PSTDICC025	PL097642.D	28 Oct 2025 15:38	AR\AJ	Ok,M
9	PSTDICC005	PL097643.D	28 Oct 2025 15:52	AR\AJ	Ok,M
10	PCHLORICC1000	PL097644.D	28 Oct 2025 16:05	AR\AJ	Ok
11	PCHLORICC750	PL097645.D	28 Oct 2025 16:19	AR\AJ	Ok
12	PCHLORICC500	PL097646.D	28 Oct 2025 16:32	AR\AJ	Ok
13	PCHLORICC250	PL097647.D	28 Oct 2025 16:46	AR\AJ	Ok
14	PCHLORICC050	PL097648.D	28 Oct 2025 17:00	AR\AJ	Ok,M
15	PTOXICC1000	PL097649.D	28 Oct 2025 17:13	AR\AJ	Ok
16	PTOXICC750	PL097650.D	28 Oct 2025 17:27	AR\AJ	Ok
17	PTOXICC500	PL097651.D	28 Oct 2025 17:40	AR\AJ	Ok
18	PTOXICC250	PL097652.D	28 Oct 2025 17:54	AR\AJ	Ok
19	PTOXICC100	PL097653.D	28 Oct 2025 18:08	AR\AJ	Ok,M
20	PSTDICV050	PL097654.D	28 Oct 2025 18:21	AR\AJ	Ok
21	PCHLORICV500	PL097655.D	28 Oct 2025 18:35	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL102825

Review By	Abdul	Review On	10/29/2025 8:12:28 AM
Supervise By	Sohil	Supervise On	10/29/2025 4:12:00 PM
SubDirectory	PL102825	HP Acquire Method	HP Processing Method pl102825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL097656.D	28 Oct 2025 18:48	AR\AJ	Ok
23	I.BLK	PL097657.D	28 Oct 2025 19:02	AR\AJ	Ok
24	PEM	PL097658.D	28 Oct 2025 19:16	AR\AJ	Ok,M
25	PSTDCCC050	PL097659.D	28 Oct 2025 19:29	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL103125

Review By	rahul	Review On	10/31/2025 2:27:02 PM
Supervise By	yogesh	Supervise On	11/3/2025 8:43:32 AM
SubDirectory	PL103125	HP Acquire Method	HP Processing Method pl102825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL097698.D	31 Oct 2025 09:28	AR\AJ	Ok
2	I.BLK	PL097699.D	31 Oct 2025 09:42	AR\AJ	Ok
3	PEM	PL097700.D	31 Oct 2025 09:56	AR\AJ	Ok,M
4	PSTDCCC050	PL097701.D	31 Oct 2025 11:02	AR\AJ	Ok,M
5	Q3488-02	PL097702.D	31 Oct 2025 12:03	AR\AJ	Not Ok
6	Q3490-01	PL097703.D	31 Oct 2025 12:17	AR\AJ	Ok,M
7	PB170340BL	PL097704.D	31 Oct 2025 13:06	AR\AJ	Ok
8	PB170340BS	PL097705.D	31 Oct 2025 13:19	AR\AJ	Ok
9	Q3495-02	PL097706.D	31 Oct 2025 13:33	AR\AJ	Ok,M
10	Q3495-02MS	PL097707.D	31 Oct 2025 13:47	AR\AJ	Ok,M
11	Q3495-02MSD	PL097708.D	31 Oct 2025 14:00	AR\AJ	Ok,M
12	Q3496-01	PL097709.D	31 Oct 2025 14:14	AR\AJ	Ok,M
13	Q3497-01	PL097710.D	31 Oct 2025 14:27	AR\AJ	Ok,M
14	Q3500-01	PL097711.D	31 Oct 2025 14:41	AR\AJ	Ok,M
15	Q3501-01	PL097712.D	31 Oct 2025 14:55	AR\AJ	Ok,M
16	Q3502-01	PL097713.D	31 Oct 2025 15:08	AR\AJ	Ok,M
17	I.BLK	PL097714.D	31 Oct 2025 15:49	AR\AJ	Ok
18	PSTDCCC050	PL097715.D	31 Oct 2025 16:03	AR\AJ	Ok,M

M : Manual Integration