

Surrogate Summary

SDG No.: **Q3495**

Client: **Remington & Vernick Engineers**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		57	171
		Tetrachloro-m-xyl	1	20	19.4	97		61	148
		Decachlorobiphen	2	20	20.3	102		57	171
		Tetrachloro-m-xyl	2	20	20.0	100		61	148
I.BLK-PD090924.D	PIBLK-PD090924.D	Decachlorobiphen	1	20	22.3	111		57	171
		Tetrachloro-m-xyl	1	20	20.4	102		61	148
		Decachlorobiphen	2	20	28.7	143		57	171
		Tetrachloro-m-xyl	2	20	23.8	119		61	148
PB170348BL	PB170348BL	Decachlorobiphen	1	20	21.9	110		57	171
		Tetrachloro-m-xyl	1	20	18.8	94		61	148
		Decachlorobiphen	2	20	27.2	136		57	171
		Tetrachloro-m-xyl	2	20	22.4	112		61	148
Q3495-03	TW-1025-1	Decachlorobiphen	1	20	13.4	67		57	171
		Tetrachloro-m-xyl	1	20	16.6	83		61	148
		Decachlorobiphen	2	20	16.8	84		57	171
		Tetrachloro-m-xyl	2	20	19.9	100		61	148
I.BLK-PD090932.D	PIBLK-PD090932.D	Decachlorobiphen	1	20	21.9	110		57	171
		Tetrachloro-m-xyl	1	20	20.3	102		61	148
		Decachlorobiphen	2	20	28.1	140		57	171
		Tetrachloro-m-xyl	2	20	23.1	116		61	148
I.BLK-PD090948.D	PIBLK-PD090948.D	Decachlorobiphen	1	20	17.7	89		57	171
		Tetrachloro-m-xyl	1	20	19.2	96		61	148
		Decachlorobiphen	2	20	12.9	65		57	171
		Tetrachloro-m-xyl	2	20	19.9	99		61	148
PB170348BSD	PB170348BSD	Decachlorobiphen	1	20	25.0	125		57	171
		Tetrachloro-m-xyl	1	20	26.8	134		61	148
		Decachlorobiphen	2	20	21.3	106		57	171
		Tetrachloro-m-xyl	2	20	28.5	142		61	148
PB170348BS	PB170348BS	Decachlorobiphen	1	20	24.8	124		57	171
		Tetrachloro-m-xyl	1	20	27.6	138		61	148
		Decachlorobiphen	2	20	22.9	115		57	171
		Tetrachloro-m-xyl	2	20	28.8	144		61	148
I.BLK-PD090956.D	PIBLK-PD090956.D	Decachlorobiphen	1	20	19.2	96		57	171
		Tetrachloro-m-xyl	1	20	19.6	98		61	148
		Decachlorobiphen	2	20	17.8	89		57	171
		Tetrachloro-m-xyl	2	20	20.0	100		61	148
I.BLK-PL097636.D	PIBLK-PL097636.D	Decachlorobiphen	1	20	20.2	101		57	171
		Tetrachloro-m-xyl	1	20	18.0	90		61	148
		Decachlorobiphen	2	20	20.1	101		57	171
		Tetrachloro-m-xyl	2	20	18.7	94		61	148
I.BLK-PL097699.D	PIBLK-PL097699.D	Decachlorobiphen	1	20	22.7	113		57	171

Surrogate Summary

SDG No.: **Q3495**

Client: **Remington & Vernick Engineers**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL097699.D	PIBLK-PL097699.D	Tetrachloro-m-xyl	1	20	23.6	118		61	148
		Decachlorobiphen	2	20	20.1	100		57	171
PB170340BL	PB170340BL	Tetrachloro-m-xyl	2	20	22.9	115		61	148
		Decachlorobiphen	1	20	18.9	94		20	144
		Tetrachloro-m-xyl	1	20	19.4	97		19	148
		Decachlorobiphen	2	20	15.9	80		20	144
		Tetrachloro-m-xyl	2	20	20.0	100		19	148
		Decachlorobiphen	1	20	23.4	117		20	144
PB170340BS	PB170340BS	Tetrachloro-m-xyl	1	20	23.7	118		19	148
		Decachlorobiphen	2	20	19.8	99		20	144
		Tetrachloro-m-xyl	2	20	23.1	116		19	148
		Decachlorobiphen	1	20	16.1	81		20	144
Q3495-02	SB-1025-2	Tetrachloro-m-xyl	1	20	19.7	99		19	148
		Decachlorobiphen	2	20	14.3	72		20	144
		Tetrachloro-m-xyl	2	20	18.9	94		19	148
		Decachlorobiphen	1	20	14.8	74		20	144
Q3495-02MS	SB-1025-2MS	Tetrachloro-m-xyl	1	20	16.8	84		19	148
		Decachlorobiphen	2	20	12.8	64		20	144
		Tetrachloro-m-xyl	2	20	16.7	83		19	148
		Decachlorobiphen	1	20	14.9	74		20	144
Q3495-02MSD	SB-1025-2MSD	Tetrachloro-m-xyl	1	20	17.0	85		19	148
		Decachlorobiphen	2	20	13.4	67		20	144
		Tetrachloro-m-xyl	2	20	17.2	86		19	148
		Decachlorobiphen	1	20	23.1	115		57	171
I.BLK-PL097714.D	PIBLK-PL097714.D	Tetrachloro-m-xyl	1	20	23.0	115		61	148
		Decachlorobiphen	2	20	20.5	103		57	171
		Tetrachloro-m-xyl	2	20	23.6	118		61	148
		Decachlorobiphen	1	20	23.1	115		57	171