

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3495	Analytical Method:	8081B
Client:	Remington & Vernick Engineers	Datafile :	PD090950.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170348BSD (Column 1)	alpha-BHC	0.5	0.52	ug/L	104	0			85	130	20
	beta-BHC	0.5	0.49	ug/L	97	2			83	126	20
	delta-BHC	0.5	0.51	ug/L	103	0			69	141	20
	gamma-BHC (Lindane)	0.5	0.51	ug/L	103	0			82	129	20
	Heptachlor	0.5	0.63	ug/L	127	1			79	127	20
	Aldrin	0.5	0.51	ug/L	102	1			79	126	20
	Heptachlor epoxide	0.5	0.49	ug/L	97	4			81	124	20
	Endosulfan I	0.5	0.48	ug/L	97	1			85	122	20
	Dieldrin	0.5	0.49	ug/L	98	0			83	125	20
	4,4'-DDE	0.5	0.48	ug/L	97	1			80	127	20
	Endrin	0.5	0.47	ug/L	95	2			81	128	20
	Endosulfan II	0.5	0.48	ug/L	95	2			82	123	20
	4,4'-DDD	0.5	0.49	ug/L	98	1			77	131	20
	Endosulfan sulfate	0.5	0.45	ug/L	91	1			76	129	20
	4,4'-DDT	0.5	0.48	ug/L	95	1			80	133	20
	Methoxychlor	0.5	0.47	ug/L	94	2			78	108	20
	Endrin ketone	0.5	0.46	ug/L	91	1			80	131	20
	Endrin aldehyde	0.5	0.44	ug/L	89	1			82	127	20
	alpha-Chlordane	0.5	0.47	ug/L	93	4			82	125	20
	gamma-Chlordane	0.5	0.48	ug/L	97	3			82	125	20
	alpha-BHC	0.5	0.51	ug/L	102	1			85	130	20
	beta-BHC	0.5	0.50	ug/L	99	1			83	126	20
	delta-BHC	0.5	0.50	ug/L	101	1			69	141	20
	gamma-BHC (Lindane)	0.5	0.51	ug/L	101	1			82	129	20
	Heptachlor	0.5	0.53	ug/L	106	1			79	127	20
	Aldrin	0.5	0.50	ug/L	101	0			79	126	20
PB170348BSD (Column 2)	Heptachlor epoxide	0.5	0.50	ug/L	100	0			81	124	20
	Endosulfan I	0.5	0.49	ug/L	98	3			85	122	20
	Dieldrin	0.5	0.48	ug/L	96	3			83	125	20
	4,4'-DDE	0.5	0.48	ug/L	96	3			80	127	20
	Endrin	0.5	0.46	ug/L	92	4			81	128	20
	Endosulfan II	0.5	0.45	ug/L	91	3			82	123	20
	4,4'-DDD	0.5	0.47	ug/L	94	4			77	131	20
	Endosulfan sulfate	0.5	0.44	ug/L	88	4			76	129	20
	4,4'-DDT	0.5	0.46	ug/L	92	2			80	133	20
	Methoxychlor	0.5	0.43	ug/L	87	3			78	108	20
	Endrin ketone	0.5	0.41	ug/L	82	7			80	131	20
	Endrin aldehyde	0.5	0.45	ug/L	89	1			82	127	20
	alpha-Chlordane	0.5	0.49	ug/L	98	2			82	125	20
	gamma-Chlordane	0.5	0.49	ug/L	99	2			82	125	20