

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3495	Analytical Method:	8081B
Client:	Remington & Vernick Engineers	Datafile :	PD090952.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170348BS (Column 1)	alpha-BHC	0.5	0.52	ug/L	104				85	130	
	beta-BHC	0.5	0.50	ug/L	99				83	126	
	delta-BHC	0.5	0.51	ug/L	103				69	141	
	gamma-BHC (Lindane)	0.5	0.51	ug/L	103				82	129	
	Heptachlor	0.5	0.63	ug/L	126				79	127	
	Aldrin	0.5	0.51	ug/L	103				79	126	
	Heptachlor epoxide	0.5	0.50	ug/L	101				81	124	
	Endosulfan I	0.5	0.49	ug/L	98				85	122	
	Dieldrin	0.5	0.49	ug/L	98				83	125	
	4,4'-DDE	0.5	0.48	ug/L	96				80	127	
	Endrin	0.5	0.48	ug/L	97				81	128	
	Endosulfan II	0.5	0.48	ug/L	97				82	123	
	4,4'-DDD	0.5	0.50	ug/L	99				77	131	
	Endosulfan sulfate	0.5	0.46	ug/L	92				76	129	
	4,4'-DDT	0.5	0.48	ug/L	96				80	133	
	Methoxychlor	0.5	0.48	ug/L	96				78	108	
	Endrin ketone	0.5	0.46	ug/L	92				80	131	
	Endrin aldehyde	0.5	0.45	ug/L	90				82	127	
	alpha-Chlordane	0.5	0.49	ug/L	97				82	125	
	gamma-Chlordane	0.5	0.50	ug/L	100				82	125	
PB170348BS (Column 2)	alpha-BHC	0.5	0.51	ug/L	103				85	130	
	beta-BHC	0.5	0.50	ug/L	100				83	126	
	delta-BHC	0.5	0.51	ug/L	102				69	141	
	gamma-BHC (Lindane)	0.5	0.51	ug/L	102				82	129	
	Heptachlor	0.5	0.54	ug/L	107				79	127	
	Aldrin	0.5	0.51	ug/L	101				79	126	
	Heptachlor epoxide	0.5	0.50	ug/L	100				81	124	
	Endosulfan I	0.5	0.50	ug/L	101				85	122	
	Dieldrin	0.5	0.49	ug/L	99				83	125	
	4,4'-DDE	0.5	0.50	ug/L	99				80	127	
	Endrin	0.5	0.48	ug/L	96				81	128	
	Endosulfan II	0.5	0.47	ug/L	94				82	123	
	4,4'-DDD	0.5	0.49	ug/L	98				77	131	
	Endosulfan sulfate	0.5	0.46	ug/L	92				76	129	
	4,4'-DDT	0.5	0.47	ug/L	94				80	133	
	Methoxychlor	0.5	0.45	ug/L	90				78	108	
	Endrin ketone	0.5	0.44	ug/L	88				80	131	
	Endrin aldehyde	0.5	0.45	ug/L	90				82	127	
	alpha-Chlordane	0.5	0.50	ug/L	100				82	125	
	gamma-Chlordane	0.5	0.50	ug/L	101				82	125	