

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG102425

Review By	rahul	Review On	10/27/2025 10:16:00 AM
Supervise By	mohammad	Supervise On	11/5/2025 12:11:16 AM
SubDirectory	BG102425	HP Acquire Method	BNA_G
		HP Processing Method	BG102425
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13179,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BG064568.D	24 Oct 2025 8:20	RC/JU	Ok
2	SSTDICC2.5	BG064569.D	24 Oct 2025 9:00	RC/JU	Ok
3	SSTDICC005	BG064570.D	24 Oct 2025 9:41	RC/JU	Ok,M
4	SSTDICC010	BG064571.D	24 Oct 2025 11:02	RC/JU	Ok,M
5	SSTDICC020	BG064572.D	24 Oct 2025 12:23	RC/JU	Ok,M
6	SSTDICCC040	BG064573.D	24 Oct 2025 13:03	RC/JU	Ok,M
7	SSTDICC060	BG064574.D	24 Oct 2025 13:44	RC/JU	Ok,M
8	SSTDICC080	BG064575.D	24 Oct 2025 14:24	RC/JU	Ok,M
9	SSTDICC050	BG064576.D	24 Oct 2025 15:32	RC/JU	Ok,M
10	SSTDICV040	BG064577.D	24 Oct 2025 16:47	RC/JU	Ok
11	PB170222BL	BG064578.D	24 Oct 2025 17:28	RC/JU	Not Ok
12	Q3405-02	BG064579.D	24 Oct 2025 18:08	RC/JU	Ok
13	Q3405-03	BG064580.D	24 Oct 2025 18:49	RC/JU	Ok
14	Q3405-04	BG064581.D	24 Oct 2025 19:30	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG103125

Review By	rahul	Review On	11/3/2025 3:20:11 PM
Supervise By	mohammad	Supervise On	11/5/2025 12:11:48 AM
SubDirectory	BG103125	HP Acquire Method	BNA_G
		HP Processing Method	BG102425
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13181,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BG064653.D	31 Oct 2025 12:42	RC/JU	Ok
2	SSTDCCC040	BG064654.D	31 Oct 2025 13:24	RC/JU	Ok
3	PB170342BL	BG064655.D	31 Oct 2025 14:04	RC/JU	Not Ok
4	Q3447-08	BG064656.D	31 Oct 2025 14:50	RC/JU	Ok
5	Q3497-01	BG064657.D	31 Oct 2025 15:31	RC/JU	Ok
6	Q3504-01	BG064658.D	31 Oct 2025 16:12	RC/JU	Ok
7	Q3496-01	BG064659.D	31 Oct 2025 16:54	RC/JU	Ok
8	Q3477-03	BG064660.D	31 Oct 2025 17:35	RC/JU	Ok
9	Q3495-01	BG064661.D	31 Oct 2025 18:16	RC/JU	Ok
10	Q3495-03	BG064662.D	31 Oct 2025 18:57	RC/JU	Ok
11	Q3503-05	BG064663.D	31 Oct 2025 19:38	RC/JU	Ok
12	Q3503-01	BG064664.D	31 Oct 2025 20:19	RC/JU	Ok
13	Q3486-07	BG064665.D	31 Oct 2025 21:00	RC/JU	Ok,M
14	Q3486-03DL	BG064666.D	31 Oct 2025 21:41	RC/JU	Ok,M
15	Q3503-03	BG064667.D	31 Oct 2025 22:22	RC/JU	Ok
16	Q3501-01	BG064668.D	31 Oct 2025 23:04	RC/JU	Ok
17	Q3499-01	BG064669.D	31 Oct 2025 23:45	RC/JU	Ok
18	Q3486-07DL	BG064670.D	1 Nov 2025 00:27	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP102925

Review By	rahul	Review On	10/30/2025 8:52:33 AM
Supervise By	Jagrut	Supervise On	11/4/2025 3:53:07 PM
SubDirectory	BP102925	HP Acquire Method	BNA_P
		HP Processing Method	BP102925
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13180,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP026027.D	29 Oct 2025 08:45	RC/JU	Ok
2	SSTDICC2.5	BP026028.D	29 Oct 2025 09:26	RC/JU	Ok
3	SSTDICC005	BP026029.D	29 Oct 2025 10:07	RC/JU	Ok
4	SSTDICC010	BP026030.D	29 Oct 2025 10:48	RC/JU	Ok,M
5	SSTDICC020	BP026031.D	29 Oct 2025 11:29	RC/JU	Ok,M
6	SSTDICCC040	BP026032.D	29 Oct 2025 12:10	RC/JU	Ok
7	SSTDICC050	BP026033.D	29 Oct 2025 12:51	RC/JU	Ok,M
8	SSTDICC060	BP026034.D	29 Oct 2025 13:32	RC/JU	Ok
9	SSTDICC080	BP026035.D	29 Oct 2025 14:13	RC/JU	Ok,M
10	SSTDICV040	BP026036.D	29 Oct 2025 16:09	RC/JU	Ok,M
11	PB170259BL	BP026037.D	29 Oct 2025 16:50	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP110325

Review By	rahul	Review On	11/3/2025 3:19:18 PM
Supervise By	mohammad	Supervise On	11/6/2025 12:53:42 AM
SubDirectory	BP110325	HP Acquire Method	BNA_P
		HP Processing Method	BP102925
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13181,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP026052.D	03 Nov 2025 10:08	RC/JU	Ok
2	SSTDCCC040	BP026053.D	03 Nov 2025 11:06	RC/JU	Ok
3	PB170342BL	BP026054.D	03 Nov 2025 11:47	RC/JU	Ok
4	PB170342BS	BP026055.D	03 Nov 2025 12:29	RC/JU	Ok,M
5	PB170346BL	BP026056.D	03 Nov 2025 13:21	RC/JU	Ok
6	PB170346BS	BP026057.D	03 Nov 2025 14:03	RC/JU	Ok,M
7	PB170346BSD	BP026058.D	03 Nov 2025 14:44	RC/JU	Ok,M
8	Q3494-01	BP026059.D	03 Nov 2025 15:32	RC/JU	Ok
9	Q3495-02	BP026060.D	03 Nov 2025 16:13	RC/JU	Ok
10	Q3495-02MS	BP026061.D	03 Nov 2025 16:54	RC/JU	Ok
11	Q3495-02MSD	BP026062.D	03 Nov 2025 17:35	RC/JU	Ok,M
12	Q3519-01	BP026063.D	03 Nov 2025 18:16	RC/JU	Ok,M
13	Q3519-01MS	BP026064.D	03 Nov 2025 18:57	RC/JU	Ok,M
14	Q3519-01MSD	BP026065.D	03 Nov 2025 19:39	RC/JU	Ok
15	Q3515-01	BP026066.D	03 Nov 2025 20:20	RC/JU	Dilution
16	Q3520-01	BP026067.D	03 Nov 2025 21:01	RC/JU	Not Ok

M : Manual Integration