

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG102425

Review By	rahul	Review On	10/27/2025 10:16:00 AM
Supervise By	mohammad	Supervise On	11/5/2025 12:11:16 AM
SubDirectory	BG102425	HP Acquire Method	BNA_G
		HP Processing Method	BG102425
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13179,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064568.D	24 Oct 2025 8:20		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BG064569.D	24 Oct 2025 9:00		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BG064570.D	24 Oct 2025 9:41		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BG064571.D	24 Oct 2025 11:02		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BG064572.D	24 Oct 2025 12:23	Compound #32 kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BG064573.D	24 Oct 2025 13:03	Compound #26,48,51,54,57,65 are kept on LR	RC/JU	Ok,M
7	SSTDICC060	SSTDICC060	BG064574.D	24 Oct 2025 13:44		RC/JU	Ok,M
8	SSTDICC080	SSTDICC080	BG064575.D	24 Oct 2025 14:24		RC/JU	Ok,M
9	SSTDICC050	SSTDICC050	BG064576.D	24 Oct 2025 15:32		RC/JU	Ok,M
10	SSTDICV040	ICVBG102425	BG064577.D	24 Oct 2025 16:47	compound #77 failed in the ICV.	RC/JU	Ok
11	PB170222BL	PB170222BL	BG064578.D	24 Oct 2025 17:28	Analyzed for contamination check	RC/JU	Not Ok
12	Q3405-02	PAD-10-20-25-B	BG064579.D	24 Oct 2025 18:08		RC/JU	Ok
13	Q3405-03	PAD-10-20-25-C	BG064580.D	24 Oct 2025 18:49		RC/JU	Ok
14	Q3405-04	SW-10-20-25	BG064581.D	24 Oct 2025 19:30		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG103125

Review By	rahul	Review On	11/3/2025 3:20:11 PM
Supervise By	mohammad	Supervise On	11/5/2025 12:11:48 AM
SubDirectory	BG103125	HP Acquire Method	BNA_G
		HP Processing Method	BG102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064653.D	31 Oct 2025 12:42		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064654.D	31 Oct 2025 13:24		RC/JU	Ok
3	PB170342BL	PB170342BL	BG064655.D	31 Oct 2025 14:04	Analyzed for contamination check	RC/JU	Not Ok
4	Q3447-08	P001-TW4-251023-01	BG064656.D	31 Oct 2025 14:50		RC/JU	Ok
5	Q3497-01	OK-01-103025	BG064657.D	31 Oct 2025 15:31		RC/JU	Ok
6	Q3504-01	72-11986	BG064658.D	31 Oct 2025 16:12		RC/JU	Ok
7	Q3496-01	DRILL-CUTTINGS	BG064659.D	31 Oct 2025 16:54		RC/JU	Ok
8	Q3477-03	SOUTH-MAHWAH-WA	BG064660.D	31 Oct 2025 17:35		RC/JU	Ok
9	Q3495-01	SB-1025-1	BG064661.D	31 Oct 2025 18:16		RC/JU	Ok
10	Q3495-03	TW-1025-1	BG064662.D	31 Oct 2025 18:57		RC/JU	Ok
11	Q3503-05	324	BG064663.D	31 Oct 2025 19:38		RC/JU	Ok
12	Q3503-01	322	BG064664.D	31 Oct 2025 20:19		RC/JU	Ok
13	Q3486-07	TP6	BG064665.D	31 Oct 2025 21:00		RC/JU	Ok,M
14	Q3486-03DL	TP4DL	BG064666.D	31 Oct 2025 21:41		RC/JU	Ok,M
15	Q3503-03	323	BG064667.D	31 Oct 2025 22:22		RC/JU	Ok
16	Q3501-01	461	BG064668.D	31 Oct 2025 23:04		RC/JU	Ok
17	Q3499-01	RBR251235	BG064669.D	31 Oct 2025 23:45		RC/JU	Ok

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG103125

Review By	rahul	Review On	11/3/2025 3:20:11 PM		
Supervise By	mohammad	Supervise On	11/5/2025 12:11:48 AM		
SubDirectory	BG103125	HP Acquire Method	BNA_G	HP Processing Method	BG102425

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q3486-07DL	TP6DL	BG064670.D	1 Nov 2025 00:27	Not Required	RC/JU	Not Ok
----	------------	-------	------------	------------------	--------------	-------	--------

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP102925

Review By	rahul	Review On	10/30/2025 8:52:33 AM
Supervise By	Jagrut	Supervise On	11/4/2025 3:53:07 PM
SubDirectory	BP102925	HP Acquire Method	BNA_P
		HP Processing Method	BP102925

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13180,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026027.D	29 Oct 2025 08:45		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP026028.D	29 Oct 2025 09:26		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP026029.D	29 Oct 2025 10:07		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BP026030.D	29 Oct 2025 10:48		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP026031.D	29 Oct 2025 11:29		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP026032.D	29 Oct 2025 12:10	Compound #26,48,51,53,57,80,84 are kept on LR & Compound #32,54,65 are kept on QR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP026033.D	29 Oct 2025 12:51		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP026034.D	29 Oct 2025 13:32		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BP026035.D	29 Oct 2025 14:13	Compound #26 is removed from 80ppm & 60PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP102925	BP026036.D	29 Oct 2025 16:09		RC/JU	Ok,M
11	PB170259BL	PB170259BL	BP026037.D	29 Oct 2025 16:50		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP110325

Review By	rahul	Review On	11/3/2025 3:19:18 PM
Supervise By	mohammad	Supervise On	11/6/2025 12:53:42 AM
SubDirectory	BP110325	HP Acquire Method	BNA_P
		HP Processing Method	BP102925

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13181,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026052.D	03 Nov 2025 10:08		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP026053.D	03 Nov 2025 11:06		RC/JU	Ok
3	PB170342BL	PB170342BL	BP026054.D	03 Nov 2025 11:47		RC/JU	Ok
4	PB170342BS	PB170342BS	BP026055.D	03 Nov 2025 12:29		RC/JU	Ok,M
5	PB170346BL	PB170346BL	BP026056.D	03 Nov 2025 13:21		RC/JU	Ok
6	PB170346BS	PB170346BS	BP026057.D	03 Nov 2025 14:03		RC/JU	Ok,M
7	PB170346BSD	PB170346BSD	BP026058.D	03 Nov 2025 14:44		RC/JU	Ok,M
8	Q3494-01	MW3	BP026059.D	03 Nov 2025 15:32		RC/JU	Ok
9	Q3495-02	SB-1025-2	BP026060.D	03 Nov 2025 16:13		RC/JU	Ok
10	Q3495-02MS	SB-1025-2MS	BP026061.D	03 Nov 2025 16:54		RC/JU	Ok
11	Q3495-02MSD	SB-1025-2MSD	BP026062.D	03 Nov 2025 17:35		RC/JU	Ok,M
12	Q3519-01	PL-01-103125	BP026063.D	03 Nov 2025 18:16		RC/JU	Ok,M
13	Q3519-01MS	PL-01-103125MS	BP026064.D	03 Nov 2025 18:57		RC/JU	Ok,M
14	Q3519-01MSD	PL-01-103125MSD	BP026065.D	03 Nov 2025 19:39		RC/JU	Ok
15	Q3515-01	SU-04-103125	BP026066.D	03 Nov 2025 20:20	Need further 2X dilution	RC/JU	Dilution
16	Q3520-01	EO-02-103125	BP026067.D	03 Nov 2025 21:01	Analyze with a lower DF	RC/JU	Not Ok

M : Manual Integration