

Surrogate Summary

SW-846

SDG No.: Q3495

Client: Remington & Vernick Engineers

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170342BL	PB170342BL	Nitrobenzene-d5	100	86.0	86		18	107
		2-Fluorobiphenyl	100	82.9	83		20	109
		Terphenyl-d14	100	85.3	85		10	105
PB170342BS	PB170342BS	Nitrobenzene-d5	100	79.2	79		18	107
		2-Fluorobiphenyl	100	76.7	77		20	109
		Terphenyl-d14	100	80.8	81		10	105
Q3495-01	SB-1025-1	Nitrobenzene-d5	100	39.5	39		18	107
		2-Fluorobiphenyl	100	35.9	36		20	109
		Terphenyl-d14	100	33.9	34		10	105
Q3495-02MS	SB-1025-2MS	Nitrobenzene-d5	100	49.0	49		18	107
		2-Fluorobiphenyl	100	43.1	43		20	109
		Terphenyl-d14	100	43.2	43		10	105
Q3495-02MSD	SB-1025-2MSD	Nitrobenzene-d5	100	46.1	46		18	107
		2-Fluorobiphenyl	100	40.7	41		20	109
		Terphenyl-d14	100	42.9	43		10	105