

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3495

Client ID : SSTDCCC040

Date Analyzed: 10/31/2025

Lab File ID: BG064654.D

Time Analyzed: 13:24

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	29760	8.216	121809	11.04	99146	14.84
UPPER LIMIT	59520	8.716	243618	11.542	198292	15.338
LOWER LIMIT	14880	7.716	60904.5	10.542	49573	14.338
EPA SAMPLE NO.						
01 SB-1025-1	34654	8.22	149932	11.04	117242	14.84
02 TW-1025-1	37784	8.22	154992	11.04	121331	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.