284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: REMI02

Lab Code: ACE SDG No.: Q3495

Instrument ID: BNA_P Calibration Date(s): 10/29/2025 10/29/2025

Calibration Time(s): 09:26 14:13

LAB FILE ID: RRF2.5 = BP026028.D RRF005 = BP026029.D RRF010 = BP026030.D
RRF020 = BP026031.D RRF040 = BP026032.D RRF050 = BP026033.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.055	1.255	1.234	1.249	1.123	1.212	7.4
Benzaldehyde			0.855	0.938	0.809	0.670	0.802	11.5
Phenol-d6		1.366	1.580	1.574	1.606	1.425	1.543	6.8
Phenol		1.542	1.771	1.750	1.773	1.568	1.713	6.4
bis(2-Chloroethyl)ether		1.271	1.434	1.372	1.373	1.228	1.352	5.6
2-Chlorophenol		1.116	1.354	1.356	1.403	1.262	1.342	9.0
2-Methylphenol		0.958	1.129	1.145	1.175	1.057	1.121	7.8
2,2-oxybis(1-Chloropropane)		1.978	2.210	2.102	2.099	1.846	2.043	5.8
Acetophenone		0.481	0.548	0.519	0.518	0.455	0.506	6.1
3+4-Methylphenols			1.521	1.557	1.613	1.430	1.560	5.0
n-Nitroso-di-n-propylamine	0.795	0.894	1.013	1.016	1.027	0.904	0.958	8.8
Nitrobenzene-d5		0.256	0.317	0.327	0.338	0.306	0.321	10.3
Hexachloroethane		0.481	0.562	0.535	0.552	0.490	0.533	6.5
Nitrobenzene		0.282	0.345	0.346	0.357	0.317	0.339	8.8
Isophorone		0.606	0.721	0.705	0.720	0.632	0.687	7.0
2-Nitrophenol		0.069	0.092	0.107	0.128	0.124	0.104	23.2
2,4-Dimethylphenol		0.243	0.306	0.281	0.302	0.269	0.289	8.9
bis(2-Chloroethoxy)methane		0.406	0.466	0.436	0.447	0.396	0.433	5.8
2,4-Dichlorophenol		0.241	0.302	0.308	0.327	0.292	0.305	10.8
Naphthalene		1.039	1.181	1.121	1.115	0.978	1.091	6.3
4-Chloroaniline		0.374	0.440	0.427	0.437	0.384	0.419	6.8
Hexachlorobutadiene		0.198	0.231	0.209	0.216	0.190	0.211	6.7
Caprolactam			0.099	0.103	0.113	0.099	0.107	7.0
4-Chloro-3-methylphenol		0.257	0.314	0.325	0.345	0.303	0.320	10.3
2-Methylnaphthalene		0.717	0.823	0.771	0.795	0.694	0.763	6.0
Hexachlorocyclopentadiene			0.200	0.198	0.237	0.218	0.227	11.5
2,4,6-Trichlorophenol		0.256	0.342	0.372	0.402	0.360	0.367	15.7
2-Fluorobiphenyl		1.350	1.549	1.454	1.421	1.231	1.392	7.5
2,4,5-Trichlorophenol		0.309	0.414	0.425	0.456	0.405	0.419	12.9
1,1-Biphenyl		1.454	1.673	1.577	1.560	1.378	1.531	6.4
2-Chloronaphthalene		1.130	1.304	1.248	1.229	1.090	1.205	6.2
2-Nitroaniline		0.166	0.226	0.269	0.304	0.277	0.272	21.9
Dimethylphthalate		1.282	1.574	1.491	1.522	1.323	1.455	7.8
Acenaphthylene		1.547	1.845	1.832	1.832	1.618	1.756	7.2

All other compounds must meet a minimum RRF of 0.010.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: REMI02

Lab Code: ACE SDG No.: Q3495

Instrument ID: BNA_P Calibration Date(s): 10/29/2025 10/29/2025

Calibration Time(s): 09:26 14:13

LAB FILE ID: RRF2.5 = BP026028.D RRF005 = BP026029.D RRF010 = BP026030.D
RRF020 = BP026031.D RRF040 = BP026032.D RRF050 = BP026033.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.142	0.196	0.234	0.264	0.245	0.237	22.6
3-Nitroaniline		0.177	0.254	0.296	0.324	0.295	0.291	20.7
Acenaphthene		1.101	1.285	1.247	1.249	1.089	1.206	6.8
2,4-Dinitrophenol			0.069	0.082	0.100	0.096	0.100	22.4
4-Nitrophenol			0.277	0.309	0.285	0.258	0.290	7.0
Dibenzofuran		1.725	1.997	1.886	1.869	1.614	1.822	6.9
2,4-Dinitrotoluene		0.195	0.288	0.340	0.393	0.360	0.347	24.0
Diethylphthalate		1.240	1.570	1.484	1.549	1.350	1.463	8.7
4-Chlorophenyl-phenylether		0.699	0.803	0.736	0.731	0.623	0.713	8.3
Fluorene		1.347	1.621	1.501	1.450	1.260	1.427	8.8
4-Nitroaniline		0.207	0.294	0.328	0.342	0.298	0.311	16.9
4,6-Dinitro-2-methylphenol			0.053	0.065	0.076	0.075	0.077	22.1
n-Nitrosodiphenylamine		0.574	0.678	0.641	0.647	0.567	0.625	6.8
2,4,6-Tribromophenol		0.155	0.211	0.218	0.236	0.209	0.216	14.2
4-Bromophenyl-phenylether		0.206	0.234	0.222	0.228	0.204	0.222	5.8
Hexachlorobenzene		0.239	0.273	0.256	0.258	0.226	0.253	6.5
Atrazine		0.177	0.213	0.221	0.219	0.193	0.211	9.6
Pentachlorophenol			0.126	0.143	0.161	0.148	0.154	12.4
Phenanthrene		1.133	1.288	1.209	1.189	1.031	1.164	7.1
Anthracene		1.064	1.262	1.195	1.177	1.044	1.154	7.1
Carbazole		0.985	1.172	1.164	1.129	0.969	1.092	7.7
Di-n-butylphthalate		0.937	1.219	1.241	1.333	1.196	1.222	11.6
Fluoranthene		1.259	1.469	1.426	1.411	1.224	1.358	7.1
Pyrene		1.201	1.488	1.395	1.427	1.217	1.353	8.2
Terphenyl-d14		0.953	1.125	1.013	1.014	0.860	0.984	8.8
Butylbenzylphthalate		0.235	0.337	0.407	0.496	0.468	0.432	26.4
3,3-Dichlorobenzidine			0.390	0.424	0.462	0.416	0.443	8.9
Benzo(a)anthracene		1.279	1.497	1.407	1.420	1.224	1.374	7.0
Chrysene		1.210	1.424	1.348	1.345	1.169	1.302	6.9
Bis(2-ethylhexyl)phthalate		0.453	0.613	0.690	0.829	0.768	0.718	20.3
Di-n-octyl phthalate			0.822	1.025	1.287	1.238	1.191	18.9
Benzo(b)fluoranthene		1.074	1.277	1.278	1.306	1.149	1.242	7.7
Benzo(k)fluoranthene		1.119	1.365	1.306	1.331	1.141	1.268	7.9
Benzo(a)pyrene		0.941	1.169	1.171	1.210	1.054	1.141	9.4
Indeno(1,2,3-cd)pyrene		1.190	1.486	1.539	1.564	1.371	1.476	10.1

All other compounds must meet a minimum RRF of 0.010.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: REMI02

Lab Code: ACE SDG No.: Q3495

Instrument ID: BNA_P Calibration Date(s): 10/29/2025 10/29/2025

Calibration Time(s): 09:26 14:13

LAB FILE ID: RRF2.5 = BP026028.D RRF005 = BP026029.D RRF010 = BP026030.D
RRF020 = BP026031.D RRF040 = BP026032.D RRF050 = BP026033.D

COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo(a,h)anthracene		0.978	1.219	1.249	1.274	1.117	1.202	9.7
Benzo(g,h,i)perylene		0.966	1.167	1.219	1.212	1.068	1.156	8.9
1,2,4,5-Tetrachlorobenzene		0.574	0.670	0.637	0.632	0.565	0.621	6.3
1,4-Dioxane		0.472	0.567	0.519	0.515	0.453	0.511	7.6
2,3,4,6-Tetrachlorophenol		0.223	0.303	0.332	0.364	0.328	0.330	16.7

All other compounds must meet a minimum RRF of 0.010.

Response Factor Report GC/MS Ins

Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
Method File : 8270-BG102425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Nov 03 18:16:26 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BG064569.D 5 =BG064570.D 10 =BG064571.D 20 =BG064572.D 40 =BG064573.D 50 =BG064576.D 60 =BG064574.D 80 =BG0645
75.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------ISTD----------									
2)		1,4-Dioxane		0.570	0.581	0.524	0.525	0.474	0.559	0.460	0.528	8.85
3)		Pyridine		1.630	1.608	1.552	1.669	1.471	1.733	1.427	1.584	6.84
4)		n-Nitrosodimet...			0.912	0.807	0.900	0.780	0.921	0.776	0.850	8.08
5)	S	2-Fluorophenol		1.098	1.207	1.155	1.292	1.139	1.318	1.132	1.192	7.09
6)		Aniline		2.203	2.320	2.131	2.386	2.090	2.440	2.004	2.225	7.28
7)	S	Phenol-d6		1.469	1.654	1.566	1.771	1.597	1.864	1.524	1.635	8.55
8)		2-Chlorophenol		1.051	1.207	1.197	1.373	1.225	1.461	1.216	1.247	10.65
9)		Benzaldehyde			0.942	0.994	0.940	0.836	0.981	0.662	0.893	14.09
10)	C	Phenol		1.665	1.820	1.748	1.944	1.761	2.027	1.693	1.808	7.36
11)		bis(2-Chloroet...		1.307	1.409	1.292	1.383	1.273	1.443	1.198	1.329	6.47
12)		1,3-Dichlorobe...		1.435	1.516	1.394	1.520	1.342	1.542	1.299	1.435	6.62
13)	C	1,4-Dichlorobe...		1.431	1.546	1.442	1.534	1.381	1.561	1.323	1.460	6.21
14)		1,2-Dichlorobe...		1.371	1.518	1.364	1.483	1.316	1.520	1.283	1.408	6.97
15)		Benzyl Alcohol			1.281	1.263	1.425	1.306	1.533	1.283	1.348	7.97
16)		2,2'-oxybis(1-...		3.421	3.604	3.396	3.733	3.301	3.802	3.154	3.487	6.74
17)		2-Methylphenol		1.103	1.225	1.178	1.273	1.159	1.345	1.120	1.201	7.22
18)		Hexachloroethane		0.468	0.535	0.505	0.541	0.496	0.557	0.483	0.512	6.41
19)	P	n-Nitroso-di-n...	1.044	1.038	1.141	1.129	1.227	1.115	1.289	1.057	1.130	7.93
20)		3+4-Methylphenols			1.652	1.581	1.818	1.612	1.895	1.571	1.688	8.04
21)	I	Naphthalene-d8	----------ISTD----------									
22)		Acetophenone		0.535	0.565	0.528	0.584	0.516	0.590	0.509	0.547	6.04
23)	S	Nitrobenzene-d5		0.239	0.286	0.314	0.377	0.356	0.403	0.356	0.333	17.02
24)		Nitrobenzene		0.274	0.312	0.344	0.410	0.377	0.438	0.380	0.362	15.64
25)		Isophorone		0.751	0.796	0.771	0.869	0.764	0.891	0.751	0.799	7.25
26)	C	2-Nitrophenol		0.076	0.095	0.110	0.137	0.135	0.152	0.144	0.121	23.20
27)		2,4-Dimethylph...		0.236	0.294	0.285	0.335	0.297	0.343	0.290	0.297	11.93
28)		bis(2-Chloroet...		0.452	0.464	0.443	0.490	0.426	0.493	0.417	0.455	6.48
29)	C	2,4-Dichloroph...		0.243	0.299	0.315	0.360	0.327	0.373	0.322	0.320	13.34
30)		1,2,4-Trichlor...		0.338	0.369	0.348	0.391	0.349	0.399	0.349	0.363	6.44
31)		Naphthalene		1.083	1.144	1.084	1.177	1.036	1.201	1.037	1.109	5.96
32)		Benzoic acid			0.142	0.164	0.223	0.215	0.260	0.237	0.207	21.74
33)		4-Chloroaniline		0.390	0.456	0.411	0.459	0.410	0.486	0.405	0.431	8.29
34)	C	Hexachlorobuta...		0.266	0.291	0.282	0.299	0.271	0.306	0.273	0.284	5.33
35)		Caprolactam			0.112	0.114	0.134	0.118	0.142	0.116	0.123	9.85
36)	C	4-Chloro-3-met...		0.323	0.381	0.361	0.426	0.376	0.454	0.370	0.384	11.18
37)		2-Methylnaphth...		0.764	0.858	0.797	0.865	0.772	0.888	0.755	0.814	6.74
38)		1-Methylnaphth...		0.798	0.849	0.776	0.853	0.752	0.878	0.735	0.806	6.85

Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
Method File : 8270-BG102425.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.657	0.684	0.648	0.702	0.633	0.717	0.657	0.671	4.55
41)	P	Hexachlorocycl...		0.268	0.279	0.324	0.297	0.340	0.305	0.302	8.97
42)	S	2,4,6-Tribromo...	0.221	0.266	0.272	0.318	0.292	0.346	0.306	0.289	14.04
43)	C	2,4,6-Trichlor...	0.294	0.333	0.375	0.423	0.399	0.456	0.408	0.384	14.41
44)		2,4,5-Trichlor...	0.339	0.425	0.439	0.484	0.443	0.517	0.454	0.443	12.48
45)	S	2-Fluorobiphenyl	1.310	1.389	1.306	1.384	1.235	1.384	1.228	1.319	5.24
46)		1,1'-Biphenyl	1.456	1.450	1.401	1.489	1.326	1.498	1.333	1.422	4.98
47)		2-Chloronaphth...	1.065	1.082	1.053	1.132	1.013	1.145	1.023	1.073	4.71
48)		2-Nitroaniline	0.201	0.234	0.277	0.348	0.343	0.404	0.367	0.311	24.06
49)		Acenaphthylene	1.667	1.726	1.647	1.786	1.593	1.827	1.604	1.693	5.30
50)		Dimethylphthalate	1.389	1.570	1.488	1.641	1.462	1.699	1.462	1.530	7.21
51)		2,6-Dinitrotol...	0.127	0.180	0.193	0.247	0.244	0.283	0.254	0.218	24.69
52)	C	Acenaphthene	1.126	1.178	1.115	1.223	1.093	1.261	1.101	1.157	5.63
53)		3-Nitroaniline	0.180	0.211	0.252	0.294	0.281	0.325	0.297	0.263	19.70
54)	P	2,4-Dinitrophenol		0.086	0.110	0.137	0.139	0.165	0.154	0.132	22.11
55)		Dibenzofuran	1.881	1.972	1.826	1.963	1.734	1.988	1.725	1.870	5.96
56)	P	4-Nitrophenol		0.191	0.213	0.261	0.250	0.283	0.267	0.244	14.40
57)		2,4-Dinitrotol...	0.223	0.269	0.300	0.385	0.373	0.434	0.396	0.340	22.67
58)		Fluorene	1.471	1.559	1.426	1.531	1.356	1.549	1.329	1.460	6.38
59)		2,3,4,6-Tetrac...	0.312	0.385	0.402	0.466	0.444	0.508	0.449	0.424	15.04
60)		Diethylphthalate	1.417	1.645	1.514	1.657	1.486	1.718	1.502	1.563	7.06
61)		4-Chlorophenyl...	0.800	0.864	0.792	0.866	0.763	0.888	0.763	0.820	6.38
62)		4-Nitroaniline	0.194	0.252	0.286	0.327	0.310	0.355	0.324	0.293	18.55
63)		Azobenzene	1.362	1.504	1.377	1.538	1.338	1.564	1.353	1.434	6.80
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.055	0.064	0.083	0.083	0.097	0.093	0.079	21.00
66)	c	n-Nitrosodiphe...	0.510	0.551	0.525	0.583	0.518	0.594	0.500	0.540	6.79
67)		4-Bromophenyl-...	0.218	0.236	0.230	0.258	0.229	0.264	0.224	0.237	7.28
68)		Hexachlorobenzene	0.254	0.266	0.262	0.292	0.260	0.302	0.256	0.270	6.99
69)		Atrazine	0.215	0.227	0.249	0.245	0.219	0.261	0.211	0.232	8.27
70)	C	Pentachlorophenol		0.132	0.152	0.180	0.172	0.199	0.177	0.169	13.91
71)		Phenanthrene	1.084	1.133	1.073	1.162	1.020	1.179	1.005	1.094	6.17
72)		Anthracene	1.087	1.144	1.101	1.187	1.041	1.204	1.025	1.113	6.17
73)		Carbazole	0.982	1.037	0.982	1.031	0.910	1.028	0.897	0.981	5.90
74)		Di-n-butylphth...	1.011	1.156	1.179	1.240	1.084	1.198	1.075	1.134	7.11
75)	C	Fluoranthene	1.385	1.463	1.380	1.440	1.273	1.403	1.247	1.370	5.91
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine		0.392	0.475	0.490	0.416	0.533	0.383	0.448	13.40
78)		Pyrene	1.287	1.331	1.284	1.411	1.233	1.406	1.196	1.307	6.24
79)	S	Terphenyl-d14	1.074	1.120	1.066	1.142	0.987	1.102	0.928	1.060	7.21
80)		Butylbenzylpht...	0.284	0.359	0.388	0.450	0.408	0.467	0.410	0.395	15.43
81)		Benzo(a)anthra...	1.294	1.355	1.292	1.416	1.253	1.430	1.243	1.326	5.71
82)		3,3'-Dichlorob...		0.418	0.428	0.477	0.424	0.486	0.424	0.443	6.88
83)		Chrysene	1.249	1.291	1.229	1.344	1.188	1.359	1.173	1.262	5.76
84)		Bis(2-ethylhex...	0.484	0.555	0.581	0.671	0.585	0.667	0.588	0.590	10.95
85)	c	Di-n-octyl pht...		0.871	0.932	1.087	0.976	1.117	0.986	0.995	9.32

Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
Method File : 8270-BG102425.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.383	1.491	1.410	1.568	1.396	1.642	1.410	1.471	6.80
88)		Benzo(b)fluora...	1.138	1.246	1.169	1.322	1.173	1.353	1.183	1.226	6.77
89)		Benzo(k)fluora...	1.175	1.260	1.206	1.285	1.151	1.364	1.149	1.227	6.51
90)	C	Benzo(a)pyrene	1.061	1.144	1.093	1.202	1.067	1.246	1.079	1.127	6.42
91)		Dibenzo(a,h)an...	1.120	1.223	1.158	1.268	1.130	1.324	1.143	1.195	6.55
92)		Benzo(g,h,i)pe...	1.080	1.162	1.089	1.213	1.078	1.280	1.092	1.142	6.95

(#) = Out of Range

Response Factor Report MSBNA_P

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP102925.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Oct 30 11:51:34 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BP026028.D 5 =BP026029.D 10 =BP026030.D 20 =BP026031.D 40 =BP026032.D 50 =BP026033.D 60 =BP026034.D 80 =BP0260
35.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------	ISTD	----------							
2)		1,4-Dioxane		0.472	0.567	0.519	0.515	0.453	0.541	0.510	0.511	7.57
3)		Pyridine		1.330	1.538	1.488	1.481	1.304	1.552	1.475	1.453	6.69
4)		n-Nitrosodimet...			0.601	0.591	0.584	0.522	0.611	0.584	0.582	5.39
5)	S	2-Fluorophenol		1.055	1.255	1.234	1.249	1.123	1.318	1.250	1.212	7.44
6)		Aniline		1.893	2.157	2.117	2.112	1.885	2.158	2.091	2.059	5.76
7)	S	Phenol-d6		1.366	1.580	1.574	1.606	1.425	1.644	1.604	1.543	6.77
8)		2-Chlorophenol		1.116	1.354	1.356	1.403	1.262	1.478	1.426	1.342	8.97
9)		Benzaldehyde			0.856	0.937	0.809	0.670	0.796	0.744	0.802	11.45
10)	C	Phenol		1.542	1.771	1.750	1.773	1.568	1.823	1.761	1.713	6.44
11)		bis(2-Chloroet...		1.271	1.434	1.372	1.373	1.228	1.419	1.363	1.352	5.57
12)		1,3-Dichlorobe...		1.466	1.660	1.572	1.573	1.385	1.610	1.524	1.541	6.00
13)	C	1,4-Dichlorobe...		1.491	1.681	1.585	1.578	1.399	1.633	1.542	1.558	5.97
14)		1,2-Dichlorobe...		1.412	1.595	1.517	1.515	1.338	1.550	1.481	1.487	5.83
15)		Benzyl Alcohol			1.070	1.099	1.140	1.020	1.179	1.165	1.112	5.47
16)		2,2'-oxybis(1-...		1.978	2.210	2.102	2.099	1.846	2.083	1.982	2.043	5.75
17)		2-Methylphenol		0.958	1.129	1.145	1.175	1.057	1.210	1.176	1.121	7.75
18)		Hexachloroethane		0.481	0.562	0.535	0.552	0.490	0.568	0.546	0.533	6.49
19)	P	n-Nitroso-di-n...	0.795	0.894	1.013	1.016	1.027	0.904	1.025	0.986	0.958	8.83
20)		3+4-Methylphenols			1.521	1.557	1.613	1.430	1.646	1.595	1.560	4.96
21)	I	Naphthalene-d8	----------	ISTD	----------							
22)		Acetophenone		0.481	0.548	0.519	0.518	0.455	0.525	0.494	0.506	6.14
23)	S	Nitrobenzene-d5		0.256	0.317	0.327	0.338	0.306	0.359	0.343	0.321	10.34
24)		Nitrobenzene		0.282	0.345	0.346	0.357	0.317	0.371	0.354	0.339	8.83
25)		Isophorone		0.606	0.721	0.705	0.720	0.632	0.730	0.693	0.687	7.03
26)	C	2-Nitrophenol		0.069	0.092	0.107	0.128	0.124			0.104	23.21
27)		2,4-Dimethylph...		0.243	0.306	0.281	0.302	0.269	0.315	0.305	0.289	8.90
28)		bis(2-Chloroet...		0.406	0.466	0.436	0.447	0.396	0.451	0.427	0.433	5.79
29)	C	2,4-Dichloroph...		0.241	0.302	0.308	0.327	0.292	0.340	0.324	0.305	10.75
30)		1,2,4-Trichlor...		0.306	0.361	0.337	0.339	0.301	0.347	0.328	0.331	6.52
31)		Naphthalene		1.039	1.181	1.121	1.115	0.978	1.137	1.063	1.091	6.28
32)		Benzoic acid			0.100	0.127	0.168	0.162	0.206	0.213	0.163	26.84
33)		4-Chloroaniline		0.374	0.440	0.427	0.437	0.384	0.445	0.427	0.419	6.80
34)	C	Hexachlorobuta...		0.198	0.231	0.209	0.216	0.190	0.222	0.207	0.211	6.70
35)		Caprolactam			0.099	0.103	0.113	0.099	0.116	0.112	0.107	6.98
36)	C	4-Chloro-3-met...		0.257	0.314	0.325	0.345	0.303	0.356	0.337	0.320	10.33
37)		2-Methylnaphth...		0.717	0.823	0.771	0.795	0.694	0.793	0.750	0.763	6.02
38)		1-Methylnaphth...		0.709	0.801	0.754	0.768	0.663	0.770	0.726	0.742	6.19

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP102925.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.574	0.670	0.637	0.632	0.565	0.654	0.619	0.621	6.33
41)	P	Hexachlorocycl...		0.200	0.198	0.237	0.218	0.260	0.251	0.227	11.46
42)	S	2,4,6-Tribromo...	0.155	0.211	0.218	0.236	0.209	0.250	0.234	0.216	14.19
43)	C	2,4,6-Trichlor...	0.256	0.342	0.372	0.402	0.360	0.427	0.413	0.367	15.70
44)		2,4,5-Trichlor...	0.309	0.414	0.425	0.456	0.405	0.469	0.456	0.419	12.90
45)	S	2-Fluorobiphenyl	1.350	1.549	1.454	1.421	1.231	1.432	1.307	1.392	7.53
46)		1,1'-Biphenyl	1.454	1.673	1.577	1.560	1.378	1.585	1.493	1.531	6.36
47)		2-Chloronaphth...	1.130	1.304	1.248	1.229	1.090	1.253	1.180	1.205	6.23
48)		2-Nitroaniline	0.166	0.226	0.269	0.304	0.277	0.332	0.329	0.272	21.95
49)		Acenaphthylene	1.547	1.845	1.832	1.832	1.618	1.876	1.742	1.756	7.23
50)		Dimethylphthalate	1.282	1.574	1.491	1.522	1.323	1.554	1.438	1.455	7.81
51)		2,6-Dinitrotol...	0.142	0.196	0.234	0.264	0.245	0.294	0.287	0.237	22.61
52)	C	Acenaphthene	1.101	1.285	1.247	1.249	1.089	1.278	1.192	1.206	6.77
53)		3-Nitroaniline	0.177	0.254	0.296	0.324	0.295	0.351	0.342	0.291	20.65
54)	P	2,4-Dinitrophenol		0.069	0.082	0.100	0.096	0.123	0.126	0.100	22.44
55)		Dibenzofuran	1.725	1.997	1.886	1.869	1.614	1.884	1.776	1.822	6.94
56)	P	4-Nitrophenol		0.277	0.309	0.285	0.258	0.310	0.301	0.290	7.04
57)		2,4-Dinitrotol...	0.195	0.288	0.340	0.393	0.360	0.433	0.418	0.347	23.95
58)		Fluorene	1.347	1.621	1.501	1.450	1.260	1.492	1.319	1.427	8.77
59)		2,3,4,6-Tetrac...	0.223	0.303	0.332	0.364	0.328	0.386	0.370	0.330	16.67
60)		Diethylphthalate	1.240	1.570	1.484	1.549	1.350	1.591	1.454	1.463	8.74
61)		4-Chlorophenyl...	0.699	0.803	0.736	0.731	0.623	0.739	0.658	0.713	8.29
62)		4-Nitroaniline	0.207	0.294	0.328	0.342	0.298	0.368	0.339	0.311	16.92
63)		Azobenzene	1.206	1.475	1.365	1.379	1.188	1.387	1.279	1.325	7.90
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.053	0.065	0.076	0.075	0.096	0.096	0.077	22.06
66)	c	n-Nitrosodiphe...	0.574	0.678	0.641	0.647	0.567	0.659	0.611	0.625	6.80
67)		4-Bromophenyl-...	0.206	0.234	0.222	0.228	0.204	0.237	0.222	0.222	5.77
68)		Hexachlorobenzene	0.239	0.273	0.256	0.258	0.226	0.268	0.247	0.253	6.53
69)		Atrazine	0.177	0.213	0.221	0.219	0.193	0.239	0.216	0.211	9.57
70)	C	Pentachlorophenol		0.126	0.143	0.161	0.148	0.178	0.170	0.154	12.39
71)		Phenanthrene	1.133	1.288	1.209	1.189	1.031	1.195	1.105	1.164	7.12
72)		Anthracene	1.064	1.262	1.195	1.177	1.044	1.220	1.115	1.154	7.09
73)		Carbazole	0.985	1.172	1.164	1.129	0.969	1.148	1.078	1.092	7.74
74)		Di-n-butylphth...	0.937	1.219	1.241	1.333	1.196	1.378	1.252	1.222	11.57
75)	C	Fluoranthene	1.259	1.469	1.426	1.411	1.224	1.422	1.295	1.358	7.09
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine		0.411	0.508	0.517	0.451	0.598	0.560	0.508	13.50
78)		Pyrene	1.201	1.488	1.395	1.427	1.217	1.431	1.310	1.353	8.25
79)	S	Terphenyl-d14	0.953	1.125	1.013	1.014	0.860	1.019	0.908	0.984	8.77
80)		Butylbenzylpht...	0.235	0.337	0.407	0.496	0.468	0.559	0.520	0.432	26.44
81)		Benzo(a)anthra...	1.279	1.497	1.407	1.420	1.224	1.443	1.350	1.374	6.97
82)		3,3'-Dichlorob...		0.390	0.424	0.462	0.416	0.497	0.467	0.443	8.87
83)		Chrysene	1.210	1.424	1.348	1.345	1.169	1.349	1.268	1.302	6.92
84)		Bis(2-ethylhex...	0.453	0.613	0.690	0.829	0.768	0.866	0.808	0.718	20.27
85)	c	Di-n-octyl pht...		0.822	1.025	1.287	1.238	1.407	1.364	1.191	18.86

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP102925.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.190	1.491	1.538	1.563	1.369	1.629	1.555	1.477	10.15
88)		Benzo(b)fluora...	1.074	1.277	1.278	1.306	1.149	1.345	1.262	1.242	7.67
89)		Benzo(k)fluora...	1.119	1.365	1.306	1.331	1.141	1.347	1.267	1.268	7.86
90)	C	Benzo(a)pyrene	0.941	1.169	1.171	1.210	1.054	1.253	1.186	1.141	9.38
91)		Dibenzo(a,h)an...	0.978	1.219	1.249	1.274	1.117	1.320	1.255	1.202	9.72
92)		Benzo(g,h,i)pe...	0.966	1.167	1.219	1.212	1.068	1.259	1.204	1.156	8.92

(#) = Out of Range