

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064653.D
Instrument ID: BNA_G

Contract: REMI02
SDG NO.: Q3495
DFTPP Injection Date: 10/31/2025
DFTPP Injection Time: 12:42

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.2) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.2
365	Greater than 1% of mass 198	5.2
441	Present, but less than mass 443	81
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	20 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG064654.D	10/31/2025	13:24
TW-1025-1	Q3495-03	BG064662.D	10/31/2025	18:57