

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP026052.D
Instrument ID: BNA_P

Contract: REMI02
SDG NO.: Q3495
DFTPP Injection Date: 11/03/2025
DFTPP Injection Time: 10:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	80.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.1 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP026053.D	11/03/2025	11:06
PB170342BL	PB170342BL	BP026054.D	11/03/2025	11:47
PB170342BS	PB170342BS	BP026055.D	11/03/2025	12:29
PB170346BL	PB170346BL	BP026056.D	11/03/2025	13:21
PB170346BS	PB170346BS	BP026057.D	11/03/2025	14:03
PB170346BSD	PB170346BSD	BP026058.D	11/03/2025	14:44
SB-1025-2	Q3495-02	BP026060.D	11/03/2025	16:13
SB-1025-2MS	Q3495-02MS	BP026061.D	11/03/2025	16:54
SB-1025-2MSD	Q3495-02MSD	BP026062.D	11/03/2025	17:35