

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI02
Lab Code: ACE SDG NO.: Q3495
Lab File ID: VX048416.D Date Analyzed: 10/31/2025
Instrument ID: MSVOA_X Time Analyzed: 09:18
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	152790	5.53	255523	6.74	229228	10.04
UPPER LIMIT	305580	6.031	511046	7.239	458456	10.537
LOWER LIMIT	76395	5.031	127762	6.239	114614	9.537
EPA SAMPLE NO.						
TW-1025-1	121387	5.54	239256	6.75	246924	10.04
VX1031WBL01	109923	5.54	221003	6.75	229713	10.04
VX1031WBS01	179964	5.53	315014	6.74	292819	10.04
VX1031WBSD01	161570	5.53	274506	6.74	244643	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.